

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121620\
 Data File : BM028136.D
 Acq On : 15 Dec 2020 20:34
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICV040

Quant Time: Dec 16 02:01:01 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 16 01:58:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	92	0.00
2	1,4-Dioxane	40.000	34.242	14.4	88	0.00
3	Pyridine	40.000	35.984	10.0	93	0.00
4	n-Nitrosodimethylamine	40.000	34.660	13.4	91	0.00
5 S	2-Fluorophenol	80.000	71.664	10.4	92	0.00
6	Aniline	40.000	36.747	8.1	95	0.00
7 S	Phenol-d6	80.000	73.797	7.8	95	0.00
8	2-Chlorophenol	40.000	36.430	8.9	93	0.00
9	Benzaldehyde	40.000	33.418	16.5	94	0.00
10 C	Phenol	40.000	36.824	7.9	95	0.00
11	bis(2-Chloroethyl)ether	40.000	35.824	10.4	93	0.00
12	1,3-Dichlorobenzene	40.000	35.655	10.9	91	0.00
13 C	1,4-Dichlorobenzene	40.000	35.702	10.7	91	0.00
14	1,2-Dichlorobenzene	40.000	35.699	10.8	92	0.00
15	Benzyl Alcohol	40.000	37.901	5.2	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.426	11.4	92	0.00
17	2-Methylphenol	40.000	37.192	7.0	95	0.00
18	Hexachloroethane	40.000	35.548	11.1	89	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.614	6.0	97	0.00
20	3+4-Methylphenols	40.000	38.017	5.0	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	35.229	11.9	97	0.00
23 S	Nitrobenzene-d5	80.000	71.055	11.2	96	0.00
24	Nitrobenzene	40.000	35.327	11.7	96	0.00
25	Isophorone	40.000	36.606	8.5	98	0.00
26 C	2-Nitrophenol	40.000	37.535	6.2	97	0.00
27	2,4-Dimethylphenol	40.000	35.546	11.1	97	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.625	13.4	95	0.00
29 C	2,4-Dichlorophenol	40.000	36.375	9.1	97	0.00
30	1,2,4-Trichlorobenzene	40.000	34.297	14.3	94	0.00
31	Naphthalene	40.000	34.832	12.9	95	0.00
32	Benzoic acid	40.000	39.773	0.6	105	0.00
33	4-Chloroaniline	40.000	36.123	9.7	99	0.00
34 C	Hexachlorobutadiene	40.000	33.787	15.5	92	0.00
35	Caprolactam	40.000	40.401	-1.0	108	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.983	7.5	101	0.00
37	2-Methylnaphthalene	40.000	35.442	11.4	97	0.00
38	1-Methylnaphthalene	40.000	35.462	11.3	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	34.239	14.4	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	35.187	12.0	93	0.00
42 S	2,4,6-Tribromophenol	80.000	73.314	8.4	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.266	9.3	100	0.00
44	2,4,5-Trichlorophenol	40.000	36.234	9.4	101	0.00
45 S	2-Fluorobiphenyl	80.000	68.510	14.4	97	0.00
46	1,1'-Biphenyl	40.000	34.293	14.3	98	0.00
47	2-Chloronaphthalene	40.000	34.479	13.8	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121620\
 Data File : BM028136.D
 Acq On : 15 Dec 2020 20:34
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICV040

Quant Time: Dec 16 02:01:01 2020
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 16 01:58:38 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	37.592	6.0	103	0.00
49	Acenaphthylene	40.000	35.473	11.3	100	0.00
50	Dimethylphthalate	40.000	35.113	12.2	101	0.00
51	2,6-Dinitrotoluene	40.000	36.660	8.4	103	0.00
52 C	Acenaphthene	40.000	35.244	11.9	101	0.00
53	3-Nitroaniline	40.000	37.515	6.2	105	0.00
54 P	2,4-Dinitrophenol	40.000	37.378	6.6	106	0.00
55	Dibenzofuran	40.000	34.643	13.4	100	0.00
56 P	4-Nitrophenol	40.000	38.341	4.1	108	0.00
57	2,4-Dinitrotoluene	40.000	37.822	5.4	103	0.00
58	Fluorene	40.000	35.088	12.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.152	9.6	101	0.00
60	Diethylphthalate	40.000	35.340	11.6	100	0.00
61	4-Chlorophenyl-phenylether	40.000	34.678	13.3	99	0.00
62	4-Nitroaniline	40.000	38.241	4.4	107	0.00
63	Azobenzene	40.000	35.417	11.5	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.513	8.7	105	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.036	12.4	102	0.00
67	4-Bromophenyl-phenylether	40.000	34.383	14.0	99	0.00
68	Hexachlorobenzene	40.000	34.731	13.2	101	0.00
69	Atrazine	40.000	35.871	10.3	102	0.00
70 C	Pentachlorophenol	40.000	37.490	6.3	104	0.00
71	Phenanthrene	40.000	34.795	13.0	103	0.00
72	Anthracene	40.000	35.283	11.8	103	0.00
73	Carbazole	40.000	35.839	10.4	106	0.00
74	Di-n-butylphthalate	40.000	35.842	10.4	100	0.00
75 C	Fluoranthene	40.000	35.549	11.1	106	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.00
77	Benzidine	40.000	40.582	-1.5	107	0.00
78	Pyrene	40.000	35.543	11.1	106	0.00
79 S	Terphenyl-d14	80.000	69.227	13.5	104	0.00
80	Butylbenzylphthalate	40.000	36.865	7.8	104	0.00
81	Benzo(a)anthracene	40.000	35.424	11.4	110	0.00
82	3,3'-Dichlorobenzidine	40.000	36.493	8.8	112	0.00
83	Chrysene	40.000	34.983	12.5	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.431	8.9	103	0.00
85 c	Di-n-octyl phthalate	40.000	36.554	8.6	106	0.00
86 I	Perylene-d12	20.000	20.000	0.0	111	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	32.593	18.5	102	0.00
88	Benzo(b)fluoranthene	40.000	35.378	11.6	110	0.00
89	Benzo(k)fluoranthene	40.000	36.875	7.8	112	0.00
90 C	Benzo(a)pyrene	40.000	35.569	11.1	110	0.00
91	Dibenzo(a,h)anthracene	40.000	32.687	18.3	102	0.00
92	Benzo(g,h,i)perylene	40.000	32.273	19.3	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121620\
Data File : BM028136.D
Acq On : 15 Dec 2020 20:34
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICV040

Quant Time: Dec 16 02:01:01 2020
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM121620.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 16 01:58:38 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0