

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121715\
 Data File : BM003599.D
 Acq On : 17 Dec 2015 15:02
 Operator : UM/IZ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.459

Quant Time: Dec 18 00:31:18 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 00:27:21 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	60	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.175	-2.3	61	0.00
3	1,4-Dioxane	0.218	0.217	0.5	60	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	61	0.00
5	Naphthalene	0.901	0.958	-6.3	64	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.486	-1.5	61	0.00
7	2-Methylnaphthalene	0.605	0.628	-3.8	62	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	60	0.00
9	Acenaphthylene	1.528	1.662	-8.8	65	0.00
10 C	Acenaphthene	1.232	1.306	-6.0	62	0.00
11	Fluorene	1.365	1.526	-11.8	65	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	57	0.00
13	Pentachlorophenol	0.057	0.071	-24.6#	72	0.00
14	Phenanthrene	1.032	1.253	-21.4#	67	0.00
15	Anthracene	0.893	1.029	-15.2	64	0.00
16 SURR	Fluoranthene-d10	0.849	0.948	-11.7	61	0.00
17 C	Fluoranthene	1.120	1.344	-20.0	65	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	56	0.00
19	Pyrene	1.080	1.313	-21.6#	64	0.00
20	Benzo(a)anthracene	0.977	1.172	-20.0	65	0.00
21	Chrysene	1.119	1.139	-1.8	59	0.00
22 I	Perylene-d12	1.000	1.000	0.0	76	0.00
23	Benzo(b)fluoranthene	1.365	1.220	10.6	65	0.00
24	Benzo(k)fluoranthene	1.237	1.192	3.6	74	0.00
25 C	Benzo(a)pyrene	1.152	1.188	-3.1	78	0.00
26	Indeno(1,2,3-cd)pyrene	1.335	1.385	-3.7	83	0.00
27	Dibenzo(a,h)anthracene	1.075	1.125	-4.7	84	0.00
28	Benzo(q,h,i)perylene	1.162	1.192	-2.6	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0