

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121715\
 Data File : BM003608.D
 Acq On : 18 Dec 2015 01:49
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.460

Quant Time: Dec 18 04:19:49 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 00:41:46 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2 SURR	1,4-Dioxane-d8	0.171	0.166	2.9	66	0.00
3	1,4-Dioxane	0.218	0.206	5.5	65	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
5	Naphthalene	0.901	0.918	-1.9	75	0.00
6 SURR	2-Methylnaphthalene-d10	0.479	0.498	-4.0	77	0.00
7	2-Methylnaphthalene	0.605	0.653	-7.9	79	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	83	0.00
9	Acenaphthylene	1.528	1.526	0.1	83	0.00
10 C	Acenaphthene	1.232	1.303	-5.8	85	0.00
11	Fluorene	1.365	1.510	-10.6	89	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
13	Pentachlorophenol	0.057	0.047	17.5	75	0.00
14	Phenanthrene	1.032	1.121	-8.6	94	0.00
15	Anthracene	0.893	0.972	-8.8	95	0.00
16 SURR	Fluoranthene-d10	0.849	0.957	-12.7	96	0.00
17 C	Fluoranthene	1.120	1.380	-23.2#	104	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
19	Pyrene	1.080	1.267	-17.3	99	0.00
20	Benzo(a)anthracene	0.977	1.113	-13.9	99	0.00
21	Chrysene	1.119	1.142	-2.1	94	0.00
22 I	Perylene-d12	1.000	1.000	0.0	98	-0.01
23	Benzo(b)fluoranthene	1.365	1.354	0.8	93	0.00
24	Benzo(k)fluoranthene	1.237	1.242	-0.4	99	-0.01
25 C	Benzo(a)pyrene	1.152	1.184	-2.8	99	-0.01
26	Indeno(1,2,3-cd)pyrene	1.335	1.178	11.8	91	-0.01
27	Dibenzo(a,h)anthracene	1.075	0.948	11.8	91	0.00
28	Benzo(q,h,i)perylene	1.162	1.002	13.8	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1