

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121715\
 Data File : BM003608.D
 Acq On : 18 Dec 2015 01:49
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.460

Quant Time: Dec 18 04:19:49 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 00:41:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	3676	0.40	ng/ul	0.00
4) Naphthalene-d8	10.50	136	15188	0.40	ng/ul	0.00
8) Acenaphthene-d10	14.35	164	8754	0.40	ng/ul	0.00
12) Phenanthrene-d10	17.11	188	20654	0.40	ng/ul	0.00
18) Chrysene-d12	21.31	240	21567	0.40	ng/ul	0.00
22) Perylene-d12	23.55	264	16451	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
2) 1,4-Dioxane-d8	3.17	96	3047	1.94	ng/uL	0.00
6) 2-Methylnaphthalene-d10	12.09	152	7563	0.42	ng/ul	0.00
16) Fluoranthene-d10	19.14	212	19768	0.45	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) 1,4-Dioxane	3.20	88	3789	1.89	ng/ul#	19
5) Naphthalene	10.55	128	13949	0.41	ng/ul	98
7) 2-Methylnaphthalene	12.16	142	9923	0.43	ng/ul	99
9) Acenaphthylene	14.07	152	13357	0.40	ng/ul#	91
10) Acenaphthene	14.42	153	11408	0.42	ng/ul	96
11) Fluorene	15.41	166	13216	0.44	ng/ul	95
13) Pentachlorophenol	16.76	266	977	0.33	ng/ul	96
14) Phenanthrene	17.14	178	23145	0.43	ng/ul	96
15) Anthracene	17.23	178	20075	0.44	ng/ul	96
17) Fluoranthene	19.17	202	28496	0.49	ng/ul	92
19) Pyrene	19.54	202	27315	0.47	ng/ul	93
20) Benzo(a)anthracene	21.29	228	23998	0.46	ng/ul	98
21) Chrysene	21.34	228	24633	0.41	ng/ul	98
23) Benzo(b)fluoranthene	22.87	252	22271	0.40	ng/ul	97
24) Benzo(k)fluoranthene	22.91	252	20432	0.40	ng/ul	97
25) Benzo(a)pyrene	23.44	252	19478	0.41	ng/ul	97
26) Indeno(1,2,3-cd)pyrene	25.81	276	19374	0.35	ng/ul#	91
27) Dibenzo(a,h)anthracene	25.83	278	15589	0.35	ng/ul	98
28) Benzo(g,h,i)perylene	26.51	276	16476	0.34	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121715\
Data File : BM003608.D
Acq On : 18 Dec 2015 01:49
Operator : UM/IZ
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.460

Quant Time: Dec 18 04:19:49 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM121515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 18 00:41:46 2015
Response via : Initial Calibration

