

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121719\
 Data File : BM024153.D
 Acq On : 18 Dec 2019 15:37
 Operator : JU
 Sample : K6170-15
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BW3

Quant Time: Dec 18 16:17:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 17 15:54:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.45	152	310182	20.00	ng/ul	0.00
18) Naphthalene-d8	10.22	136	1453797	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.11	164	967204	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.87	188	2076226	20.00	ng/ul	0.00
78) Chrysene-d12	21.08	240	1866670	20.00	ng/ul	0.00
86) Perylene-d12	23.22	264	1817231	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.01	96	47344	6.74	ng/uL	0.00
5) Phenol-d5	6.65	99	189268	6.44	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.80	67	480499	27.28	ng/ul	0.00
9) 2-Chlorophenol-d4	6.99	132	484476	23.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.17	113	360508	15.02	ng/ul	0.00
19) Nitrobenzene-d5	8.60	128	299165	28.64	ng/ul	0.00
22) 2-Nitrophenol-d4	9.32	143	337550	29.39	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.86	165	597573	26.76	ng/ul	0.00
29) 4-Chloroaniline-d4	10.40	131	27071	1.00	ng/ul	0.02
44) Dimethylphthalate-d6	13.53	166	2326898	32.53	ng/ul	0.00
47) Acenaphthylene-d8	13.80	160	2629117	30.68	ng/ul	0.00
52) 4-Nitrophenol-d4	14.37	143	55823	4.35	ng/ul	0.02
58) Fluorene-d10	15.12	176	2023469	32.27	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.26	200	343130	27.09	ng/ul	0.00
71) Anthracene-d10	16.97	188	3363653	35.61	ng/ul	0.00
79) Pyrene-d10	19.27	212	3716293	41.79	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.08	264	3360141	37.23	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM121719\
Data File : BM024153.D
Acq On : 18 Dec 2019 15:37
Operator : JU
Sample : K6170-15
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BW3

Quant Time: Dec 18 16:17:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 17 15:54:50 2019
Response via : Initial Calibration

