

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121724\
 Data File : BM049040.D
 Acq On : 18 Dec 2024 02:35
 Operator : RC/JU
 Sample : P5155-03
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E28Q8

Quant Time: Dec 18 04:08:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 17 10:03:27 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.573	152	13321	0.400	ng/ul	0.00
4) Naphthalene-d8	10.336	136	43060	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.205	164	24246	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.950	188	49435m	0.400	ng/ul	0.00
17) Chrysene-d12	21.146	240	35518m	0.400	ng/ul	0.00
23) Perylene-d12	23.306	264	36086m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	42617	2.737	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.930	152	12143	0.221	ng/ul	0.00
18) Fluoranthene-d10	18.987	212	24743m	0.233	ng/ul	0.00
Target Compounds						Qvalue
5) Naphthalene	10.380	128	2946	0.027	ng/ul#	76
7) 2-Methylnaphthalene	12.007	142	3367	0.050	ng/ul	96
8) 1-Methylnaphthalene	12.227	142	3186	0.046	ng/ul	95
15) Phenanthrene	16.992	178	7554m	0.056	ng/ul	
19) Fluoranthene	19.015	202	2984m	0.020	ng/ul	
20) Pyrene	19.377	202	4516m	0.028	ng/ul	
22) Chrysene	21.181	228	4811m	0.034	ng/ul	
29) Benzo(g,h,i)perylene	26.101	276	6288m	0.048	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121724\
Data File : BM049040.D
Acq On : 18 Dec 2024 02:35
Operator : RC/JU
Sample : P5155-03
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E28Q8

Quant Time: Dec 18 04:08:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM120924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 17 10:03:27 2024
Response via : Initial Calibration

