

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121815\
 Data File : BM003610.D
 Acq On : 18 Dec 2015 13:06
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 18 16:16:58 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 16:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	54	-0.02
2	1,4-Dioxane	0.448	0.408	8.9	55	-0.03
3	n-Nitrosodimethylamine	0.461	0.467	-1.3	57	-0.03
4 S	2-Fluorophenol	0.997	0.973	2.4	59	-0.02
5 S	Phenol-d6	1.227	1.168	4.8	54	0.00
6	bis(2-Chloroethyl)ether	1.193	1.094	8.3	52	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	52	-0.02
8 S	Nitrobenzene-d5	0.233	0.238	-2.1	57	-0.02
9	Nitrobenzene	0.252	0.256	-1.6	55	-0.02
10	Naphthalene	1.077	1.022	5.1	53	-0.02
11	Hexachlorobutadiene	0.164	0.165	-0.6	56	-0.02
12	2-Methylnaphthalene	0.738	0.723	2.0	53	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	51	-0.02
14 S	2,4,6-Tribromophenol	0.104	0.119	-14.4	58	0.00
15 S	2-Fluorobiphenyl	1.505	1.413	6.1	53	-0.02
16	Acenaphthylene	1.976	2.049	-3.7	55	-0.02
17	Acenaphthene	1.321	1.275	3.5	55	-0.02
18	Fluorene	1.490	1.468	1.5	53	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	42#	-0.03
20	4-Bromophenyl-phenylether	0.144	0.218	-51.4#	74	-0.03
21	Hexachlorobenzene	0.173	0.210	-21.4	59	0.00
22	Pentachlorophenol	0.055	0.055	0.0	45#	-0.03
23	Phenanthrene	0.964	1.407	-46.0#	69	-0.03
24	Anthracene	1.016	1.023	-0.7	43#	-0.03
25	Fluoranthene	0.999	1.396	-39.7#	58	-0.02
26 I	Chrysene-d12	1.000	1.000	0.0	63	-0.02
27	Benzydine	0.409	0.395	3.4	69	0.00
28	Pvrene	1.812	1.487	17.9	56	-0.02
29 S	Terphenyl-d14	0.811	0.697	14.1	57	-0.02
30	Benzo(a)anthracene	1.350	1.329	1.6	65	-0.02
31	3,3'-Dichlorobenzidine	0.349	0.414	-18.6	78	-0.02
32	Chrysene	1.563	1.245	20.3	53	-0.02
33	Bis(2-ethylhexyl)phthalate	0.684	0.714	-4.4	66	-0.02
34	Indeno(1,2,3-cd)pyrene	1.163	0.917	21.2	64	-0.03
35 I	Perylene-d12	1.000	1.000	0.0	59	-0.02
36	Benzo(b)fluoranthene	1.522	1.515	0.5	62	-0.02
37	Benzo(k)fluoranthene	1.366	1.351	1.1	61	-0.02
38 C	Benzo(a)pyrene	1.249	1.249	0.0	64	-0.02
39	Dibenzo(a,h)anthracene	1.136	0.981	13.6	64	-0.03
40	Benzo(a,h,i)perylene	1.199	1.002	16.4	64	-0.03

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121815\
Data File : BM003610.D
Acq On : 18 Dec 2015 13:06
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 18 16:16:58 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 16:19:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0