

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121815\
 Data File : BM003610.D
 Acq On : 18 Dec 2015 13:06
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 18 16:16:58 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 02 16:19:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	54	-0.02
2	1,4-Dioxane	1.000	0.910	9.0	55	-0.03
3	n-Nitrosodimethylamine	1.000	1.013	-1.3	57	-0.03
4 S	2-Fluorophenol	1.000	0.976	2.4	59	-0.02
5 S	Phenol-d6	1.000	0.951	4.9	54	0.00
6	bis(2-Chloroethyl)ether	1.000	0.917	8.3	52	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	52	-0.02
8 S	Nitrobenzene-d5	1.000	1.022	-2.2	57	-0.02
9	Nitrobenzene	1.000	1.014	-1.4	55	-0.02
10	Naphthalene	1.000	0.949	5.1	53	-0.02
11	Hexachlorobutadiene	1.000	1.004	-0.4	56	-0.02
12	2-Methylnaphthalene	1.000	0.979	2.1	53	-0.02
13 I	Acenaphthene-d10	5.000	5.000	0.0	51	-0.02
14 S	2,4,6-Tribromophenol	1.000	1.105	-10.5	58	0.00
15 S	2-Fluorobiphenyl	1.000	0.939	6.1	53	-0.02
16	Acenaphthylene	1.000	1.037	-3.7	55	-0.02
17	Acenaphthene	1.000	0.965	3.5	55	-0.02
18	Fluorene	1.000	0.985	1.5	53	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	42	-0.03
20	4-Bromophenyl-phenylether	1.000	1.515	-51.5#	74	-0.03
21	Hexachlorobenzene	1.000	1.290	-29.0#	59	0.00
22	Pentachlorophenol	1.000	1.017	-1.7	45	-0.03
23	Phenanthrene	1.000	1.460	-46.0#	69	-0.03
24	Anthracene	1.000	1.007	-0.7	43	-0.03
25	Fluoranthene	1.000	1.398	-39.8#	58	-0.02
26 I	Chrysene-d12	5.000	5.000	0.0	63	-0.02
27	Benzydine	1.000	0.974	2.6	69	0.00
28	Pvrene	1.000	0.821	17.9	56	-0.02
29 S	Terphenyl-d14	1.000	0.859	14.1	57	-0.02
30	Benzo(a)anthracene	1.000	0.985	1.5	65	-0.02
31	3,3'-Dichlorobenzidine	1.000	1.104	-10.4	78	-0.02
32	Chrysene	1.000	0.796	20.4	53	-0.02
33	Bis(2-ethylhexyl)phthalate	1.000	1.029	-2.9	66	-0.02
34	Indeno(1,2,3-cd)pyrene	1.000	0.788	21.2	64	-0.03
35 I	Perylene-d12	5.000	5.000	0.0	59	-0.02
36	Benzo(b)fluoranthene	1.000	0.995	0.5	62	-0.02
37	Benzo(k)fluoranthene	1.000	0.989	1.1	61	-0.02
38 C	Benzo(a)pyrene	1.000	1.000	0.0	64	-0.02
39	Dibenzo(a,h)anthracene	1.000	0.863	13.7	64	-0.03
40	Benzo(a,h,i)perylene	1.000	0.835	16.5	64	-0.03

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121815\
Data File : BM003610.D
Acq On : 18 Dec 2015 13:06
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 18 16:16:58 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM120215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 02 16:19:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0