

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121816\
 Data File : BM008434.D
 Acq On : 18 Dec 2016 12:46
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02065

Quant Time: Dec 19 01:23:14 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 01:18:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.499	0.535	-7.2	130	0.00
3 S	1,4-Dioxane-d8	0.426	0.446	-4.7	126	0.00
4	Benzaldehyde	1.039	1.197	-15.2	127	0.00
5 S	Phenol-d5	1.565	1.613	-3.1	128	0.00
6	Phenol	1.623	1.676	-3.3	129	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.901	0.937	-4.0	126	0.00
8	Bis(2-Chloroethyl)ether	1.212	1.270	-4.8	129	0.00
9 S	2-Chlorophenol-d4	1.174	1.261	-7.4	131	0.00
10	2-Chlorophenol	1.188	1.300	-9.4	132	0.00
11	2-Methylphenol	1.186	1.225	-3.3	129	0.00
12	2,2'-oxybis(1-Chloropropane	1.387	1.464	-5.6	127	0.00
13 S	4-Methylphenol-d8	1.207	1.253	-3.8	129	0.00
14	Acetophenone	1.978	1.987	-0.5	125	0.00
15 P	N-Nitroso-di-n-propylamine	1.013	1.060	-4.6	126	0.00
16	4-Methylphenol	1.281	1.317	-2.8	130	0.00
17	Hexachloroethane	0.542	0.585	-7.9	130	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
19 S	Nitrobenzene-d5	0.144	0.153	-6.3	128	0.00
20	Nitrobenzene	0.370	0.404	-9.2	131	0.00
21	Isophorone	0.671	0.716	-6.7	127	0.00
22 S	2-Nitrophenol-d4	0.159	0.171	-7.5	131	0.00
23 C	2-Nitrophenol	0.163	0.177	-8.6	131	0.00
24	2,4-Dimethylphenol	0.367	0.395	-7.6	128	0.00
25	Bis(2-Chloroethoxy)methane	0.383	0.409	-6.8	124	0.00
26 S	2,4-Dichlorophenol-d3	0.293	0.324	-10.6	129	0.00
27 C	2,4-Dichlorophenol	0.295	0.319	-8.1	127	0.00
28	Naphthalene	0.947	1.019	-7.6	127	0.00
29 S	4-Chloroaniline-d4	0.335	0.366	-9.3	124	0.00
30	4-Chloroaniline	0.340	0.385	-13.2	129	0.00
31 C	Hexachlorobutadiene	0.240	0.259	-7.9	131	0.00
32	Caprolactam	0.092	0.094	-2.2	126	0.00
33 C	4-Chloro-3-methylphenol	0.319	0.345	-8.2	127	0.00
34	2-Methylnaphthalene	0.684	0.736	-7.6	128	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
36	1,2,4,5-Tetrachlorobenzene	0.627	0.689	-9.9	128	0.00
37	Hexachlorocyclopentadiene	0.250	0.330	-32.0#	205#	0.00
38 C	2,4,6-Trichlorophenol	0.356	0.395	-11.0	130	0.00
39	2,4,5-Trichlorophenol	0.385	0.407	-5.7	126	0.00
40	1,1'-Biphenyl	1.336	1.415	-5.9	123	0.00
41	2-Chloronaphthalene	1.025	1.118	-9.1	127	0.00
42	2-Nitroaniline	0.302	0.338	-11.9	130	0.00
43 S	Dimethylphthalate-d6	1.332	1.422	-6.8	123	0.00
44	Dimethylphthalate	1.305	1.398	-7.1	123	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121816\
 Data File : BM008434.D
 Acq On : 18 Dec 2016 12:46
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02065

Quant Time: Dec 19 01:23:14 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 01:18:57 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.252	0.285	-13.1	131	0.00
46 S	Acenaphthylene-d8	1.571	1.727	-9.9	127	0.00
47	Acenaphthylene	1.606	1.725	-7.4	123	0.00
48	3-Nitroaniline	0.253	0.274	-8.3	129	0.00
49 C	Acenaphthene	1.106	1.197	-8.2	125	0.00
50	2,4-Dinitrophenol	0.148	0.145	2.0	141	0.00
51 S	4-Nitrophenol-d4	0.196	0.155	20.9#	104	0.00
52	4-Nitrophenol	0.189	0.147	22.2#	101	0.00
53	Dibenzofuran	1.607	1.730	-7.7	124	0.00
54	2,4-Dinitrotoluene	0.379	0.415	-9.5	125	0.00
55	2,3,4,6-Tetrachlorophenol	0.358	0.384	-7.3	124	0.00
56	Diethylphthalate	1.298	1.379	-6.2	123	0.00
57 S	Fluorene-d10	1.159	1.235	-6.6	123	0.00
58	Fluorene	1.316	1.426	-8.4	126	0.00
59	4-Chlorophenyl-phenylether	0.703	0.759	-8.0	124	0.00
60	4-Nitroaniline	0.244	0.255	-4.5	127	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.111	6.7	117	0.00
63	4,6-Dinitro-2-methylphenol	0.122	0.113	7.4	115	0.00
64	N-Nitrosodiphenylamine	0.543	0.590	-8.7	123	0.00
65	4-Bromophenyl-phenylether	0.219	0.240	-9.6	123	0.00
66	Hexachlorobenzene	0.249	0.266	-6.8	121	0.00
67	Atrazine	0.222	0.233	-5.0	124	0.00
68 C	Pentachlorophenol	0.134	0.126	6.0	117	0.00
69	Phenanthrene	1.027	1.105	-7.6	122	0.00
70 S	Anthracene-d10	0.871	0.959	-10.1	125	0.00
71	Anthracene	1.041	1.137	-9.2	122	0.00
72	Carbazole	0.914	0.983	-7.5	125	0.00
73	Di-n-butylphthalate	1.022	1.112	-8.8	123	0.00
74 C	Fluoranthene	1.258	1.342	-6.7	122	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
76 S	Pyrene-d10	0.772	0.839	-8.7	123	0.00
77	Pyrene	0.981	1.060	-8.1	123	0.00
78	Butylbenzylphthalate	0.351	0.387	-10.3	126	0.00
79	3,3'-Dichlorobenzidine	0.350	0.373	-6.6	119	0.00
80	Benzo(a)anthracene	1.030	1.121	-8.8	124	0.00
81	Bis(2-ethylhexyl)phthalate	0.513	0.569	-10.9	125	0.00
82	Chrysene	0.996	1.077	-8.1	121	0.00
83 I	Perylene-d12	1.000	1.000	0.0	124	0.00
84	Di-n-octyl phthalate	0.957	1.016	-6.2	127	0.00
85	Benzo(b)fluoranthene	1.076	1.161	-7.9	125	0.00
86	Benzo(k)fluoranthene	1.031	1.129	-9.5	121	0.00
87 S	Benzo(a)pyrene-d12	0.834	0.916	-9.8	126	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121816\
Data File : BM008434.D
Acq On : 18 Dec 2016 12:46
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02065

Quant Time: Dec 19 01:23:14 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-2016\SOM02.2-EPA-BM121416.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 19 01:18:57 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.042	1.135	-8.9	122	0.00
89	Indeno(1,2,3-cd)pyrene	1.262	1.376	-9.0	124	0.00
90	Dibenzo(a,h)anthracene	1.059	1.142	-7.8	123	0.00
91	Benzo(a,h,i)perylene	1.052	1.141	-8.5	124	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0