


Data Path : Z:\HPCHEM1\BNA_M\Data\BM121817\
 Data File : BM013489.D
 Acq On : 18 Dec 2017 08:51
 Operator : SJ/JU
 Sample : DFTPP65
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP65

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
 Title : SVOA CALIBRATION
 Last Update : Wed Dec 20 03:10:37 2017



AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.0	104570	PASS
68	69	0.00	2	1.8	2025	PASS
69	198	0.00	100	42.1	110016	PASS
70	69	0.00	2	0.5	499	PASS
127	198	10	80	46.7	121925	PASS
197	198	0.00	2	0.2	614	PASS
198	198	100	100	100.0	261354	PASS
199	198	5	9	6.7	17574	PASS
275	198	10	60	28.3	73930	PASS
365	198	1	100	3.8	9918	PASS
441	443	0.01	100	73.4	29355	PASS
442	198	50	100	81.2	212290	PASS
443	442	15	24	18.8	40000	PASS

m/z	Abundance
37.00	732.0
38.00	1459.0
39.10	7216.0
40.00	940.0
43.90	1646.0
45.10	446.0
49.10	649.0
50.00	26560.0
51.10	89456.0
52.10	4843.0
55.10	650.0
55.90	3181.0
57.00	5436.0
60.90	990.0
62.00	1959.0
63.10	2949.0
64.00	681.0
65.00	1304.0
68.00	1880.0
69.00	91296.0
70.00	440.0
73.00	1550.0
74.00	13076.0
75.10	13834.0
76.00	5387.0
77.00	102288.0
78.00	7238.0
79.00	8377.0
80.10	5046.0
81.00	8933.0
81.90	2470.0
83.00	2228.0
83.90	724.0
84.90	1760.0
86.00	2860.0
86.90	1061.0
88.00	838.0
91.00	1699.0
92.10	1875.0
93.00	15090.0
94.00	1193.0
96.10	694.0
98.00	10547.0
99.00	8417.0
99.90	1061.0
101.00	5258.0
103.20	1144.0
104.10	2849.0
105.00	1905.0
106.10	1118.0
107.00	28480.0
107.90	3745.0
109.00	1047.0
110.00	51936.0
111.00	9112.0
112.10	711.0
113.10	692.0
116.10	1589.0
117.00	28928.0
118.00	2805.0
118.90	391.0
120.00	670.0
121.90	1737.0
123.00	4182.0
124.00	1205.0
125.00	1597.0
126.20	633.0
127.00	96600.0
128.00	7628.0
129.00	38464.0
130.00	3791.0
131.00	662.0
131.90	482.0
134.00	1675.0
135.00	2528.0
135.90	1452.0
137.00	1873.0
137.90	501.0
140.00	740.0
141.00	5548.0
141.90	1831.0
143.00	1158.0
146.00	868.0
147.00	3534.0
148.00	6959.0
148.90	1082.0
151.00	409.0
151.40	479.0
152.10	710.0
153.10	2015.0
153.90	1199.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

155.10	3599.0
156.00	5033.0
157.00	1181.0
157.80	1598.0
159.10	564.0
160.10	1999.0
160.90	3089.0
162.00	528.0
164.00	308.0
165.00	2990.0
166.00	1463.0
167.00	10120.0
168.10	6320.0
169.00	1082.0
171.10	533.0
171.80	1848.0
173.00	1626.0
174.00	2317.0
175.00	3935.0
176.00	881.0
177.00	3267.0
178.10	925.0
178.90	6421.0
180.00	5557.0
180.90	2222.0
181.70	689.0
184.00	859.0
185.00	4510.0
186.00	29560.0
187.10	8452.0
188.00	954.0
188.90	2108.0
191.00	1134.0
191.90	2782.0
193.10	3301.0
194.00	869.0
195.90	4196.0
198.00	187072.0
199.00	11796.0
199.80	883.0
201.00	541.0
201.70	825.0
202.90	2545.0
204.00	6919.0
205.10	12511.0
206.00	51224.0
207.00	9542.0
208.00	3421.0
209.00	1421.0
210.00	1383.0
211.00	2199.0
212.00	478.0
212.90	301.0
214.90	519.0
216.00	1117.0
216.90	14145.0
218.00	1226.0
220.30	598.0
221.00	8085.0
222.00	1209.0
223.10	2613.0
224.00	28024.0
225.00	7120.0
226.00	898.0
227.00	13808.0
227.90	2708.0
229.00	3620.0
229.90	330.0
231.00	1006.0
234.00	690.0
234.80	1320.0
235.90	986.0
237.00	1310.0
239.00	500.0
239.90	587.0
240.90	997.0
242.00	1549.0
243.00	1591.0
244.00	20152.0
245.00	3406.0
246.00	4560.0
247.00	1154.0
247.80	340.0
248.90	787.0
251.90	322.0
252.70	382.0
252.90	392.0
254.10	471.0
255.00	116200.0
256.00	17616.0
256.90	1676.0
258.00	6886.0
259.00	945.0
260.90	448.0
265.00	2695.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

265.90	721.0
270.80	910.0
272.10	489.0
273.00	3253.0
274.00	8352.0
275.00	51144.0
276.00	6042.0
277.00	5333.0
277.90	832.0
278.80	337.0
281.00	700.0
282.90	892.0
284.00	441.0
285.20	906.0
289.00	474.0
289.90	546.0
292.80	864.0
294.00	391.0
296.10	14763.0
296.90	2020.0
298.00	412.0
303.00	1457.0
304.00	463.0
313.90	1063.0
314.90	2031.0
316.00	880.0
321.00	414.0
322.00	387.0
323.10	4837.0
324.00	1325.0
326.70	709.0
327.90	376.0
334.00	2441.0
335.10	667.0
340.90	539.0
346.00	1098.0
351.90	1522.0
353.10	715.0
354.00	947.0
354.90	466.0
365.00	7329.0
365.90	1236.0
372.00	1934.0
373.00	591.0
383.10	733.0
383.90	466.0
389.90	445.0
391.00	302.0
401.90	1210.0
403.00	915.0
404.00	565.0
421.00	951.0
422.00	964.0
423.00	5911.0
424.00	2157.0
441.10	14155.0
442.10	110856.0
443.10	20464.0
444.10	1869.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

m/z	Abundance
37.10	795.0
38.10	1722.0
39.00	9678.0
40.00	808.0
43.10	345.0
43.90	1736.0
45.10	338.0
49.10	1458.0
50.10	38552.0
51.10	124920.0
52.00	6744.0
55.00	1050.0
56.00	3595.0
57.00	8610.0
61.00	1867.0
62.10	2097.0
63.10	5098.0
64.00	530.0
65.20	2253.0
68.10	2504.0
69.00	132352.0
69.90	663.0
73.00	1976.0
74.00	16157.0
75.00	24448.0
76.00	6765.0
77.10	140992.0
78.10	9179.0
79.00	12036.0
80.00	8978.0
81.00	11829.0
82.00	2581.0
82.90	2238.0
85.00	2199.0
86.00	3058.0
87.00	1630.0
88.10	753.0
91.00	3049.0
92.00	3017.0
93.00	20632.0
94.00	1731.0
95.90	1116.0
98.00	15629.0
99.00	12592.0
100.00	998.0
101.00	7313.0
102.00	574.0
103.00	1691.0
104.10	3866.0
105.00	3918.0
106.00	1318.0
107.00	47464.0
108.00	7028.0
109.00	1049.0
110.00	77976.0
111.00	13200.0
111.90	969.0
113.00	623.0
115.10	619.0
116.10	2463.0
117.00	47496.0
118.00	2825.0
118.90	467.0
119.90	681.0
121.90	2901.0
123.00	4945.0
123.80	2051.0
124.90	1681.0
126.20	820.0
127.00	143232.0
128.00	10886.0
129.00	61632.0
130.00	5579.0
130.90	720.0
131.80	316.0
132.20	498.0
132.80	446.0
134.10	2042.0
135.10	4527.0
136.00	2053.0
137.00	2613.0
138.00	883.0
139.90	938.0
141.00	7760.0
142.00	3858.0
142.90	2122.0
144.10	529.0
145.10	518.0
146.00	1944.0
147.00	4629.0
148.00	11161.0

Instrument :

BNA_M

ClientSampleId :

DFTPP65

149.10	2439.0
150.00	644.0
151.10	831.0
151.90	707.0
152.10	694.0
152.90	3001.0
154.00	2137.0
155.00	4345.0
156.00	8642.0
157.00	1569.0
157.80	1101.0
159.00	1434.0
160.00	2437.0
161.00	3355.0
161.90	1017.0
163.10	404.0
165.00	4148.0
166.10	2496.0
167.00	18224.0
168.00	8415.0
168.90	1506.0
169.90	707.0
170.90	465.0
171.90	1642.0
173.00	2162.0
174.00	4555.0
175.10	6227.0
176.00	1282.0
177.00	3677.0
178.00	833.0
179.00	14409.0
180.00	9841.0
181.00	3953.0
182.20	854.0
183.00	316.0
184.00	913.0
185.00	4935.0
186.00	44592.0
187.00	11473.0
188.00	1730.0
189.00	4044.0
189.90	665.0
191.00	1739.0
192.00	3835.0
193.00	4990.0
194.10	695.0
194.90	996.0
196.00	7413.0
196.80	1738.0
197.90	313280.0
199.00	22848.0
199.90	1532.0
201.40	1351.0
202.20	387.0
202.90	3009.0
204.00	12429.0
205.00	21184.0
206.00	94664.0
207.00	15119.0
208.00	3670.0
208.90	1554.0
210.00	1421.0
211.10	3068.0
211.60	1557.0
212.40	311.0
215.00	2171.0
215.90	2727.0
217.00	25384.0
218.20	2900.0
221.00	14527.0
221.90	4857.0
223.10	5186.0
224.00	49928.0
225.10	14951.0
226.10	1207.0
227.00	21608.0
228.00	3814.0
229.10	3322.0
229.90	662.0
231.10	2318.0
233.90	1508.0
235.00	1843.0
236.10	554.0
237.00	1711.0
239.10	1107.0
240.00	765.0
240.90	1795.0
242.10	2892.0
243.00	3120.0
244.10	37744.0
245.10	3890.0
246.00	7720.0
246.90	1449.0
248.10	505.0
248.90	1971.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

250.10	439.0
252.00	850.0
253.00	778.0
254.00	1379.0
255.00	204928.0
256.00	32336.0
257.00	2825.0
258.00	14371.0
259.00	2441.0
265.00	5496.0
265.90	694.0
269.90	563.0
271.00	604.0
271.90	442.0
273.00	7621.0
274.00	14739.0
275.00	87016.0
276.00	13164.0
277.00	10293.0
278.00	1383.0
278.90	516.0
280.90	831.0
282.20	387.0
283.00	1618.0
284.00	823.0
285.10	1574.0
289.00	369.0
290.00	391.0
293.00	1598.0
294.10	339.0
295.10	472.0
296.00	32040.0
296.90	4338.0
301.90	672.0
303.10	2739.0
304.00	353.0
307.80	320.0
309.00	381.0
310.10	338.0
312.90	542.0
313.90	1665.0
315.10	3351.0
316.00	2451.0
317.00	317.0
320.80	493.0
321.20	776.0
322.10	418.0
323.00	8135.0
324.00	1653.0
327.00	1658.0
327.90	1274.0
332.00	594.0
332.80	780.0
334.10	5126.0
335.00	1443.0
339.20	372.0
340.90	908.0
345.90	2070.0
352.10	2408.0
352.90	1159.0
354.10	1809.0
354.90	840.0
365.00	10590.0
366.00	1896.0
370.90	561.0
372.10	3696.0
373.10	501.0
383.00	1382.0
384.00	468.0
389.90	459.0
391.00	446.0
401.00	361.0
402.00	1310.0
403.10	2836.0
404.00	1136.0
421.10	1670.0
422.00	1679.0
423.10	11814.0
424.10	4210.0
425.30	352.0
441.10	34672.0
442.10	237632.0
443.10	44664.0
444.00	4780.0
444.90	377.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

m/z	Abundance
38.00	1662.0
39.00	7494.0
40.00	554.0
44.00	2309.0
49.10	849.0
50.00	28544.0
51.10	99336.0
52.00	4017.0
52.90	323.0
55.10	653.0
56.00	2916.0
57.00	6534.0
58.10	558.0
61.00	1779.0
62.00	1221.0
63.00	4707.0
64.00	602.0
65.20	2257.0
68.10	1693.0
69.00	106400.0
70.10	395.0
72.90	1307.0
74.00	12094.0
75.00	16254.0
76.10	6763.0
77.10	117064.0
78.10	7517.0
79.10	8397.0
80.00	6144.0
81.10	9766.0
82.10	2484.0
82.90	1668.0
85.10	1470.0
85.90	2948.0
87.00	1487.0
88.10	688.0
88.90	452.0
91.10	2624.0
91.90	2689.0
93.00	16119.0
93.90	1351.0
94.90	536.0
95.90	502.0
96.80	435.0
97.90	12868.0
99.00	11293.0
100.10	881.0
101.00	5366.0
102.10	472.0
102.90	1906.0
103.90	3642.0
105.00	2418.0
105.80	766.0
106.20	557.0
107.00	40480.0
108.00	5410.0
108.90	720.0
110.00	65728.0
111.00	9458.0
112.10	1943.0
113.00	515.0
115.00	313.0
116.20	1920.0
117.00	42936.0
118.00	2887.0
119.00	578.0
120.10	395.0
121.90	2870.0
123.00	2511.0
124.10	2194.0
124.90	1509.0
126.20	799.0
127.00	125944.0
128.00	9177.0
129.00	52792.0
130.00	4267.0
131.00	542.0
132.10	643.0
132.70	710.0
133.90	1617.0
135.00	3803.0
136.00	1806.0
137.00	2246.0
138.00	585.0
139.80	717.0
140.90	8438.0
142.00	2352.0
143.00	1274.0
144.10	357.0
145.10	318.0
145.90	1812.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

147.00	4463.0
148.00	10051.0
149.00	2118.0
150.20	673.0
151.10	991.0
151.60	719.0
153.00	2952.0
154.00	2011.0
155.00	2933.0
156.00	5419.0
156.90	976.0
158.00	888.0
159.00	989.0
159.90	1797.0
161.10	4253.0
161.90	1609.0
162.80	591.0
165.00	3454.0
166.00	2957.0
167.10	15721.0
168.00	7332.0
169.00	1166.0
169.80	475.0
170.80	817.0
172.00	1413.0
173.10	1244.0
174.00	3328.0
175.00	5113.0
176.00	1492.0
176.90	2330.0
177.90	1207.0
179.00	10329.0
180.00	7670.0
181.00	3705.0
181.90	641.0
182.30	365.0
182.90	793.0
184.00	733.0
185.00	5031.0
186.00	38696.0
187.00	13640.0
188.10	1839.0
189.10	3036.0
190.20	485.0
191.00	1751.0
192.00	5025.0
193.00	6097.0
194.10	1158.0
196.00	9619.0
197.00	1844.0
198.00	283712.0
198.90	18080.0
199.90	1404.0
201.60	1291.0
202.20	747.0
203.00	2599.0
204.00	10954.0
205.00	21328.0
206.00	78672.0
207.00	13787.0
208.00	3402.0
208.90	1841.0
209.90	1317.0
211.10	4113.0
211.80	733.0
215.00	1788.0
216.00	2335.0
217.00	24072.0
218.00	2768.0
218.90	356.0
221.00	9339.0
221.70	2946.0
223.00	5463.0
224.00	47560.0
225.00	11855.0
226.20	1339.0
227.00	22472.0
228.00	3362.0
228.90	4024.0
231.00	1514.0
233.90	1768.0
234.90	1862.0
236.00	1345.0
237.00	1923.0
239.10	835.0
239.90	820.0
241.00	1379.0
242.10	3677.0
243.00	2578.0
244.00	33688.0
245.00	4398.0
246.00	8856.0
246.90	1368.0
248.00	323.0
249.00	1433.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

252.10	529.0
252.90	996.0
254.10	1446.0
255.00	188352.0
256.00	29608.0
257.00	2256.0
258.00	13791.0
259.00	2424.0
260.00	513.0
261.20	377.0
265.00	4477.0
265.90	747.0
270.00	344.0
271.00	554.0
271.90	444.0
273.00	6125.0
274.00	16496.0
275.00	83632.0
276.10	12095.0
277.00	7159.0
278.00	1622.0
281.10	313.0
281.80	407.0
283.10	1231.0
283.80	424.0
284.80	836.0
285.90	455.0
293.00	2144.0
293.90	430.0
295.00	828.0
296.00	29736.0
297.00	4135.0
297.90	457.0
300.90	595.0
301.90	643.0
303.00	3491.0
304.00	1471.0
307.80	676.0
308.90	410.0
310.20	423.0
311.80	301.0
312.90	380.0
314.10	1594.0
315.00	3148.0
316.00	1392.0
317.10	462.0
321.10	664.0
321.80	845.0
323.10	9298.0
324.00	929.0
326.90	1823.0
328.00	1023.0
331.90	616.0
332.90	826.0
334.00	5696.0
335.10	1265.0
341.00	1261.0
342.10	347.0
346.10	2014.0
347.00	447.0
352.10	3155.0
353.10	1900.0
354.10	3211.0
355.20	576.0
365.00	11835.0
366.00	1652.0
370.10	462.0
371.10	712.0
372.00	4754.0
372.90	1057.0
383.00	1606.0
384.00	322.0
390.00	376.0
391.90	344.0
402.10	2067.0
403.10	1854.0
404.10	971.0
421.10	2527.0
422.00	1767.0
423.10	14693.0
424.00	3824.0
424.90	540.0
441.10	39240.0
442.10	288384.0
443.10	54872.0
444.00	5630.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65

Scan 2446 (17.474 min): BM013489.D
DFTPP65

m/z	Abundance
39.90	810.0
44.00	1538.0
73.00	1019.0
133.10	516.0
207.00	2095.0
207.90	380.0
209.00	414.0

Instrument :
BNA_M
ClientSampleId :
DFTPP65