

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121915\
 Data File : BM003638.D
 Acq On : 20 Dec 2015 02:04
 Operator : UM/SJ
 Sample : G4801-32
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 G9C80

Quant Time: Dec 20 02:54:00 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Dec 20 01:33:56 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	34828	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	170946	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.35	164	109590	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.10	188	265545	20.00	ng/ul	0.00
78) Chrysene-d12	21.30	240	341654	20.00	ng/ul	0.00
86) Perylene-d12	23.54	264	351893	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.16	96	3001	4.64	ng/uL	0.00
5) Phenol-d5	6.87	99	70130	23.89	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	40564	25.94	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	60325	25.18	ng/ul	0.00
13) 4-Methylphenol-d8	8.40	113	63378	24.93	ng/ul	0.00
19) Nitrobenzene-d5	8.86	128	31561	25.45	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	35306	25.30	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.11	165	61250	24.05	ng/ul	0.00
29) 4-Chloroaniline-d4	10.63	131	80175	23.64	ng/ul	0.00
44) Dimethylphthalate-d6	13.76	166	237364	26.21	ng/ul	0.00
47) Acenaphthylene-d8	14.04	160	272851	26.46	ng/ul	0.00
52) 4-Nitrophenol-d4	14.56	143	41052	23.27	ng/ul	0.00
58) Fluorene-d10	15.35	176	200212	26.72	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.47	200	28952	19.34	ng/ul	0.00
71) Anthracene-d10	17.20	188	329656	27.28	ng/ul	0.00
79) Pyrene-d10	19.50	212	399316	26.76	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.40	264	483080	27.64	ng/ul	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
45) Dimethylphthalate	13.80	163	102277	11.04	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM121915\
Data File : BM003638.D
Acq On : 20 Dec 2015 02:04
Operator : UM/SJ
Sample : G4801-32
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
G9C80

Quant Time: Dec 20 02:54:00 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM121915.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Dec 20 01:33:56 2015
Response via : Initial Calibration





#45

Dimethylphthalate

Concen: 11.04 ng/ul

RT: 13.80 min Scan# 1822

Delta R.T. -0.01 min

Lab File: BM003638.D

Acq: 20 Dec 2015 02:04

Instrument :

BNA_M

ClientSampleId :

G9C80

Tgt Ion:163 Resp: 102277

Ion Ratio Lower Upper

163 100

194 3.7 3.2 4.8

164 10.2 7.8 11.8

