

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121915\
 Data File : BM003621.D
 Acq On : 19 Dec 2015 15:57
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Dec 20 01:31:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM121915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Dec 20 01:27:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.436	0.453	-3.9	92	0.00
3 S	1,4-Dioxane-d8	0.371	0.369	0.5	90	0.00
4	Benzaldehyde	0.972	1.093	-12.4	90	0.00
5 S	Phenol-d5	1.685	1.720	-2.1	89	0.00
6	Phenol	1.775	1.820	-2.5	89	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.898	0.934	-4.0	90	0.00
8	Bis(2-Chloroethyl)ether	1.312	1.367	-4.2	89	0.00
9 S	2-Chlorophenol-d4	1.376	1.424	-3.5	90	0.00
10	2-Chlorophenol	1.430	1.504	-5.2	90	0.00
11	2-Methylphenol	1.405	1.451	-3.3	90	0.00
12	2,2'-oxybis(1-Chloropropane	1.324	1.386	-4.7	89	0.00
13 S	4-Methylphenol-d8	1.460	1.505	-3.1	90	0.00
14	Acetophenone	2.280	2.441	-7.1	91	0.00
15 P	N-Nitroso-di-n-propylamine	1.105	1.181	-6.9	89	0.00
16	4-Methylphenol	1.567	1.629	-4.0	90	0.00
17	Hexachloroethane	0.544	0.563	-3.5	90	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
19 S	Nitrobenzene-d5	0.145	0.151	-4.1	91	0.00
20	Nitrobenzene	0.335	0.347	-3.6	89	0.00
21	Isophorone	0.668	0.696	-4.2	89	0.00
22 S	2-Nitrophenol-d4	0.163	0.167	-2.5	89	0.00
23 C	2-Nitrophenol	0.181	0.188	-3.9	90	0.00
24	2,4-Dimethylphenol	0.371	0.391	-5.4	90	0.00
25	Bis(2-Chloroethoxy)methane	0.403	0.415	-3.0	90	0.00
26 S	2,4-Dichlorophenol-d3	0.298	0.310	-4.0	90	0.00
27 C	2,4-Dichlorophenol	0.310	0.325	-4.8	91	0.00
28	Naphthalene	1.030	1.069	-3.8	90	0.00
29 S	4-Chloroaniline-d4	0.397	0.433	-9.1	90	0.00
30	4-Chloroaniline	0.401	0.439	-9.5	91	0.00
31 C	Hexachlorobutadiene	0.191	0.194	-1.6	90	0.00
32	Caprolactam	0.116	0.119	-2.6	90	0.00
33 C	4-Chloro-3-methylphenol	0.363	0.381	-5.0	89	0.00
34	2-Methylnaphthalene	0.770	0.802	-4.2	90	0.00
35	1-Methylnaphthalene	0.731	0.766	-4.8	89	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
37	1,2,4,5-Tetrachlorobenzene	0.605	0.614	-1.5	89	0.00
38	Hexachlorocyclopentadiene	0.324	0.309	4.6	93	0.00
39 C	2,4,6-Trichlorophenol	0.409	0.411	-0.5	88	0.00
40	2,4,5-Trichlorophenol	0.452	0.459	-1.5	90	0.00
41	1,1'-Biphenyl	1.624	1.665	-2.5	90	0.00
42	2-Chloronaphthalene	1.220	1.249	-2.4	90	0.00
43	2-Nitroaniline	0.338	0.351	-3.8	90	0.00
44 S	Dimethylphthalate-d6	1.653	1.714	-3.7	91	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121915\
 Data File : BM003621.D
 Acq On : 19 Dec 2015 15:57
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02053

Quant Time: Dec 20 01:31:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM121915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sun Dec 20 01:27:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.691	1.761	-4.1	91	0.00
46	2,6-Dinitrotoluene	0.329	0.344	-4.6	90	0.00
47 S	Acenaphthylene-d8	1.882	1.955	-3.9	91	0.00
48	Acenaphthylene	2.058	2.140	-4.0	90	0.00
49	3-Nitroaniline	0.368	0.398	-8.2	92	0.00
50 C	Acenaphthene	1.374	1.423	-3.6	90	0.00
51	2,4-Dinitrophenol	0.175	0.165	5.7	100	0.00
52 S	4-Nitrophenol-d4	0.322	0.338	-5.0	93	0.00
53	4-Nitrophenol	0.272	0.285	-4.8	92	0.00
54	Dibenzofuran	1.962	2.041	-4.0	91	0.00
55	2,4-Dinitrotoluene	0.500	0.542	-8.4	93	0.00
56	2,3,4,6-Tetrachlorophenol	0.389	0.403	-3.6	90	0.00
57	Diethylphthalate	1.789	1.861	-4.0	91	0.00
58 S	Fluorene-d10	1.368	1.436	-5.0	92	0.00
59	Fluorene	1.574	1.672	-6.2	91	0.00
60	4-Chlorophenyl-phenylether	0.764	0.803	-5.1	92	0.00
61	4-Nitroaniline	0.408	0.446	-9.3	94	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.113	0.115	-1.8	95	0.00
64	4,6-Dinitro-2-methylphenol	0.121	0.126	-4.1	97	0.00
65	N-Nitrosodiphenylamine	0.585	0.603	-3.1	90	0.00
66	4-Bromophenyl-phenylether	0.199	0.201	-1.0	91	0.00
67	Hexachlorobenzene	0.221	0.231	-4.5	93	0.00
68	Atrazine	0.225	0.238	-5.8	92	0.00
69 C	Pentachlorophenol	0.125	0.123	1.6	91	0.00
70	Phenanthrene	1.106	1.152	-4.2	92	0.00
71 S	Anthracene-d10	0.910	0.947	-4.1	92	0.00
72	Anthracene	1.128	1.181	-4.7	92	0.00
73	1,2,3,4-Tetrachlorobenzene	0.274	0.274	0.0	90	0.00
74	Pentachlorobenzene	0.258	0.262	-1.6	91	0.00
75	Carbazole	1.030	1.131	-9.8	93	0.00
76	Di-n-butylphthalate	1.340	1.372	-2.4	91	0.00
77 C	Fluoranthene	1.272	1.418	-11.5	94	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
79 S	Pyrene-d10	0.874	0.867	0.8	93	0.00
80	Pyrene	1.151	1.139	1.0	93	0.00
81	Butylbenzylphthalate	0.544	0.539	0.9	92	0.00
82	3,3'-Dichlorobenzidine	0.382	0.418	-9.4	95	0.00
83	Benzo(a)anthracene	1.206	1.235	-2.4	95	0.00
84	Bis(2-ethylhexyl)phthalate	0.810	0.831	-2.6	95	0.00
85	Chrysene	1.163	1.188	-2.1	95	0.00
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Di-n-octyl phthalate	1.382	1.392	-0.7	95	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121915\
Data File : BM003621.D
Acq On : 19 Dec 2015 15:57
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02053

Quant Time: Dec 20 01:31:43 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM121915.M
Quant Title : SVOA CALIBRATION
QLast Update : Sun Dec 20 01:27:01 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.206	1.209	-0.2	96	0.00
89	Benzo(k)fluoranthene	1.142	1.161	-1.7	97	0.00
90 S	Benzo(a)pyrene-d12	0.994	1.028	-3.4	98	0.00
91 C	Benzo(a)pyrene	1.153	1.191	-3.3	97	0.01
92	Indeno(1,2,3-cd)pyrene	1.241	1.399	-12.7	99	0.01
93	Dibenzo(a,h)anthracene	1.041	1.182	-13.5	100	0.01
94	Benzo(a,h,i)perylene	1.020	1.168	-14.5	100	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0