

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121916\
 Data File : BM008449.D
 Acq On : 19 Dec 2016 19:49
 Operator : UM/SJ
 Sample : H5938-17
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA839

Manual Integrations
 APPROVED

sohil
 12/20/2016 6:31:08 PM

Quant Time: Dec 20 04:01:29 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 20 03:40:00 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	335	0.40	ng/ul	-0.05
2) Naphthalene-d8	10.43	136	1161	0.40	ng/ul	-0.06
6) Acenaphthene-d10	14.30	164	755	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.06	188	1643m	0.40	ng/ul	-0.03
16) Chrysene-d12	21.27	240	1546	0.40	ng/ul	-0.01
20) Perylene-d12	23.49	264	1426m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.04	152	623	0.35	ng/ul	-0.06
14) Fluoranthene-d10	19.10	212	1528	0.34	ng/ul	-0.02
Target Compounds						Ovalue
3) Naphthalene	10.48	128	736	0.23	ng/ul#	80
5) 2-Methylnaphthalene	12.11	142	488	0.22	ng/ul	96
8) Acenaphthene	14.35	153	155	0.06	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121916\
Data File : BM008449.D
Acq On : 19 Dec 2016 19:49
Operator : UM/SJ
Sample : H5938-17
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA839

Manual Integrations
APPROVED

sohil
12/20/2016 6:31:08 PM

Quant Time: Dec 20 04:01:29 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 20 03:40:00 2016
Response via : Initial Calibration

