

Method Path : Z:\HPCHEM1\BNA M\METHODS\
 Method File : SOM-EPA-SIM-BM121916.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 20 04:48:59 2016
 Response Via : Initial Calibration

Calibration Files

0.1 =BM008440.D 0.2 =BM008441.D 0.4 =BM008460.D
 0.8 =BM008443.D 1.6 =BM008444.D 3.2 =BM008445.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	Avg	%RSD

1) I	1,4-Dichlorobenzene-d	-----ISTD-----							
2) I	Naphthalene-d8	-----ISTD-----							
3)	Naphthalene	1.226	1.145	1.099	1.090	1.067		1.125	5.58
4)	SURR2-Methylnaphthale	0.683	0.634	0.583	0.591	0.576		0.613	7.34
5)	2-Methylnaphthale	0.824	0.767	0.730	0.740	0.727		0.757	5.30
6) I	Acenaphthene-d10	-----ISTD-----							
7)	Acenaphthylene	1.873	1.818	1.744	1.854	1.759		1.809	3.15
8) C	Acenaphthene	1.381	1.365	1.332	1.385	1.341		1.361	1.73
9)	Fluorene	1.612	1.547	1.531	1.575	1.502		1.553	2.70
10) I	Phenanthrene-d10	-----ISTD-----							
11)	Pentachlorophenol		0.117	0.135	0.114	0.119	0.137	0.124	8.70
12)	Phenanthrene	1.227	1.223	1.209	1.227	1.221		1.221	0.61
13)	Anthracene	1.203	1.191	1.137	1.196	1.204		1.186	2.35
14)	SURRFluoranthene-d10	1.152	1.091	1.094	1.089	1.080		1.101	2.62
15) C	Fluoranthene	1.570	1.510	1.466	1.475	1.466		1.498	2.97
16) I	Chrysene-d12	-----ISTD-----							
17)	Pyrene	1.616	1.535	1.521	1.547	1.512		1.546	2.66
18)	Benzo(a)anthracen	1.474	1.239	1.247	1.339	1.268		1.313	7.47
19)	Chrysene	1.525	1.484	1.460	1.349	1.372		1.438	5.20
20) I	Perylene-d12	-----ISTD-----							
21)	Benzo(b)fluoranth	1.779	1.579	1.544	1.436	1.542		1.576	7.97
22)	Benzo(k)fluoranth	1.514	1.629	1.548	1.553	1.439		1.537	4.49
23) C	Benzo(a)pyrene	1.544	1.470	1.447	1.421	1.420		1.461	3.50
24)	Indeno(1,2,3-cd)p	1.779	1.744	1.768	1.689	1.724		1.741	2.05
25)	Dibenzo(a,h)anthr	1.422	1.398	1.414	1.344	1.376		1.391	2.29
26)	Benzo(a,h,i)peryl	1.485	1.515	1.527	1.454	1.483		1.493	1.93

(#) = Out of Range