

m/z	Abundance
39.00	664.0
40.00	2006.0
41.10	1208.0
42.10	471.0
43.00	906.0
44.00	1623.0
49.00	835.0
51.10	380.0
55.10	1117.0
56.00	364.0
57.00	745.0
66.10	306.0
68.10	324.0
69.00	599.0
70.10	688.0
73.10	1054.0
77.10	322.0
82.00	432.0
83.20	343.0
84.00	1208.0
85.10	426.0
86.00	391.0
91.10	385.0
95.00	373.0
96.00	518.0
97.00	416.0
98.00	413.0
133.00	385.0
149.00	457.0
152.10	318.0
153.00	331.0
207.00	1305.0
208.20	326.0
280.90	539.0
282.00	303.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
37.00	495.0
38.00	1352.0
39.10	6102.0
40.00	2351.0
41.10	860.0
42.10	410.0
43.00	706.0
44.00	1209.0
45.00	315.0
49.00	1451.0
50.10	23664.0
51.10	90168.0
52.10	4387.0
53.00	364.0
54.10	402.0
54.90	1150.0
56.10	2717.0
57.00	6779.0
58.10	492.0
61.00	1284.0
62.10	1379.0
63.10	3665.0
64.10	856.0
65.00	1863.0
66.00	486.0
67.00	437.0
68.10	2234.0
69.00	99480.0
70.10	855.0
71.00	466.0
73.10	1529.0
74.10	10741.0
75.10	17648.0
76.10	5764.0
77.10	111288.0
78.10	7259.0
79.00	9115.0
80.00	5700.0
81.00	8448.0
82.00	2153.0
83.00	2411.0
84.00	1462.0
85.00	1963.0
86.00	3542.0
87.00	1531.0
87.90	439.0
89.10	501.0
91.00	2514.0
92.10	2091.0
93.00	14733.0
94.00	1345.0
95.00	547.0
96.00	1061.0
96.90	342.0
97.20	367.0
98.00	11300.0
99.00	8317.0
100.00	725.0
101.00	4511.0
102.10	369.0
103.00	1488.0
104.00	3288.0
105.00	3615.0
106.00	864.0
107.00	36816.0
108.00	5240.0
109.20	1299.0
110.00	63848.0
111.10	9521.0
112.00	1088.0
113.00	522.0
115.10	399.0
116.10	2158.0
117.10	33464.0
118.00	2698.0
119.00	453.0
120.10	510.0
122.00	2517.0
123.00	3796.0
124.00	1515.0
125.10	1449.0
127.00	109760.0
128.00	9457.0
129.00	46744.0
130.00	4522.0
131.00	910.0
132.20	432.0
134.00	1309.0
135.00	4303.0
136.00	1672.0
137.00	1797.0
137.70	430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

138.10	401.0
139.00	388.0
139.90	660.0
141.00	6778.0
142.00	2600.0
143.00	1609.0
144.10	426.0
146.00	1515.0
147.00	3577.0
148.00	8451.0
149.00	1856.0
150.00	359.0
151.00	951.0
153.10	1934.0
154.00	1752.0
155.00	3544.0
156.10	5715.0
157.00	935.0
158.00	1296.0
159.00	931.0
160.00	2076.0
161.00	3242.0
162.10	868.0
163.00	324.0
163.90	399.0
165.00	2655.0
166.10	1791.0
167.00	12922.0
168.00	6671.0
169.00	1136.0
170.00	704.0
170.90	482.0
172.00	1279.0
173.00	1507.0
174.00	2796.0
175.00	5356.0
176.00	1180.0
177.00	2503.0
178.00	1063.0
179.00	9892.0
180.00	6956.0
181.10	3157.0
182.00	429.0
183.00	484.0
184.00	829.0
185.00	4747.0
186.10	34424.0
187.00	11101.0
188.10	1193.0
189.00	2443.0
190.10	630.0
191.10	1665.0
192.00	3451.0
193.00	3374.0
194.00	705.0
195.00	453.0
196.00	7160.0
197.10	1605.0
198.00	244672.0
199.00	17568.0
200.00	1464.0
200.90	484.0
201.60	815.0
203.00	2128.0
204.10	10499.0
205.00	16896.0
206.10	68128.0
207.10	10807.0
208.00	3023.0
209.00	1334.0
210.10	1246.0
211.00	2521.0
213.00	353.0
215.00	1317.0
216.00	1327.0
217.00	19176.0
218.00	2936.0
221.00	9177.0
221.90	2884.0
223.00	4192.0
224.10	38312.0
225.00	9854.0
226.10	1047.0
227.00	17864.0
228.00	2752.0
229.00	3547.0
229.90	558.0
231.00	1519.0
233.20	374.0
234.00	1070.0
235.10	899.0
236.00	821.0
237.00	1370.0
239.00	692.0
240.00	735.0
240.90	1193.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

242.00	1887.0
243.10	2101.0
244.10	28056.0
245.10	4133.0
246.00	6166.0
246.90	1576.0
249.00	1037.0
253.00	535.0
254.00	1180.0
255.00	153472.0
256.00	23112.0
257.00	1829.0
258.00	10359.0
259.00	1214.0
264.10	320.0
265.00	3908.0
266.00	587.0
267.00	306.0
270.90	660.0
272.00	434.0
273.00	4664.0
274.00	11618.0
275.00	66344.0
276.00	8822.0
277.00	5946.0
278.00	1560.0
281.00	629.0
282.10	421.0
283.00	709.0
284.00	350.0
285.10	1302.0
293.10	1557.0
295.00	454.0
296.00	21208.0
297.00	3051.0
302.00	332.0
303.00	2264.0
304.00	802.0
308.00	347.0
314.00	930.0
315.10	2387.0
316.00	1079.0
321.00	598.0
323.10	6205.0
324.00	1017.0
327.00	1562.0
328.10	807.0
331.90	363.0
333.10	597.0
334.10	3991.0
335.00	964.0
341.00	711.0
342.00	308.0
346.00	1493.0
352.00	1697.0
352.90	1137.0
354.00	2233.0
355.20	494.0
365.00	8846.0
366.00	1439.0
371.00	396.0
372.10	2458.0
373.20	689.0
382.90	840.0
390.00	306.0
402.00	1482.0
403.00	1527.0
404.10	625.0
421.00	1452.0
422.10	959.0
423.10	9155.0
424.00	2406.0
441.10	23896.0
442.10	163072.0
443.10	33528.0
444.00	3146.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
37.10	572.0
38.10	1366.0
39.10	8555.0
40.00	2145.0
41.10	1208.0
42.00	700.0
43.10	938.0
44.00	1204.0
45.10	424.0
49.10	1600.0
50.10	30880.0
51.10	115784.0
52.10	6543.0
53.10	488.0
54.00	401.0
55.10	1551.0
56.00	3576.0
57.00	7945.0
58.10	605.0
59.10	330.0
61.10	1643.0
62.10	2053.0
63.10	5460.0
64.00	937.0
65.10	2564.0
66.00	387.0
67.10	480.0
68.10	2911.0
69.00	135488.0
70.10	943.0
72.10	466.0
73.00	1296.0
74.10	14320.0
75.00	22032.0
76.10	6858.0
77.10	155200.0
78.10	9884.0
79.10	11783.0
80.00	8207.0
81.00	12162.0
82.10	2921.0
83.00	2780.0
84.00	1690.0
85.00	2967.0
86.00	3950.0
87.10	1722.0
88.00	624.0
89.10	519.0
91.00	3563.0
92.00	3201.0
93.00	20248.0
94.00	1584.0
95.00	685.0
96.00	1082.0
97.10	482.0
98.00	15027.0
99.00	11446.0
100.00	1275.0
101.00	6346.0
102.00	399.0
103.00	2590.0
104.00	3985.0
105.00	4134.0
106.10	1424.0
107.00	52232.0
108.00	8238.0
109.10	1239.0
110.00	87200.0
111.10	13546.0
112.00	1686.0
113.00	795.0
115.00	406.0
116.00	3186.0
117.00	48688.0
118.10	3169.0
119.10	548.0
119.90	663.0
121.00	555.0
122.00	2911.0
123.00	4547.0
124.00	2391.0
125.10	2270.0
126.20	619.0
127.00	160768.0
128.00	12577.0
129.00	66296.0
130.00	5650.0
131.00	1314.0
132.10	676.0
133.20	491.0
134.00	2144.0
135.00	6321.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

136.00	2141.0
137.10	3072.0
138.00	855.0
139.00	641.0
140.10	894.0
141.00	9374.0
142.10	2712.0
143.00	2111.0
144.00	855.0
145.10	610.0
146.00	1710.0
147.00	5016.0
148.00	12460.0
149.00	2277.0
150.00	577.0
151.10	1227.0
152.10	825.0
153.00	3367.0
154.00	2217.0
155.10	5186.0
156.00	7502.0
157.10	1714.0
157.90	1624.0
159.00	1271.0
160.00	3642.0
161.10	4154.0
162.00	1422.0
163.10	640.0
164.10	512.0
165.00	3648.0
166.00	3297.0
167.00	20672.0
168.00	8986.0
169.00	1588.0
170.00	610.0
171.00	1034.0
172.00	1634.0
173.00	2231.0
174.00	4365.0
175.10	7685.0
176.10	2374.0
177.00	4029.0
178.10	1704.0
179.00	15561.0
180.00	10804.0
181.10	4755.0
182.10	662.0
183.10	470.0
183.90	1659.0
185.00	8082.0
186.00	56224.0
187.10	17528.0
188.00	1452.0
189.00	3966.0
190.00	705.0
191.00	2274.0
192.00	4951.0
193.10	5486.0
194.10	1129.0
195.10	698.0
196.00	11618.0
197.10	2465.0
198.00	389632.0
199.00	24488.0
200.10	1873.0
201.50	1354.0
203.00	3104.0
204.00	15738.0
205.10	25840.0
206.10	110208.0
207.10	15544.0
208.00	4344.0
209.00	1465.0
210.20	1908.0
211.10	3826.0
212.00	842.0
213.10	379.0
214.90	1489.0
216.00	2146.0
217.00	30544.0
218.10	3222.0
219.00	681.0
221.00	13722.0
221.90	3622.0
223.00	6104.0
224.10	62000.0
225.10	17344.0
226.00	1820.0
227.00	26480.0
228.10	3809.0
229.00	6423.0
230.00	720.0
231.00	2194.0
232.00	598.0
233.00	615.0
234.00	1797.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

235.00	1987.0
236.00	1268.0
237.00	2154.0
238.00	382.0
239.00	988.0
240.00	610.0
240.90	1754.0
242.10	3448.0
243.10	3040.0
244.10	44192.0
245.00	6619.0
246.00	10553.0
247.00	2109.0
248.10	416.0
249.00	1658.0
250.00	420.0
251.10	420.0
252.10	493.0
253.00	894.0
254.20	1588.0
255.00	257280.0
256.00	37880.0
257.00	3090.0
258.00	16050.0
259.00	2592.0
259.90	755.0
261.00	637.0
264.00	454.0
265.00	6164.0
266.10	1063.0
267.00	310.0
270.00	358.0
271.10	672.0
272.00	787.0
273.00	8067.0
274.00	19792.0
275.00	116048.0
276.10	15011.0
277.00	10921.0
278.00	1897.0
281.00	607.0
283.00	1664.0
284.00	686.0
285.00	1884.0
286.00	461.0
289.00	567.0
289.90	464.0
292.00	580.0
293.00	2461.0
294.00	538.0
295.00	786.0
296.00	37048.0
297.00	5064.0
298.00	381.0
301.00	507.0
302.10	580.0
303.00	4258.0
304.00	1186.0
308.00	758.0
310.10	332.0
313.00	726.0
314.00	2152.0
315.00	4654.0
316.00	1965.0
317.10	463.0
321.00	923.0
322.00	395.0
323.10	10619.0
324.00	1608.0
327.00	2286.0
328.00	1352.0
332.00	760.0
333.00	1360.0
334.10	6606.0
335.10	1611.0
336.10	352.0
341.00	1119.0
346.00	2408.0
347.00	491.0
351.10	336.0
352.10	3579.0
353.10	2472.0
354.10	3620.0
355.10	623.0
365.00	16720.0
366.00	2641.0
371.10	931.0
372.10	5622.0
373.00	1358.0
383.10	1196.0
384.00	371.0
390.10	890.0
390.90	405.0
392.10	467.0
400.90	325.0
402.10	1768.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

403.10	3383.0
404.10	1292.0
421.00	2295.0
422.00	2508.0
423.10	16720.0
424.00	3769.0
425.00	462.0
441.10	50296.0
442.10	326784.0
443.10	63832.0
444.10	6419.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
37.00	342.0
38.10	1098.0
39.10	6504.0
40.00	1987.0
41.10	989.0
42.00	459.0
43.10	745.0
44.10	1313.0
45.10	332.0
49.00	1302.0
50.10	21864.0
51.10	90408.0
52.10	4300.0
53.10	367.0
53.90	396.0
55.00	1084.0
56.00	3640.0
57.00	6261.0
58.00	356.0
61.00	1225.0
62.10	1471.0
63.10	3943.0
64.10	764.0
65.10	2018.0
66.00	354.0
67.00	499.0
68.10	2374.0
69.00	101088.0
70.20	797.0
71.00	328.0
73.20	1378.0
74.10	12784.0
75.00	18448.0
76.10	6599.0
77.10	118472.0
78.10	7497.0
79.00	9873.0
80.00	6682.0
81.00	9979.0
82.00	2764.0
83.00	2003.0
84.00	1518.0
85.00	1895.0
86.10	3520.0
87.00	1489.0
88.00	512.0
89.00	472.0
91.10	2798.0
92.00	2401.0
93.00	16448.0
94.00	1325.0
95.10	407.0
95.90	923.0
97.00	705.0
98.00	13194.0
99.00	8831.0
100.00	664.0
101.00	5465.0
102.00	330.0
103.10	1832.0
104.00	3801.0
105.00	3534.0
106.20	1116.0
107.00	41624.0
108.10	5697.0
109.10	1126.0
110.00	69056.0
111.00	11342.0
112.10	1418.0
112.90	729.0
115.00	344.0
116.00	2667.0
117.00	36520.0
118.00	2976.0
119.10	372.0
120.00	654.0
121.10	422.0
122.00	2978.0
123.00	4784.0
124.00	2438.0
125.00	1997.0
126.20	962.0
127.00	134336.0
128.10	9859.0
129.00	55416.0
130.00	5136.0
131.10	1232.0
132.00	426.0
134.00	1729.0
135.00	4621.0
136.10	2315.0
137.00	2608.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

138.00	619.0
138.90	519.0
140.10	947.0
141.00	8485.0
142.10	2333.0
143.00	1779.0
144.00	577.0
145.00	414.0
146.00	1352.0
147.10	4130.0
148.00	9863.0
149.00	1729.0
150.00	475.0
151.00	1217.0
151.60	947.0
153.00	2582.0
154.10	1876.0
155.10	4555.0
156.10	6661.0
157.10	1820.0
158.00	1501.0
159.00	1200.0
160.00	3062.0
161.10	4229.0
162.00	1188.0
162.90	451.0
164.00	484.0
165.00	3940.0
166.10	2627.0
167.00	16282.0
168.10	7270.0
169.00	1239.0
169.90	645.0
171.00	715.0
172.00	2011.0
173.00	1915.0
174.00	3741.0
175.00	6549.0
176.10	1681.0
177.00	3455.0
178.00	1094.0
179.00	13582.0
180.00	8354.0
181.00	4168.0
182.00	648.0
183.00	495.0
184.00	1078.0
185.00	6538.0
186.00	50448.0
187.00	13025.0
188.00	1490.0
189.00	2883.0
190.10	491.0
191.00	1711.0
192.00	4489.0
193.00	5120.0
194.10	1112.0
195.10	860.0
196.00	8855.0
197.10	2170.0
198.00	343680.0
199.00	22512.0
200.00	2195.0
201.60	1239.0
203.00	2551.0
204.00	12622.0
205.10	22984.0
206.10	97312.0
207.10	13414.0
208.10	3826.0
209.00	1783.0
210.10	1288.0
211.00	3473.0
211.90	999.0
213.00	416.0
215.00	1019.0
216.00	2567.0
217.00	26376.0
218.00	3667.0
219.00	347.0
220.00	477.0
221.00	13073.0
222.00	3422.0
223.00	6300.0
224.10	54696.0
225.00	14976.0
226.00	1594.0
227.00	26672.0
228.00	3488.0
229.00	5068.0
230.00	672.0
231.00	2325.0
233.00	623.0
234.00	1675.0
235.10	1961.0
236.00	1029.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

237.00	1821.0
238.00	359.0
239.00	1398.0
240.00	929.0
241.00	1240.0
242.00	3114.0
243.00	3293.0
244.10	45712.0
245.10	5648.0
246.00	10272.0
247.00	1757.0
248.90	1846.0
250.00	446.0
250.90	481.0
252.00	558.0
253.10	1293.0
254.00	1421.0
255.00	243200.0
256.00	34512.0
257.00	2883.0
258.00	15622.0
259.10	2344.0
260.10	349.0
261.00	472.0
264.10	367.0
265.00	5852.0
266.00	978.0
270.00	450.0
271.10	597.0
272.00	587.0
273.00	7306.0
274.00	20672.0
275.00	112248.0
276.00	14772.0
277.00	10466.0
278.00	1953.0
279.00	457.0
280.90	490.0
282.00	526.0
282.90	1128.0
284.00	1015.0
285.00	1817.0
286.10	318.0
288.90	384.0
290.00	505.0
292.00	421.0
293.00	2292.0
294.00	626.0
295.00	655.0
296.00	35912.0
297.00	4308.0
297.90	359.0
301.00	415.0
302.00	459.0
303.10	4559.0
304.10	826.0
307.90	478.0
309.10	393.0
310.10	318.0
312.90	381.0
314.00	1918.0
315.00	4065.0
316.00	2270.0
317.10	364.0
321.10	815.0
322.10	588.0
323.10	12097.0
324.10	1974.0
327.00	2718.0
328.10	1204.0
331.90	720.0
333.00	1078.0
334.10	6654.0
335.00	1796.0
341.10	1495.0
342.00	604.0
346.00	2457.0
346.90	429.0
352.10	2985.0
353.10	1662.0
354.00	3628.0
355.00	870.0
365.00	17832.0
366.00	2699.0
370.00	374.0
371.10	807.0
372.10	5957.0
373.00	1637.0
383.00	1266.0
384.00	398.0
389.90	661.0
391.00	562.0
391.90	463.0
401.10	327.0
402.00	2133.0
403.00	3156.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

404.00	1070.0
421.00	3092.0
422.00	2754.0
423.10	19976.0
424.10	3669.0
425.00	484.0
440.00	352.0
441.10	57480.0
442.10	374016.0
443.10	69920.0
444.10	6515.0
445.00	356.0


Instrument :
BNA_M
ClientSampleId :
DFTPP41

Data Path : Z:\HPCHEM1\BNA M\DATA\BM121916\
 Data File : BM008439.D
 Acq On : 19 Dec 2016 10:11
 Operator : UM/SJ
 Sample : DFTPP41
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP41

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-TEST.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 30 00:32:18 2016



AutoFind: Scans 2531, 2532, 2533; Background Corrected with Scan 2526

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.2	98406	PASS
68	69	0.00	2	2.0	2182	PASS
69	198	0.00	100	34.2	111419	PASS
70	69	0.00	2	0.2	177	PASS
127	198	10	80	41.4	134954	PASS
197	198	0.00	2	0.6	2080	PASS
198	198	100	100	100.0	325994	PASS
199	198	5	9	6.6	21522	PASS
275	198	10	60	30.1	98213	PASS
365	198	1	100	4.4	14466	PASS
441	443	0.01	100	78.7	43890	PASS
442	198	50	100	88.3	287957	PASS
443	442	15	24	19.4	55760	PASS