

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121923\
 Data File : BM043468.D
 Acq On : 20 Dec 2023 17:41
 Operator : MA/JU
 Sample : PB157763BS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS763

Quant Time: Dec 20 23:14:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 19 23:32:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.983	152	5407	0.400	ng/ul	0.00
4) Naphthalene-d8	10.790	136	16504	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.610	164	8619	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.358	188	17986	0.400	ng/ul	0.00
17) Chrysene-d12	21.536	240	12150	0.400	ng/ul	# 0.00
23) Perylene-d12	23.950	264	13732	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.385	96	5075	0.743	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.374	152	7820	0.334	ng/ul	0.00
18) Fluoranthene-d10	19.380	212	13743	0.317	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.418	88	14422	2.105	ng/ul#	87
5) Naphthalene	10.839	128	15609	0.315	ng/ul	99
7) 2-Methylnaphthalene	12.451	142	9669	0.302	ng/ul	99
8) 1-Methylnaphthalene	12.665	142	9519	0.296	ng/ul	99
10) Acenaphthylene	14.338	152	14850	0.313	ng/ul	99
11) Acenaphthene	14.675	153	10938	0.317	ng/ul	99
12) Fluorene	15.665	166	12035	0.308	ng/ul	99
14) Pentachlorophenol	17.003	266	2830	0.710	ng/ul	98
15) Phenanthrene	17.400	178	19639	0.327	ng/ul	99
16) Anthracene	17.493	178	17803	0.327	ng/ul	99
19) Fluoranthene	19.413	202	20336	0.297	ng/ul	96
20) Pyrene	19.770	202	21053	0.309	ng/ul	94
21) Benzo(a)anthracene	21.518	228	18039	0.338	ng/ul	99
22) Chrysene	21.571	228	18910	0.336	ng/ul	100
24) Benzo(b)fluoranthene	23.213	252	19880	0.319	ng/ul	93
25) Benzo(k)fluoranthene	23.263	252	21062	0.326	ng/ul#	92
26) Benzo(a)pyrene	23.845	252	17930	0.349	ng/ul#	92
27) Indeno(1,2,3-cd)pyrene	26.468	276	24775	0.423	ng/ul#	92
28) Dibenzo(a,h)anthracene	26.495	278	19982	0.431	ng/ul	97
29) Benzo(g,h,i)perylene	27.233	276	21058	0.421	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121923\
Data File : BM043468.D
Acq On : 20 Dec 2023 17:41
Operator : MA/JU
Sample : PB157763BS
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS763

Quant Time: Dec 20 23:14:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 23:32:46 2023
Response via : Initial Calibration

