

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121923\
 Data File : BM043480.D
 Acq On : 21 Dec 2023 00:53
 Operator : MA/JU
 Sample : PB157925BS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS925

Quant Time: Dec 21 01:24:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 19 23:32:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.979	152	5426	0.400	ng/ul	0.00
4) Naphthalene-d8	10.785	136	16312	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.611	164	8009	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.358	188	16304	0.400	ng/ul	0.00
17) Chrysene-d12	21.533	240	10998	0.400	ng/ul	# 0.00
23) Perylene-d12	23.947	264	12516	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	4562	0.665	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.374	152	7051	0.305	ng/ul	0.00
18) Fluoranthene-d10	19.380	212	11922	0.303	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.414	88	14582	2.121	ng/ul#	81
5) Naphthalene	10.834	128	15377	0.314	ng/ul	99
7) 2-Methylnaphthalene	12.445	142	9244	0.292	ng/ul	100
8) 1-Methylnaphthalene	12.665	142	8983	0.283	ng/ul	100
10) Acenaphthylene	14.333	152	13568	0.308	ng/ul	99
11) Acenaphthene	14.675	153	9896	0.308	ng/ul	99
12) Fluorene	15.661	166	10821	0.298	ng/ul	100
14) Pentachlorophenol	17.004	266	2535	0.702	ng/ul	98
15) Phenanthrene	17.401	178	17182	0.316	ng/ul	99
16) Anthracene	17.493	178	15542	0.315	ng/ul	99
19) Fluoranthene	19.408	202	18535	0.299	ng/ul#	90
20) Pyrene	19.771	202	18855	0.306	ng/ul	96
21) Benzo(a)anthracene	21.515	228	15835	0.328	ng/ul	99
22) Chrysene	21.571	228	16679	0.327	ng/ul	100
24) Benzo(b)fluoranthene	23.211	252	17277	0.304	ng/ul	93
25) Benzo(k)fluoranthene	23.260	252	17885	0.304	ng/ul#	90
26) Benzo(a)pyrene	23.839	252	16345	0.349	ng/ul#	90
27) Indeno(1,2,3-cd)pyrene	26.461	276	21406	0.401	ng/ul#	93
28) Dibenzo(a,h)anthracene	26.492	278	17248	0.408	ng/ul	97
29) Benzo(g,h,i)perylene	27.229	276	18333	0.402	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121923\
Data File : BM043480.D
Acq On : 21 Dec 2023 00:53
Operator : MA/JU
Sample : PB157925BS
Misc :
ALS Vial : 58 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS925

Quant Time: Dec 21 01:24:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 23:32:46 2023
Response via : Initial Calibration

