

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121924\
 Data File : BM049074.D
 Acq On : 19 Dec 2024 17:18
 Operator : RC/JU
 Sample : P5226-13
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 E2AQ5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/21/2024
 Supervised By :mohammad ahmed 12/26/2024

Quant Time: Dec 19 22:20:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 18 15:33:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.565	152	6482	0.400	ng/u1	0.00
4) Naphthalene-d8	10.326	136	18977	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.197	164	10211	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.947	188	21737m	0.400	ng/u1	0.00
17) Chrysene-d12	21.149	240	16306m	0.400	ng/u1	0.00
23) Perylene-d12	23.306	264	15789m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	25864	3.460	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.926	152	5112	0.200	ng/u1	0.00
18) Fluoranthene-d10	18.984	212	10173	0.220	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM121924\
Data File : BM049074.D
Acq On : 19 Dec 2024 17:18
Operator : RC/JU
Sample : P5226-13
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
E2AQ5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/21/2024
Supervised By :mohammad ahmed 12/26/2024

Quant Time: Dec 19 22:20:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 18 15:33:13 2024
Response via : Initial Calibration

