

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122016\
 Data File : BM008490.D
 Acq On : 20 Dec 2016 21:42
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.451

Quant Time: Dec 21 12:51:37 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 02:43:00 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
3	Naphthalene	1.125	1.110	1.3	124	0.00
4 SURR	2-Methylnaphthalene-d10	0.613	0.572	6.7	121	0.00
5	2-Methylnaphthalene	0.757	0.705	6.9	119	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
7	Acenaphthylene	1.809	1.890	-4.5	125	0.00
8 C	Acenaphthene	1.361	1.373	-0.9	119	0.00
9	Fluorene	1.553	1.506	3.0	113	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
11	Pentachlorophenol	0.124	0.139	-12.1	113	0.00
12	Phenanthrene	1.221	1.175	3.8	107	0.00
13	Anthracene	1.186	1.174	1.0	114	0.00
14 SURR	Fluoranthene-d10	1.101	1.058	3.9	107	0.00
15 C	Fluoranthene	1.498	1.534	-2.4	115	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
17	Pyrene	1.546	1.686	-9.1	111	0.00
18	Benzo(a)anthracene	1.313	1.326	-1.0	106	0.00
19	Chrysene	1.438	1.440	-0.1	99	0.00
20 I	Perylene-d12	1.000	1.000	0.0	106	0.00
21	Benzo(b)fluoranthene	1.576	1.366	13.3	93	0.00
22	Benzo(k)fluoranthene	1.537	1.597	-3.9	109	0.00
23 C	Benzo(a)pyrene	1.461	1.469	-0.5	107	0.00
24	Indeno(1,2,3-cd)pyrene	1.741	1.801	-3.4	108	0.00
25	Dibenzo(a,h)anthracene	1.391	1.430	-2.8	107	0.00
26	Benzo(a,h,i)perylene	1.493	1.595	-6.8	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0