


Data Path : Z:\HPCHEM1\BNA_M\Data\BM122017\
 Data File : BM013522.D
 Acq On : 20 Dec 2017 09:02
 Operator : SJ/JU
 Sample : DFTPP67
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP67

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM113017.M
 Title : SVOA CALIBRATION
 Last Update : Mon Dec 25 01:37:20 2017



AutoFind: Scans 2451, 2452, 2453; Background Corrected with Scan 2446

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.3	103042	PASS
68	69	0.00	2	1.1	1147	PASS
69	198	0.00	100	40.5	103546	PASS
70	69	0.00	2	0.6	607	PASS
127	198	10	80	45.9	117378	PASS
197	198	0.00	2	0.5	1222	PASS
198	198	100	100	100.0	255765	PASS
199	198	5	9	7.0	18014	PASS
275	198	10	60	29.4	75296	PASS
365	198	1	100	4.1	10608	PASS
441	443	0.01	100	75.2	30429	PASS
442	198	50	100	81.1	207402	PASS
443	442	15	24	19.5	40445	PASS

m/z	Abundance
36.90	413.0
38.10	1414.0
39.00	8277.0
39.90	414.0
44.00	1922.0
45.00	324.0
48.80	627.0
50.10	28584.0
51.10	104160.0
52.10	5444.0
53.10	341.0
54.90	1013.0
56.10	3265.0
57.10	5350.0
61.00	1608.0
62.00	2369.0
63.10	3466.0
64.00	532.0
65.00	1841.0
68.00	1495.0
69.00	101816.0
70.10	344.0
73.10	1241.0
74.10	11477.0
75.00	17624.0
76.10	5835.0
77.10	114840.0
78.10	7685.0
79.00	8796.0
80.00	7493.0
81.00	7205.0
82.00	2840.0
83.00	2323.0
83.90	320.0
84.90	1887.0
86.00	2979.0
87.00	1669.0
88.00	611.0
91.00	2364.0
92.00	2737.0
93.00	15867.0
93.90	1131.0
96.00	442.0
97.10	382.0
98.00	12519.0
99.00	8290.0
100.10	747.0
101.00	4160.0
102.00	379.0
103.00	1067.0
104.00	4114.0
105.00	3454.0
106.00	642.0
107.00	37552.0
108.00	4997.0
108.90	710.0
110.00	54968.0
111.00	8671.0
111.90	938.0
113.00	476.0
115.00	325.0
116.00	2650.0
117.00	41592.0
118.00	2863.0
120.10	368.0
122.10	3054.0
123.00	3879.0
123.90	1576.0
125.10	988.0
125.90	810.0
127.00	108864.0
128.00	8667.0
129.00	42944.0
130.00	3788.0
130.90	1092.0
133.90	1715.0
135.10	3240.0
136.00	1587.0
136.90	2045.0
139.80	861.0
140.90	7485.0
142.00	2523.0
142.90	1581.0
144.90	360.0
146.00	1557.0
146.90	3657.0
148.00	9408.0
148.90	1947.0
149.90	374.0
151.20	712.0
151.80	428.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

152.10	423.0
152.90	1928.0
154.00	1595.0
155.00	3988.0
156.00	5890.0
156.90	644.0
157.30	697.0
157.90	1932.0
159.00	548.0
160.00	2202.0
160.90	2950.0
161.90	1656.0
164.90	1736.0
166.00	2416.0
166.90	13331.0
168.00	8667.0
169.10	993.0
170.00	510.0
170.90	594.0
171.80	1301.0
173.00	1869.0
174.10	3101.0
175.10	4172.0
175.90	1430.0
177.10	1996.0
177.90	1082.0
178.90	9785.0
180.00	5750.0
181.00	2940.0
181.80	554.0
183.10	504.0
183.90	909.0
185.00	3862.0
186.10	33408.0
187.00	9242.0
188.00	1356.0
188.90	3786.0
190.20	635.0
191.10	762.0
192.00	3094.0
193.00	2518.0
193.90	847.0
195.00	581.0
196.00	4234.0
197.00	924.0
197.90	216896.0
199.00	14764.0
200.00	1207.0
200.80	771.0
201.60	1153.0
202.90	2303.0
204.00	10125.0
205.00	14803.0
206.00	64672.0
207.10	9398.0
208.00	2303.0
208.90	1200.0
210.20	749.0
210.90	2449.0
211.70	788.0
213.10	441.0
214.90	1217.0
216.10	1852.0
217.00	19344.0
218.00	2479.0
221.00	7821.0
222.10	2679.0
223.00	4150.0
224.00	35784.0
225.00	10027.0
226.00	1390.0
227.00	17376.0
228.00	2413.0
229.00	4048.0
229.90	640.0
231.10	1302.0
231.90	405.0
234.00	1132.0
235.00	1665.0
236.00	653.0
236.90	1118.0
239.20	763.0
239.80	633.0
241.00	652.0
242.00	1666.0
243.10	1399.0
244.10	27624.0
245.00	4049.0
246.00	6116.0
247.00	1423.0
249.20	884.0
250.90	443.0
253.00	886.0
254.00	819.0
255.00	139712.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

256.00	20648.0
257.00	1734.0
258.00	8712.0
259.00	1439.0
265.00	3420.0
266.00	836.0
270.90	733.0
273.00	4539.0
274.00	10728.0
275.00	60280.0
276.00	7706.0
277.10	5139.0
278.00	871.0
279.10	315.0
281.00	330.0
282.80	370.0
283.20	397.0
284.00	595.0
284.90	1053.0
293.10	1192.0
294.80	395.0
296.00	20064.0
297.10	3096.0
301.90	507.0
303.10	1925.0
303.80	743.0
308.00	491.0
309.00	311.0
314.20	1040.0
315.00	2331.0
315.90	1413.0
321.20	499.0
323.10	4826.0
324.00	1079.0
327.10	811.0
328.00	629.0
333.10	463.0
333.90	3064.0
335.10	861.0
341.00	595.0
346.10	1185.0
352.10	1331.0
353.00	672.0
354.20	1326.0
354.90	678.0
365.00	7877.0
366.00	651.0
371.90	2534.0
373.10	497.0
383.10	696.0
402.10	627.0
403.10	1585.0
404.10	480.0
421.10	1236.0
422.10	1069.0
423.00	7147.0
424.00	1506.0
440.10	368.0
441.10	21232.0
442.10	135616.0
443.10	27400.0
444.10	2055.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

m/z	Abundance
37.10	499.0
38.00	2182.0
39.00	9486.0
39.90	665.0
43.90	2120.0
49.10	1244.0
50.00	33112.0
51.10	123432.0
52.00	6197.0
52.90	339.0
55.00	745.0
56.00	3267.0
57.00	7102.0
61.10	1552.0
62.00	1874.0
63.10	4955.0
64.00	488.0
65.00	2337.0
67.90	1946.0
69.00	119736.0
70.00	891.0
72.90	1362.0
74.00	14508.0
75.00	21384.0
76.00	6960.0
77.10	129920.0
78.00	9150.0
79.00	11514.0
80.00	7752.0
81.00	10267.0
82.00	2332.0
83.00	2706.0
84.00	521.0
85.00	1876.0
86.00	3660.0
87.00	1706.0
88.20	438.0
88.80	383.0
91.10	2588.0
92.10	2474.0
93.00	20992.0
93.90	1050.0
94.80	470.0
96.20	1083.0
98.00	15970.0
99.00	11884.0
100.00	684.0
101.00	5253.0
102.10	457.0
102.90	1447.0
104.00	3212.0
105.00	3501.0
106.00	1182.0
107.00	45568.0
108.00	7087.0
109.10	2048.0
110.00	70064.0
111.00	11036.0
111.90	1086.0
113.00	732.0
116.00	2423.0
117.00	51464.0
117.90	3201.0
118.90	603.0
120.10	423.0
121.10	471.0
122.10	2691.0
123.00	4345.0
124.00	1849.0
124.90	2219.0
126.10	996.0
127.00	132800.0
128.00	12217.0
129.00	57232.0
130.00	5257.0
131.00	989.0
132.70	344.0
134.00	1735.0
135.10	4235.0
136.00	1683.0
137.10	2899.0
138.00	656.0
139.20	392.0
139.90	1401.0
141.00	10065.0
142.00	2593.0
143.10	1935.0
143.80	341.0
145.00	737.0
145.80	918.0
146.90	4371.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

148.00	10858.0
149.00	1969.0
150.10	467.0
151.10	1299.0
152.10	773.0
153.00	2321.0
153.90	1903.0
155.00	4817.0
156.00	6037.0
157.00	1599.0
157.90	1191.0
159.10	957.0
160.00	2791.0
161.00	2490.0
162.10	1086.0
163.00	840.0
164.90	3492.0
166.00	2220.0
167.00	17320.0
168.00	11069.0
168.90	1222.0
169.90	1000.0
170.90	731.0
171.90	1696.0
173.10	2042.0
174.00	4606.0
175.00	5058.0
176.10	1673.0
177.10	2591.0
178.00	664.0
179.00	13223.0
180.00	8246.0
180.90	4159.0
181.90	540.0
182.80	653.0
183.90	1200.0
184.90	6197.0
186.00	48480.0
187.00	15380.0
188.00	1613.0
188.90	3276.0
189.80	630.0
190.20	554.0
191.00	1382.0
191.90	4225.0
193.10	4139.0
194.10	1723.0
194.90	456.0
196.00	9184.0
197.00	1413.0
198.00	303360.0
198.90	21840.0
200.00	1565.0
201.40	1705.0
202.90	2263.0
204.00	12833.0
205.00	19192.0
206.00	85304.0
207.00	13565.0
208.00	3482.0
208.90	1312.0
210.00	1923.0
211.10	3104.0
211.90	493.0
213.20	354.0
214.90	2032.0
216.10	2653.0
217.00	25200.0
218.00	3190.0
218.90	440.0
221.20	9058.0
222.00	2440.0
223.00	5509.0
224.00	50424.0
225.00	12170.0
226.00	973.0
227.00	23768.0
227.90	3440.0
229.00	4362.0
229.90	602.0
230.90	2167.0
234.00	1528.0
234.90	1256.0
235.90	868.0
237.10	1677.0
238.10	306.0
239.10	703.0
239.80	439.0
241.10	1098.0
242.10	2028.0
243.00	2562.0
244.10	38432.0
245.10	5408.0
246.00	9717.0
246.90	2144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

247.90	355.0
249.00	1221.0
250.90	399.0
251.90	801.0
252.60	390.0
253.20	520.0
254.00	1262.0
255.00	199488.0
256.00	27120.0
257.00	1977.0
258.00	12707.0
259.00	1437.0
260.00	486.0
261.00	351.0
265.00	5016.0
266.00	512.0
270.00	360.0
271.10	390.0
272.00	498.0
273.00	5376.0
274.00	15426.0
275.00	87936.0
276.00	10101.0
277.00	7260.0
277.90	1201.0
281.00	513.0
282.00	536.0
283.10	1222.0
284.00	588.0
285.10	1729.0
289.00	515.0
290.00	528.0
291.90	662.0
293.00	2081.0
293.90	576.0
296.00	31008.0
296.90	4612.0
297.80	457.0
302.00	643.0
303.10	2534.0
304.00	1469.0
309.00	460.0
313.90	1519.0
314.90	2749.0
316.10	1792.0
316.90	449.0
321.20	672.0
323.00	7447.0
324.00	1515.0
326.90	2144.0
327.90	818.0
331.90	661.0
333.10	661.0
334.00	5410.0
335.00	1437.0
341.00	1018.0
346.00	1504.0
352.00	1815.0
353.10	1404.0
354.00	2315.0
355.20	616.0
365.00	11845.0
366.00	2438.0
371.00	744.0
372.10	4353.0
372.90	1056.0
382.90	1047.0
389.90	307.0
402.00	1221.0
403.10	2254.0
404.00	1298.0
420.90	1172.0
422.00	1938.0
423.00	13294.0
424.00	3396.0
441.00	32440.0
442.10	239232.0
443.10	47216.0
444.00	5264.0
445.10	480.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

m/z	Abundance
36.90	682.0
38.00	1310.0
39.00	6147.0
40.00	768.0
44.00	1659.0
49.10	691.0
50.10	23592.0
51.10	81536.0
52.00	3887.0
53.10	341.0
54.90	454.0
56.00	2620.0
57.00	5676.0
61.00	1081.0
62.00	1531.0
63.10	4014.0
64.10	561.0
65.00	1585.0
68.20	1386.0
69.00	89088.0
70.00	586.0
73.00	1659.0
74.10	8255.0
75.00	14923.0
76.00	4305.0
77.00	91368.0
78.00	6578.0
79.00	7480.0
80.00	5249.0
81.00	7323.0
82.00	1676.0
82.90	1256.0
84.00	341.0
85.00	1443.0
86.00	2441.0
87.00	1006.0
90.90	2017.0
92.00	1557.0
93.00	16528.0
94.00	613.0
96.00	882.0
96.90	323.0
98.00	11522.0
99.00	7463.0
100.00	647.0
100.90	4490.0
103.00	1402.0
104.00	2826.0
104.90	2980.0
106.20	677.0
107.00	34208.0
108.00	5230.0
109.10	1551.0
110.00	52792.0
111.00	7939.0
112.00	806.0
115.90	2362.0
117.00	37536.0
118.00	1766.0
119.00	329.0
119.90	490.0
122.00	2629.0
123.00	3723.0
124.10	2413.0
124.90	1867.0
127.00	110472.0
128.00	8595.0
129.00	48504.0
130.00	3961.0
131.00	1481.0
132.90	571.0
134.00	923.0
135.00	3576.0
136.00	1687.0
136.90	2027.0
137.90	505.0
138.90	431.0
140.00	722.0
141.00	6620.0
142.00	1763.0
142.90	1549.0
143.90	600.0
145.10	528.0
145.90	1211.0
146.90	2960.0
148.00	7209.0
148.90	1981.0
150.00	615.0
151.10	826.0
151.70	773.0
153.10	1541.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

154.00	1804.0
155.00	2952.0
156.10	4520.0
157.10	1313.0
158.00	1498.0
159.10	959.0
160.10	1705.0
160.90	2564.0
162.20	917.0
163.00	466.0
163.90	371.0
164.90	2744.0
165.90	2610.0
167.00	14721.0
168.00	7739.0
169.00	896.0
170.90	870.0
171.90	1229.0
173.10	1904.0
174.00	3017.0
175.10	4738.0
176.10	2180.0
177.00	2972.0
177.90	1213.0
179.00	9943.0
180.00	6999.0
181.00	4311.0
182.00	853.0
184.10	947.0
185.00	5167.0
186.10	32144.0
187.00	10389.0
187.90	1559.0
189.00	3222.0
190.10	660.0
191.10	1651.0
191.90	2499.0
193.00	3661.0
193.90	864.0
195.10	782.0
195.90	7044.0
196.90	1331.0
197.90	247040.0
198.90	17440.0
199.90	1527.0
200.20	1604.0
201.40	1022.0
203.00	1587.0
204.00	10545.0
205.00	17160.0
206.00	71200.0
207.00	11080.0
207.90	3680.0
209.00	1132.0
209.90	1008.0
211.00	3569.0
215.00	1191.0
215.90	1395.0
216.90	23480.0
217.90	2304.0
218.90	368.0
220.90	9113.0
221.90	2415.0
223.10	3546.0
224.00	40696.0
225.00	10483.0
225.90	990.0
227.00	20608.0
228.00	2894.0
229.00	5081.0
230.00	906.0
231.00	1890.0
231.80	441.0
234.00	1430.0
235.00	1389.0
236.10	836.0
237.00	1350.0
238.00	385.0
239.00	833.0
239.80	690.0
240.90	811.0
242.00	2726.0
243.10	2324.0
244.10	35040.0
245.10	3994.0
246.00	8355.0
246.90	1511.0
248.10	350.0
249.00	1035.0
251.00	437.0
251.90	601.0
253.00	1120.0
254.10	1757.0
255.00	165504.0
256.00	23344.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

256.90	1628.0
258.00	9721.0
258.90	1493.0
261.00	475.0
263.90	403.0
265.00	4216.0
265.90	611.0
270.90	400.0
272.10	478.0
273.00	4868.0
274.00	14875.0
275.00	77672.0
276.00	9878.0
277.00	7010.0
278.20	878.0
278.80	432.0
281.10	466.0
282.20	462.0
282.90	855.0
283.90	500.0
285.00	990.0
286.00	334.0
289.00	561.0
292.00	343.0
293.00	1650.0
293.90	978.0
296.00	27664.0
297.00	4068.0
301.80	354.0
302.90	3191.0
304.10	999.0
313.10	341.0
314.10	1098.0
315.10	2430.0
316.00	1647.0
321.10	374.0
322.10	438.0
323.10	6398.0
324.20	1196.0
327.00	1519.0
328.00	677.0
332.00	789.0
333.00	556.0
334.00	5098.0
335.00	1108.0
341.00	1285.0
346.00	1896.0
346.90	364.0
352.00	2519.0
353.00	1419.0
354.10	1638.0
359.10	317.0
365.00	12103.0
366.00	1437.0
370.00	303.0
371.10	584.0
372.10	4316.0
372.90	1208.0
383.00	722.0
383.90	419.0
389.80	320.0
390.90	386.0
402.00	1406.0
402.90	2084.0
404.00	841.0
421.10	1618.0
422.00	1979.0
423.10	12686.0
424.00	3449.0
441.10	37616.0
442.10	247360.0
443.10	46720.0
444.10	4861.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67

Scan 2446 (17.474 min): BM013522.D
DFTPP67

m/z	Abundance
39.90	510.0
44.00	1661.0
73.20	378.0
191.00	366.0
207.10	1440.0
208.00	474.0
281.00	412.0

Instrument :
BNA_M
ClientSampleId :
DFTPP67