

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BM038180.D
 Acq On : 21 Dec 2022 20:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Dec 21 23:33:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.039	152	189389	20.000	ng/u1	0.00
20) Naphthalene-d8	10.857	136	821041	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.668	164	518749	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.409	188	1068518	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.574	240	738811	20.000	ng/u1	0.00
88) Perylene-d12	24.027	264	727835	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	29534	7.081	ng/uL	0.00
4) Pyridine-d5	3.828	84	224477	18.142	ng/u1	0.00
7) Phenol-d5	7.204	99	296159	19.244	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	172461	18.403	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	261368	20.703	ng/u1	0.00
15) 4-Methylphenol-d8	8.757	113	254472	21.183	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	133818	20.585	ng/u1	0.00
24) 2-Nitrophenol-d4	9.939	143	153842	22.583	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	292616	22.614	ng/u1	0.00
31) 4-Chloroaniline-d4	10.998	131	391885	21.940	ng/u1	0.00
46) Dimethylphthalate-d6	14.080	166	814173	22.015	ng/u1	0.00
49) Acenaphthylene-d8	14.368	160	957476	21.183	ng/u1	0.00
54) 4-Nitrophenol-d4	14.874	143	136263	21.619	ng/u1	0.00
60) Fluorene-d10	15.657	176	698592	21.202	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.780	200	94656	16.569	ng/u1	0.00
73) Anthracene-d10	17.503	188	1026423	20.582	ng/u1	0.00
81) Pyrene-d10	19.792	212	1016202	22.898	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.862	264	764243	21.022	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	30909	7.064	ng/uL	95
5) Pyridine	3.846	79	219203	17.628	ng/u1	96
6) Benzaldehyde	7.175	77	113077	19.645	ng/u1	98
8) Phenol	7.228	94	294727	19.095	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.463	93	237119	18.708	ng/u1	96
12) 2-Chlorophenol	7.604	128	268227	20.774	ng/u1	99
13) 2-Methylphenol	8.486	108	240168	20.333	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.575	45	412773	18.036	ng/u1	97
16) Acetophenone	8.875	105	399634	21.152	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.857	70	187881	21.109	ng/u1	97
18) 4-Methylphenol	8.822	108	263333	20.609	ng/u1	97
19) Hexachloroethane	9.128	117	105644	20.467	ng/u1	95
22) Nitrobenzene	9.257	77	283477	20.267	ng/u1	96
23) Isophorone	9.780	82	545480	21.282	ng/u1	99
25) 2-Nitrophenol	9.969	139	158439	22.184	ng/u1	95
26) 2,4-Dimethylphenol	10.027	107	304477	22.045	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.263	93	338886	19.890	ng/u1	96
29) 2,4-Dichlorophenol	10.504	162	282047	22.217	ng/u1	99
30) Naphthalene	10.910	128	909231	20.521	ng/u1	99
32) 4-Chloroaniline	11.022	127	380200	21.346	ng/u1	98
33) Hexachlorobutadiene	11.180	225	190213	22.776	ng/u1	99
34) Caprolactam	11.810	113	77132	23.685	ng/u1	93
35) 4-Chloro-3-methylphenol	12.139	107	275782	23.865	ng/u1	95
36) 2-Methylnaphthalene	12.504	142	610511	21.244	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BM038180.D
 Acq On : 21 Dec 2022 20:40
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Dec 21 23:33:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.727	142	625922	21.301	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.868	216	351484	20.737	ng/ul	97
40) Hexachlorocyclopentadiene	12.845	237	69558	7.174	ng/ul	98
41) 2,4,6-Trichlorophenol	13.116	196	230704	22.346	ng/ul	97
42) 2,4,5-Trichlorophenol	13.186	196	252717	23.471	ng/ul	99
43) 1,1'-Biphenyl	13.510	154	837355	20.082	ng/ul#	98
44) 2-Chloronaphthalene	13.551	162	653562	20.318	ng/ul	98
45) 2-Nitroaniline	13.763	65	168463	21.735	ng/ul	94
47) Dimethylphthalate	14.127	163	801631	21.969	ng/ul	100
48) 2,6-Dinitrotoluene	14.251	165	159788	22.545	ng/ul	95
50) Acenaphthylene	14.392	152	1027578	20.675	ng/ul	99
51) 3-Nitroaniline	14.580	138	160896	22.753	ng/ul	98
52) Acenaphthene	14.733	153	702160	20.425	ng/ul	98
53) 2,4-Dinitrophenol	14.786	184	55629	15.995	ng/ul	91
55) 4-Nitrophenol	14.886	109	109664	26.383	ng/ul	84
56) Dibenzofuran	15.068	168	982794	20.940	ng/ul	95
57) 2,4-Dinitrotoluene	15.033	165	225114	23.453	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.292	232	211176	23.315	ng/ul	95
59) Diethylphthalate	15.480	149	824048	22.670	ng/ul	98
61) Fluorene	15.709	166	809332	21.145	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.704	204	418245	22.359	ng/ul	97
63) 4-Nitroaniline	15.739	138	158492	24.719	ng/ul#	89
66) 4,6-Dinitro-2-methylph...	15.792	198	88203	15.948	ng/ul	98
67) N-Nitrosodiphenylamine	15.915	169	661799	20.533	ng/ul	100
68) 4-Bromophenyl-phenylether	16.592	248	243361	20.777	ng/ul	95
69) Hexachlorobenzene	16.709	284	292173	20.897	ng/ul#	93
70) Atrazine	16.862	200	250025	22.699	ng/ul	95
71) Pentachlorophenol	17.056	266	164495	21.161	ng/ul	100
72) Phenanthrene	17.451	178	1173834	20.117	ng/ul	99
74) Anthracene	17.539	178	1219403	20.563	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.474	216	344002	19.914	ng/uL	99
76) Pentachlorobenzene	14.986	250	344965	20.159	ng/uL	99
77) Carbazole	17.809	167	1042911	20.299	ng/ul	99
78) Di-n-butylphthalate	18.362	149	1423539	23.311	ng/ul	100
80) Fluoranthene	19.456	202	1230793	23.499	ng/ul#	94
82) Pyrene	19.821	202	1247454	22.884	ng/ul#	93
83) Butylbenzylphthalate	20.697	149	549632	27.116	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.491	252	241536	16.521	ng/ul	97
85) Benzo(a)anthracene	21.556	228	1053073	21.172	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	826795	28.038	ng/ul	100
87) Chrysene	21.615	228	954824	20.460	ng/ul	100
89) Di-n-octyl phthalate	22.403	149	1251917	26.844	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	957281	21.215	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	941285	21.524	ng/ul	99
93) Benzo(a)pyrene	23.915	252	830660	20.934	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.591	276	1076735	21.070	ng/ul	99
95) Dibenzo(a,h)anthracene	26.603	278	913247	21.198	ng/ul	98
96) Benzo(g,h,i)perylene	27.379	276	868090	21.276	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BM038180.D
 Acq On : 21 Dec 2022 20:40
 Operator : CG/JU
 Sample : SSTCCCC020
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :


Quant Time: Dec 21 23:33:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.064 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Instrument :
BNA_M
ClientSampleId :







Tgt Ion: 88 Resp: 30909
Ion Ratio Lower Upper
88 100
43 43.0 32.8 49.2
58 85.6 64.6 97.0



#5
Pyridine
Concen: 17.628 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion: 79 Resp: 219203
Ion Ratio Lower Upper
79 100
52 102.9 78.1 117.1
51 48.0 37.2 55.8









#6
Benzaldehyde
Concen: 19.645 ng/ul
RT: 7.175 min Scan# 61
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 113077
Ion Ratio Lower Upper
77 100
105 100.7 82.5 123.7
106 101.1 81.8 122.6



#8
Phenol
Concen: 19.095 ng/ul
RT: 7.228 min Scan# 687
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion: 94 Resp: 294727
Ion Ratio Lower Upper
94 100
65 29.2 22.2 33.4
66 37.8 27.3 40.9









#10
Bis(2-Chloroethyl)ether
Concen: 18.708 ng/ul
RT: 7.463 min Scan# 727
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Instrument :
BNA_M
ClientSampleId :

Tgt Ion: 93 Resp: 237119




Ion	Ratio	Lower	Upper
93	100		
63	78.6	66.3	99.5
95	33.2	25.7	38.5




#12
2-Chlorophenol
Concen: 20.774 ng/ul
RT: 7.604 min Scan# 751
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:128 Resp: 268227

Ion	Ratio	Lower	Upper
128	100		
64	43.0	35.2	52.8
130	32.6	25.9	38.9









#13
2-Methylphenol
Concen: 20.333 ng/ul
RT: 8.486 min Scan# 901
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 240168
Ion Ratio Lower Upper
108 100
107 90.9 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.036 ng/ul
RT: 8.575 min Scan# 916
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion: 45 Resp: 412773
Ion Ratio Lower Upper
45 100
77 11.9 9.0 13.4
79 9.0 6.2 9.4









#16
Acetophenone
Concen: 21.152 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 399634
Ion Ratio Lower Upper
105 100
77 73.3 59.7 89.5
51 32.8 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 21.109 ng/ul
RT: 8.857 min Scan# 964
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion: 70 Resp: 187881
Ion Ratio Lower Upper
70 100
42 75.6 62.3 93.5
101 9.5 8.0 12.0
130 23.3 17.6 26.4









#18
4-Methylphenol
Concen: 20.609 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 263333
Ion Ratio Lower Upper
108 100
107 107.0 88.1 132.1



#19
Hexachloroethane
Concen: 20.467 ng/ul
RT: 9.128 min Scan# 1010
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:117 Resp: 105644
Ion Ratio Lower Upper
117 100
201 112.7 87.3 130.9
199 69.4 50.9 76.3









#22
Nitrobenzene
Concen: 20.267 ng/ul
RT: 9.257 min Scan# 1032
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 283477
Ion Ratio Lower Upper
77 100
123 50.2 37.8 56.6
65 14.6 11.1 16.7



#23
Isophorone
Concen: 21.282 ng/ul
RT: 9.780 min Scan# 1121
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion: 82 Resp: 545480
Ion Ratio Lower Upper
82 100
95 7.2 5.2 7.8
138 19.0 14.9 22.3









#25
2-Nitrophenol
Concen: 22.184 ng/ul
RT: 9.969 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 158439
Ion Ratio Lower Upper
139 100
65 47.4 36.2 54.4
109 28.7 19.7 29.5



#26
2,4-Dimethylphenol
Concen: 22.045 ng/ul
RT: 10.027 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:107 Resp: 304477
Ion Ratio Lower Upper
107 100
121 49.8 42.0 63.0
122 87.9 71.4 107.2





#27

Bis(2-Chloroethoxy)methane

Concen: 19.890 ng/ul

RT: 10.263 min Scan# 1203

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion: 93 Resp: 338886

Ion Ratio Lower Upper

93 100

95 34.3 25.5 38.3

123 12.3 9.4 14.2



#29

2,4-Dichlorophenol

Concen: 22.217 ng/ul

RT: 10.504 min Scan# 1244

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:162 Resp: 282047


Ion Ratio Lower Upper

162 100

164 65.0 51.4 77.2

98 33.2 26.1 39.1









#30
Naphthalene
Concen: 20.521 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 909231
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.4
127 12.7 10.4 15.6



#32
4-Chloroaniline
Concen: 21.346 ng/ul
RT: 11.022 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:127 Resp: 380200
Ion Ratio Lower Upper
127 100
129 32.8 25.3 37.9









#33
Hexachlorobutadiene
Concen: 22.776 ng/ul
RT: 11.180 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 190213
Ion Ratio Lower Upper
225 100
223 64.9 50.8 76.2
227 63.7 50.5 75.7



#34
Caprolactam
Concen: 23.685 ng/ul
RT: 11.810 min Scan# 1466
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:113 Resp: 77132
Ion Ratio Lower Upper
113 100
55 198.7 172.2 258.2
56 152.5 120.7 181.1









#35
 4-Chloro-3-methylphenol
 Concen: 23.865 ng/ul
 RT: 12.139 min Scan# 1522
 Delta R.T. -0.000 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:107 Resp: 275782
 Ion Ratio Lower Upper
 107 100
 144 26.7 23.4 35.0
 142 84.5 71.0 106.4



#36
 2-Methylnaphthalene
 Concen: 21.244 ng/ul
 RT: 12.504 min Scan# 1584
 Delta R.T. -0.006 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40

Tgt Ion:142 Resp: 610511
 Ion Ratio Lower Upper
 142 100
 141 88.1 69.4 104.0





#37

1-Methylnaphthalene

Concen: 21.301 ng/ul

RT: 12.727 min Scan# 1

Delta R.T. -0.000 min


Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:142 Resp: 625922

Ion Ratio Lower Upper

142 100

141 92.9 72.4 108.6

116 3.6 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.737 ng/ul

RT: 12.868 min Scan# 1646

Delta R.T. -0.006 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Tgt Ion:216 Resp: 351484




Ion Ratio Lower Upper


216 100

214 77.2 64.2 96.4

179 20.4 15.4 23.2

108 16.7 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 7.174 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:237 Resp: 69558
Ion Ratio Lower Upper
237 100
235 61.2 50.4 75.6
272 12.4 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 22.346 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:196 Resp: 230704
Ion Ratio Lower Upper
196 100
198 96.5 75.0 112.6
200 32.5 24.5 36.7





#42

2,4,5-Trichlorophenol

Concen: 23.471 ng/ul

RT: 13.186 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:196 Resp: 252717

Ion Ratio Lower Upper

196 100

198 95.0 76.5 114.7

200 29.2 24.9 37.3



#43

1,1'-Biphenyl

Concen: 20.082 ng/ul

RT: 13.510 min Scan# 1755

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:154 Resp: 837355


Ion Ratio Lower Upper

154 100

153 39.6 31.2 46.8

76 9.3 9.3 13.9#





#44

2-Chloronaphthalene

Concen: 20.318 ng/ul

RT: 13.551 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:162 Resp: 653562

Ion Ratio Lower Upper

162 100

164 32.9 27.3 40.9

127 37.9 29.4 44.0



#45

2-Nitroaniline

Concen: 21.735 ng/ul

RT: 13.763 min Scan# 1798

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion: 65 Resp: 168463


Ion Ratio Lower Upper

65 100

92 69.2 53.5 80.3

138 122.3 91.3 136.9





#47

Dimethylphthalate

Concen: 21.969 ng/ul

RT: 14.127 min Scan# 1

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:163 Resp: 801631

Ion Ratio Lower Upper

163 100

194 3.8 3.0 4.6

164 10.1 8.2 12.4



#48

2,6-Dinitrotoluene

Concen: 22.545 ng/ul

RT: 14.251 min Scan# 1881

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:165 Resp: 159788


Ion Ratio Lower Upper

165 100

89 56.5 41.1 61.7

121 21.3 17.4 26.2









#50
Acenaphthylene
Concen: 20.675 ng/ul
RT: 14.392 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1027578
Ion Ratio Lower Upper
152 100
151 19.8 15.3 22.9
153 13.1 10.2 15.4



#51
3-Nitroaniline
Concen: 22.753 ng/ul
RT: 14.580 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:138 Resp: 160896
Ion Ratio Lower Upper
138 100
108 9.5 7.9 11.9
92 106.2 83.4 125.0









#52
Acenaphthene
Concen: 20.425 ng/ul
RT: 14.733 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:153 Resp: 702160
Ion Ratio Lower Upper
153 100
152 47.2 38.2 57.4
154 87.1 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 15.995 ng/ul
RT: 14.786 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:184 Resp: 55629
Ion Ratio Lower Upper
184 100
63 63.1 46.2 69.4
154 64.5 45.5 68.3









#55
4-Nitrophenol
Concen: 26.383 ng/ul
RT: 14.886 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 109664
Ion Ratio Lower Upper
109 100
139 126.4 119.2 178.8
65 119.2 108.7 163.1



#56
Dibenzofuran
Concen: 20.940 ng/ul
RT: 15.068 min Scan# 2020
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:168 Resp: 982794
Ion Ratio Lower Upper
168 100
139 37.9 28.1 42.1





#57

2,4-Dinitrotoluene

Concen: 23.453 ng/ul

RT: 15.033 min Scan# 2014

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:165 Resp: 225114

Ion Ratio Lower Upper

165 100

63 42.0 34.9 52.3

182 2.3 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 23.315 ng/ul

RT: 15.292 min Scan# 2058

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Tgt Ion:232 Resp: 211176




Ion Ratio Lower Upper


232 100

131 44.6 32.8 49.2

130 2.5 2.0 3.0

166 31.1 23.2 34.8









#59
Diethylphthalate
Concen: 22.670 ng/ul
RT: 15.480 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 824048
Ion Ratio Lower Upper
149 100
177 22.1 18.5 27.7
150 11.6 9.5 14.3



#61
Fluorene
Concen: 21.145 ng/ul
RT: 15.709 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:166 Resp: 809332
Ion Ratio Lower Upper
166 100
165 97.6 77.2 115.8
167 14.0 10.4 15.6









#62
 4-Chlorophenyl-phenylether
 Concen: 22.359 ng/ul
 RT: 15.704 min Scan# 2128
 Delta R.T. -0.000 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:204 Resp: 418245
 Ion Ratio Lower Upper
 204 100
 206 32.9 25.5 38.3
 141 56.2 42.8 64.2



#63
 4-Nitroaniline
 Concen: 24.719 ng/ul
 RT: 15.739 min Scan# 2134
 Delta R.T. -0.000 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40

Tgt Ion:138 Resp: 158492
 Ion Ratio Lower Upper
 138 100
 92 47.3 36.1 54.1
 108 77.8 51.6 77.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 15.948 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:198 Resp: 88203
Ion Ratio Lower Upper
198 100
182 4.9 3.8 5.8
77 25.1 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 20.533 ng/ul
RT: 15.915 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40


Tgt Ion:169 Resp: 661799
Ion Ratio Lower Upper
169 100
168 67.0 53.8 80.8
167 36.6 29.3 43.9








#68
4-Bromophenyl-phenylether
Concen: 20.777 ng/ul
RT: 16.592 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40


Instrument :
BNA_M
ClientSampleId :





Tgt Ion:248 Resp: 243361
Ion Ratio Lower Upper
248 100
250 98.4 78.7 118.1
141 76.2 52.7 79.1




#69
Hexachlorobenzene
Concen: 20.897 ng/ul
RT: 16.709 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40



Tgt Ion:284 Resp: 292173
Ion Ratio Lower Upper
284 100
142 34.9 23.1 34.7#
249 30.2 23.3 34.9









#70
Atrazine
Concen: 22.699 ng/ul
RT: 16.862 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:200 Resp: 250025
Ion Ratio Lower Upper
200 100
173 25.1 20.5 30.7
215 45.8 40.1 60.1



#71
Pentachlorophenol
Concen: 21.161 ng/ul
RT: 17.056 min Scan# 2358
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:266 Resp: 164495
Ion Ratio Lower Upper
266 100
264 62.0 50.1 75.1
268 63.4 50.6 75.8









#72
Phenanthrene
Concen: 20.117 ng/ul
RT: 17.451 min Scan# 2425
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:178 Resp: 1173834
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.2
176 19.7 15.4 23.0



#74
Anthracene
Concen: 20.563 ng/ul
RT: 17.539 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:178 Resp: 1219403
Ion Ratio Lower Upper
178 100
179 15.6 11.8 17.8
176 18.6 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.914 ng/uL

RT: 13.474 min Scan# 1

Delta R.T. -0.006 min


Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:216 Resp: 344002

Ion Ratio Lower Upper

216 100

214 77.8 62.6 93.8

218 48.7 38.3 57.5



#76

Pentachlorobenzene

Concen: 20.159 ng/uL

RT: 14.986 min Scan# 2006

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:250 Resp: 344965


Ion Ratio Lower Upper

250 100

248 63.4 51.0 76.4

252 63.1 51.8 77.6









#77
 Carbazole
 Concen: 20.299 ng/ul
 RT: 17.809 min Scan# 2487
 Delta R.T. -0.006 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:167 Resp: 1042911
 Ion Ratio Lower Upper
 167 100
 166 21.2 16.7 25.1
 139 13.0 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 23.311 ng/ul
 RT: 18.362 min Scan# 2580
 Delta R.T. -0.000 min
 Lab File: BM038180.D
 Acq: 21 Dec 2022 20:40

Tgt Ion:149 Resp: 1423539
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 4.7 3.4 5.2









#80
Fluoranthene
Concen: 23.499 ng/ul
RT: 19.456 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:202 Resp: 1230793
Ion Ratio Lower Upper
202 100
101 7.9 8.6 13.0#
100 6.5 6.6 10.0#



#82
Pyrene
Concen: 22.884 ng/ul
RT: 19.821 min Scan# 2828
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:202 Resp: 1247454
Ion Ratio Lower Upper
202 100
101 9.3 10.0 15.0#
100 7.9 7.9 11.9









#83
Butylbenzylphthalate
Concen: 27.116 ng/ul
RT: 20.697 min Scan# 2977
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 549632
Ion Ratio Lower Upper
149 100
91 64.9 54.5 81.7
206 19.1 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 16.521 ng/ul
RT: 21.491 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:252 Resp: 241536
Ion Ratio Lower Upper
252 100
254 63.8 53.2 79.8
126 9.3 7.8 11.8





#85

Benzo(a)anthracene

Concen: 21.172 ng/ul

RT: 21.556 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 1053073

Ion Ratio Lower Upper

228 100

229 19.8 15.5 23.3

226 27.1 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 28.038 ng/ul

RT: 21.468 min Scan# 3108

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:149 Resp: 826795


Ion Ratio Lower Upper

149 100

167 27.1 21.5 32.3

279 3.8 3.2 4.8









#87
Chrysene
Concen: 20.460 ng/ul
RT: 21.615 min Scan# 3133
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 954824
Ion Ratio Lower Upper
228 100
226 30.4 24.3 36.5
229 19.3 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 26.844 ng/ul
RT: 22.403 min Scan# 3267
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:149 Resp: 1251917









#90
Benzo(b)fluoranthene
Concen: 21.215 ng/ul
RT: 23.274 min Scan# 3415
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 957281
Ion Ratio Lower Upper
252 100
253 21.7 17.6 26.4
125 8.9 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 21.524 ng/ul
RT: 23.321 min Scan# 3423
Delta R.T. -0.006 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:252 Resp: 941285
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 9.1 8.1 12.1





#93

Benzo(a)pyrene

Concen: 20.934 ng/ul

RT: 23.915 min Scan# 3

Delta R.T. -0.006 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:252 Resp: 830660

Ion Ratio Lower Upper

252 100

253 21.5 17.4 26.0

125 10.6 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.070 ng/ul

RT: 26.591 min Scan# 3979

Delta R.T. -0.000 min

Lab File: BM038180.D

Acq: 21 Dec 2022 20:40




Tgt Ion:276 Resp: 1076735


Ion Ratio Lower Upper

276 100

138 23.0 19.3 28.9

277 24.7 19.8 29.8









#95
Dibenzo(a,h)anthracene
Concen: 21.198 ng/ul
RT: 26.603 min Scan# 3980
Delta R.T. -0.000 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Instrument :
BNA_M
ClientSampleId :

Tgt Ion:278 Resp: 913247
Ion Ratio Lower Upper
278 100
139 14.2 12.4 18.6
279 24.1 19.8 29.6



#96
Benzo(g,h,i)perylene
Concen: 21.276 ng/ul
RT: 27.379 min Scan# 4113
Delta R.T. -0.012 min
Lab File: BM038180.D
Acq: 21 Dec 2022 20:40

Tgt Ion:276 Resp: 868090
Ion Ratio Lower Upper
276 100
138 21.1 17.1 25.7
277 24.1 18.2 27.4

