

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BMCCC180.D
 Acq On : 21 Dec 2022 20:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Dec 21 23:34:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	217973	20.000	ng/ul	0.00
20) Naphthalene-d8	10.857	136	926565	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.669	164	573244	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.410	188	1216710	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.574	240	855318	20.000	ng/ul	0.00
88) Perylene-d12	24.027	264	838063	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	29793	6.206	ng/uL	0.00
4) Pyridine-d5	3.828	84	231302	16.242	ng/ul	0.00
7) Phenol-d5	7.204	99	315223	17.796	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	180961	16.778	ng/ul	0.00
11) 2-Chlorophenol-d4	7.569	132	281496	19.373	ng/ul	0.00
15) 4-Methylphenol-d8	8.757	113	266084	19.245	ng/ul	0.00
21) Nitrobenzene-d5	9.216	128	144217	19.658	ng/ul	0.00
24) 2-Nitrophenol-d4	9.939	143	162968	21.198	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	310492	21.263	ng/ul	0.00
31) 4-Chloroaniline-d4	10.998	131	405917	20.138	ng/ul	0.00
46) Dimethylphthalate-d6	14.080	166	838248	20.511	ng/ul	0.00
49) Acenaphthylene-d8	14.369	160	1000346	20.027	ng/ul	0.00
54) 4-Nitrophenol-d4	14.874	143	150110	21.552	ng/ul	0.00
60) Fluorene-d10	15.657	176	733345	20.141	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.780	200	96161	14.783	ng/ul	0.00
73) Anthracene-d10	17.510	188	1108993	19.530	ng/ul	0.00
81) Pyrene-d10	19.792	212	1129964	21.993	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.868	264	826657	19.748	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.428	88	31289	6.213	ng/uL	97
5) Pyridine	3.846	79	229628	16.045	ng/ul	93
6) Benzaldehyde	7.181	77	118969	17.958	ng/ul	99
8) Phenol	7.228	94	311800	17.552	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.463	93	252091	17.281	ng/ul	96
12) 2-Chlorophenol	7.604	128	285694	19.225	ng/ul	99
13) 2-Methylphenol	8.487	108	257130	18.914	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.575	45	435752	16.543	ng/ul	99
16) Acetophenone	8.875	105	418250	19.234	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.857	70	193032	18.844	ng/ul	97
18) 4-Methylphenol	8.822	108	278931	18.967	ng/ul	100
19) Hexachloroethane	9.128	117	113370	19.083	ng/ul	95
22) Nitrobenzene	9.257	77	302484	19.163	ng/ul	97
23) Isophorone	9.787	82	563525	19.482	ng/ul	98
25) 2-Nitrophenol	9.969	139	167134	20.736	ng/ul#	94
26) 2,4-Dimethylphenol	10.028	107	321664	20.637	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.263	93	354974	18.461	ng/ul	99
29) 2,4-Dichlorophenol	10.504	162	300598	20.981	ng/ul	98
30) Naphthalene	10.910	128	968887	19.377	ng/ul	99
32) 4-Chloroaniline	11.022	127	405920	20.194	ng/ul	99
33) Hexachlorobutadiene	11.181	225	208413	22.113	ng/ul	99
34) Caprolactam	11.810	113	80976	22.034	ng/ul	88
35) 4-Chloro-3-methylphenol	12.139	107	283256	21.720	ng/ul	96
36) 2-Methylnaphthalene	12.510	142	648685	20.002	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BMCCC180.D
 Acq On : 21 Dec 2022 20:03
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Dec 21 23:34:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.728	142	661207	19.939	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.869	216	369511	19.728	ng/ul	98
40) Hexachlorocyclopentadiene	12.845	237	62909	5.871	ng/ul	100
41) 2,4,6-Trichlorophenol	13.116	196	241077	21.131	ng/ul	96
42) 2,4,5-Trichlorophenol	13.186	196	260994	21.935	ng/ul	99
43) 1,1'-Biphenyl	13.510	154	883519	19.174	ng/ul#	98
44) 2-Chloronaphthalene	13.551	162	687642	19.345	ng/ul	96
45) 2-Nitroaniline	13.763	65	177353	20.707	ng/ul	97
47) Dimethylphthalate	14.127	163	829399	20.570	ng/ul	99
48) 2,6-Dinitrotoluene	14.251	165	164915	21.056	ng/ul	94
50) Acenaphthylene	14.392	152	1087543	19.801	ng/ul	100
51) 3-Nitroaniline	14.580	138	169151	21.647	ng/ul	100
52) Acenaphthene	14.733	153	738954	19.452	ng/ul	98
53) 2,4-Dinitrophenol	14.786	184	56600	14.727	ng/ul	90
55) 4-Nitrophenol	14.886	109	114203	24.863	ng/ul	86
56) Dibenzofuran	15.069	168	1028480	19.830	ng/ul	96
57) 2,4-Dinitrotoluene	15.033	165	236473	22.295	ng/ul#	100
58) 2,3,4,6-Tetrachlorophenol	15.292	232	221256	22.106	ng/ul	95
59) Diethylphthalate	15.480	149	849497	21.148	ng/ul	99
61) Fluorene	15.716	166	856804	20.257	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.704	204	436385	21.111	ng/ul	96
63) 4-Nitroaniline	15.739	138	173251	24.452	ng/ul#	85
66) 4,6-Dinitro-2-methylph...	15.798	198	90701	14.403	ng/ul#	95
67) N-Nitrosodiphenylamine	15.916	169	694133	18.913	ng/ul	99
68) 4-Bromophenyl-phenylether	16.592	248	258917	19.413	ng/ul	96
69) Hexachlorobenzene	16.715	284	311533	19.568	ng/ul	97
70) Atrazine	16.868	200	260280	20.752	ng/ul	97
71) Pentachlorophenol	17.057	266	180561	20.399	ng/ul	98
72) Phenanthrene	17.451	178	1260906	18.977	ng/ul	100
74) Anthracene	17.545	178	1297066	19.209	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.475	216	358681	18.235	ng/ul	99
76) Pentachlorobenzene	14.986	250	368111	18.892	ng/ul	97
77) Carbazole	17.815	167	1118215	19.114	ng/ul	99
78) Di-n-butylphthalate	18.362	149	1512076	21.745	ng/ul	100
80) Fluoranthene	19.456	202	1345911	22.196	ng/ul#	93
82) Pyrene	19.821	202	1385519	21.955	ng/ul#	93
83) Butylbenzylphthalate	20.698	149	604463	25.759	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.486	252	259639	15.341	ng/ul	99
85) Benzo(a)anthracene	21.556	228	1157995	20.110	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.468	149	904154	26.485	ng/ul	99
87) Chrysene	21.615	228	1073390	19.868	ng/ul	99
89) Di-n-octyl phthalate	22.403	149	1369375	25.501	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	1046634	20.144	ng/ul	99
91) Benzo(k)fluoranthene	23.321	252	1024517	20.346	ng/ul	99
93) Benzo(a)pyrene	23.915	252	914304	20.011	ng/ul	96
94) Indeno(1,2,3-cd)pyrene	26.591	276	1190629	20.234	ng/ul	99
95) Dibenzo(a,h)anthracene	26.603	278	1005963	20.279	ng/ul	97
96) Benzo(g,h,i)perylene	27.385	276	957219	20.375	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BMCCC180.D
 Acq On : 21 Dec 2022 20:03
 Operator : CG/JU
 Sample : SSTCCCC020
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :

Quant Time: Dec 21 23:34:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.213 ng/uL
RT: 3.428 min Scan# 41
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 88 Resp: 31289
Ion Ratio Lower Upper
88 100
43 41.1 32.8 49.2
58 84.6 64.6 97.0



#5
Pyridine
Concen: 16.045 ng/uL
RT: 3.846 min Scan# 112
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 79 Resp: 229628
Ion Ratio Lower Upper
79 100
52 107.3 78.1 117.1
51 47.6 37.2 55.8









#6
Benzaldehyde
Concen: 17.958 ng/ul
RT: 7.181 min Scan# 679
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 118969
Ion Ratio Lower Upper
77 100
105 102.9 82.5 123.7
106 99.4 81.8 122.6



#8
Phenol
Concen: 17.552 ng/ul
RT: 7.228 min Scan# 687
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 94 Resp: 311800
Ion Ratio Lower Upper
94 100
65 30.1 22.2 33.4
66 39.5 27.3 40.9









#10
Bis(2-Chloroethyl)ether
Concen: 17.281 ng/ul
RT: 7.463 min Scan# 727
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 93 Resp: 252091
Ion Ratio Lower Upper
93 100
63 78.8 66.3 99.5
95 34.4 25.7 38.5



#12
2-Chlorophenol
Concen: 19.225 ng/ul
RT: 7.604 min Scan# 751
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:128 Resp: 285694
Ion Ratio Lower Upper
128 100
64 43.1 35.2 52.8
130 33.0 25.9 38.9









#13
2-Methylphenol
Concen: 18.914 ng/ul
RT: 8.487 min Scan# 901
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 257130
Ion Ratio Lower Upper
108 100
107 88.2 71.9 107.9



#14
2,2'-oxybis(1-Chloropropane)
Concen: 16.543 ng/ul
RT: 8.575 min Scan# 916
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 45 Resp: 435752
Ion Ratio Lower Upper
45 100
77 11.1 9.0 13.4
79 8.5 6.2 9.4









#16
Acetophenone
Concen: 19.234 ng/ul
RT: 8.875 min Scan# 967
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:105 Resp: 418250
Ion Ratio Lower Upper
105 100
77 72.2 59.7 89.5
51 32.0 27.2 40.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.844 ng/ul
RT: 8.857 min Scan# 964
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 70 Resp: 193032
Ion Ratio Lower Upper
70 100
42 75.4 62.3 93.5
101 9.7 8.0 12.0
130 23.3 17.6 26.4









#18
4-Methylphenol
Concen: 18.967 ng/ul
RT: 8.822 min Scan# 958
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:108 Resp: 278931
Ion Ratio Lower Upper
108 100
107 110.3 88.1 132.1



#19
Hexachloroethane
Concen: 19.083 ng/ul
RT: 9.128 min Scan# 1010
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:117 Resp: 113370
Ion Ratio Lower Upper
117 100
201 112.5 87.3 130.9
199 70.5 50.9 76.3









#22
Nitrobenzene
Concen: 19.163 ng/ul
RT: 9.257 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion: 77 Resp: 302484
Ion Ratio Lower Upper
77 100
123 49.5 37.8 56.6
65 14.1 11.1 16.7



#23
Isophorone
Concen: 19.482 ng/ul
RT: 9.787 min Scan# 1122
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 82 Resp: 563525
Ion Ratio Lower Upper
82 100
95 7.1 5.2 7.8
138 19.5 14.9 22.3









#25
2-Nitrophenol
Concen: 20.736 ng/ul
RT: 9.969 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:139 Resp: 167134
Ion Ratio Lower Upper
139 100
65 47.2 36.2 54.4
109 30.3 19.7 29.5#



#26
2,4-Dimethylphenol
Concen: 20.637 ng/ul
RT: 10.028 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:107 Resp: 321664
Ion Ratio Lower Upper
107 100
121 50.0 42.0 63.0
122 87.6 71.4 107.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.461 ng/ul
 RT: 10.263 min Scan# 1203
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion: 93 Resp: 354974
 Ion Ratio Lower Upper
 93 100
 95 32.5 25.5 38.3
 123 12.5 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 20.981 ng/ul
 RT: 10.504 min Scan# 1244
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03

Tgt Ion: 162 Resp: 300598
 Ion Ratio Lower Upper
 162 100
 164 66.4 51.4 77.2
 98 32.9 26.1 39.1









#30
Naphthalene
Concen: 19.377 ng/ul
RT: 10.910 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:128 Resp: 968887
Ion Ratio Lower Upper
128 100
129 10.7 9.0 13.4
127 13.0 10.4 15.6



#32
4-Chloroaniline
Concen: 20.194 ng/ul
RT: 11.022 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:127 Resp: 405920
Ion Ratio Lower Upper
127 100
129 32.4 25.3 37.9









#33
Hexachlorobutadiene
Concen: 22.113 ng/ul
RT: 11.181 min Scan# 11
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:225 Resp: 208413
Ion Ratio Lower Upper
225 100
223 64.2 50.8 76.2
227 64.0 50.5 75.7



#34
Caprolactam
Concen: 22.034 ng/ul
RT: 11.810 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:113 Resp: 80976
Ion Ratio Lower Upper
113 100
55 190.4 172.2 258.2
56 140.8 120.7 181.1









#35
4-Chloro-3-methylphenol
Concen: 21.720 ng/ul
RT: 12.139 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:107 Resp: 283256
Ion Ratio Lower Upper
107 100
144 28.6 23.4 35.0
142 84.6 71.0 106.4



#36
2-Methylnaphthalene
Concen: 20.002 ng/ul
RT: 12.510 min Scan# 1585
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:142 Resp: 648685
Ion Ratio Lower Upper
142 100
141 87.9 69.4 104.0





#37

1-Methylnaphthalene

Concen: 19.939 ng/ul

RT: 12.728 min Scan# 1

Delta R.T. 0.000 min


Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:142 Resp: 661207

Ion Ratio Lower Upper

142 100

141 93.2 72.4 108.6

116 3.9 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.728 ng/ul

RT: 12.869 min Scan# 1646

Delta R.T. -0.006 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Tgt Ion:216 Resp: 369511




Ion Ratio Lower Upper


216 100

214 78.8 64.2 96.4

179 21.1 15.4 23.2

108 16.5 12.7 19.1









#40
Hexachlorocyclopentadiene
Concen: 5.871 ng/ul
RT: 12.845 min Scan# 1642
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:237 Resp: 62909
Ion Ratio Lower Upper
237 100
235 62.8 50.4 75.6
272 13.2 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 21.131 ng/ul
RT: 13.116 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:196 Resp: 241077
Ion Ratio Lower Upper
196 100
198 98.8 75.0 112.6
200 31.6 24.5 36.7









#42
2,4,5-Trichlorophenol
Concen: 21.935 ng/ul
RT: 13.186 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:196 Resp: 260994
Ion Ratio Lower Upper
196 100
198 96.8 76.5 114.7
200 31.1 24.9 37.3



#43
1,1'-Biphenyl
Concen: 19.174 ng/ul
RT: 13.510 min Scan# 1755
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:154 Resp: 883519
Ion Ratio Lower Upper
154 100
153 39.7 31.2 46.8
76 9.2 9.3 13.9#









#44
2-Chloronaphthalene
Concen: 19.345 ng/ul
RT: 13.551 min Scan# 1762
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:162 Resp: 687642
Ion Ratio Lower Upper
162 100
164 32.7 27.3 40.9
127 39.6 29.4 44.0



#45
2-Nitroaniline
Concen: 20.707 ng/ul
RT: 13.763 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion: 65 Resp: 177353
Ion Ratio Lower Upper
65 100
92 68.1 53.5 80.3
138 117.6 91.3 136.9









#47
Dimethylphthalate
Concen: 20.570 ng/ul
RT: 14.127 min Scan# 1860
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:163 Resp: 829399
Ion Ratio Lower Upper
163 100
194 3.7 3.0 4.6
164 9.7 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.056 ng/ul
RT: 14.251 min Scan# 1881
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:165 Resp: 164915
Ion Ratio Lower Upper
165 100
89 57.8 41.1 61.7
121 21.9 17.4 26.2









#50
Acenaphthylene
Concen: 19.801 ng/ul
RT: 14.392 min Scan# 1905
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:152 Resp: 1087543
Ion Ratio Lower Upper
152 100
151 19.0 15.3 22.9
153 12.8 10.2 15.4



#51
3-Nitroaniline
Concen: 21.647 ng/ul
RT: 14.580 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:138 Resp: 169151
Ion Ratio Lower Upper
138 100
108 9.6 7.9 11.9
92 103.8 83.4 125.0









#52
Acenaphthene
Concen: 19.452 ng/ul
RT: 14.733 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:153 Resp: 738954
Ion Ratio Lower Upper
153 100
152 47.5 38.2 57.4
154 87.1 71.4 107.0



#53
2,4-Dinitrophenol
Concen: 14.727 ng/ul
RT: 14.786 min Scan# 1972
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:184 Resp: 56600
Ion Ratio Lower Upper
184 100
63 61.4 46.2 69.4
154 67.6 45.5 68.3









#55
4-Nitrophenol
Concen: 24.863 ng/ul
RT: 14.886 min Scan# 1989
Delta R.T. 0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:109 Resp: 114203
Ion Ratio Lower Upper
109 100
139 129.9 119.2 178.8
65 119.7 108.7 163.1



#56
Dibenzofuran
Concen: 19.830 ng/ul
RT: 15.069 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:168 Resp: 1028480
Ion Ratio Lower Upper
168 100
139 37.7 28.1 42.1





#57

2,4-Dinitrotoluene

Concen: 22.295 ng/ul

RT: 15.033 min Scan# 2014

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:165 Resp: 236473

Ion Ratio Lower Upper

165 100

63 43.4 34.9 52.3

182 2.1 2.2 3.2#



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.106 ng/ul

RT: 15.292 min Scan# 2058

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Tgt Ion:232 Resp: 221256




Ion Ratio Lower Upper


232 100

131 45.3 32.8 49.2

130 2.3 2.0 3.0

166 30.8 23.2 34.8









#59
Diethylphthalate
Concen: 21.148 ng/ul
RT: 15.480 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 849497
Ion Ratio Lower Upper
149 100
177 22.3 18.5 27.7
150 12.0 9.5 14.3



#61
Fluorene
Concen: 20.257 ng/ul
RT: 15.716 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:166 Resp: 856804
Ion Ratio Lower Upper
166 100
165 98.5 77.2 115.8
167 13.3 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 21.111 ng/ul
RT: 15.704 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:204 Resp: 436385
Ion Ratio Lower Upper
204 100
206 32.3 25.5 38.3
141 58.0 42.8 64.2



#63
4-Nitroaniline
Concen: 24.452 ng/ul
RT: 15.739 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:138 Resp: 173251
Ion Ratio Lower Upper
138 100
92 48.8 36.1 54.1
108 81.0 51.6 77.4









#66
4,6-Dinitro-2-methylphenol
Concen: 14.403 ng/ul
RT: 15.798 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:198 Resp: 90701
Ion Ratio Lower Upper
198 100
182 3.0 3.8 5.8#
77 21.9 19.3 28.9



#67
N-Nitrosodiphenylamine
Concen: 18.913 ng/ul
RT: 15.916 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:169 Resp: 694133
Ion Ratio Lower Upper
169 100
168 66.8 53.8 80.8
167 37.2 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 19.413 ng/ul
RT: 16.592 min Scan# 21
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:248 Resp: 258917
Ion Ratio Lower Upper
248 100
250 99.2 78.7 118.1
141 72.3 52.7 79.1



#69
Hexachlorobenzene
Concen: 19.568 ng/ul
RT: 16.715 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:284 Resp: 311533
Ion Ratio Lower Upper
284 100
142 30.8 23.1 34.7
249 30.2 23.3 34.9





#70

Atrazine

Concen: 20.752 ng/ul

RT: 16.868 min Scan# 21

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:200 Resp: 260280

Ion Ratio Lower Upper

200 100

173 24.7 20.5 30.7

215 47.2 40.1 60.1



#71

Pentachlorophenol

Concen: 20.399 ng/ul

RT: 17.057 min Scan# 2358

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03




Tgt Ion:266 Resp: 180561


Ion Ratio Lower Upper

266 100

264 65.0 50.1 75.1

268 63.3 50.6 75.8









#72
 Phenanthrene
 Concen: 18.977 ng/ul
 RT: 17.451 min Scan# 2425
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:178 Resp: 1260906
 Ion Ratio Lower Upper
 178 100
 179 15.3 12.2 18.2
 176 19.0 15.4 23.0



#74
 Anthracene
 Concen: 19.209 ng/ul
 RT: 17.545 min Scan# 2441
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03

Tgt Ion:178 Resp: 1297066
 Ion Ratio Lower Upper
 178 100
 179 15.3 11.8 17.8
 176 18.8 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 18.235 ng/uL

RT: 13.475 min Scan# 1

Delta R.T. -0.006 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:216 Resp: 358681

Ion Ratio Lower Upper

216 100

214 76.9 62.6 93.8

218 47.9 38.3 57.5



#76

Pentachlorobenzene

Concen: 18.892 ng/uL

RT: 14.986 min Scan# 2006

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03




Tgt Ion:250 Resp: 368111


Ion Ratio Lower Upper

250 100

248 61.1 51.0 76.4

252 63.1 51.8 77.6









#77
 Carbazole
 Concen: 19.114 ng/ul
 RT: 17.815 min Scan# 2487
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03




Instrument :
 BNA_M
 ClientSampleId :


Tgt Ion:167 Resp: 1118215
 Ion Ratio Lower Upper
 167 100
 166 21.6 16.7 25.1
 139 12.6 9.8 14.6



#78
 Di-n-butylphthalate
 Concen: 21.745 ng/ul
 RT: 18.362 min Scan# 2580
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03

Tgt Ion:149 Resp: 1512076
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.3 10.9
 104 4.6 3.4 5.2









#80
Fluoranthene
Concen: 22.196 ng/ul
RT: 19.456 min Scan# 21
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:202 Resp: 1345911
Ion Ratio Lower Upper
202 100
101 7.8 8.6 13.0#
100 6.3 6.6 10.0#



#82
Pyrene
Concen: 21.955 ng/ul
RT: 19.821 min Scan# 2828
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:202 Resp: 1385519
Ion Ratio Lower Upper
202 100
101 9.2 10.0 15.0#
100 8.1 7.9 11.9









#83
Butylbenzylphthalate
Concen: 25.759 ng/ul
RT: 20.698 min Scan# 2977
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:149 Resp: 604463
Ion Ratio Lower Upper
149 100
91 63.7 54.5 81.7
206 19.0 15.8 23.8



#84
3,3'-Dichlorobenzidine
Concen: 15.341 ng/ul
RT: 21.486 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:252 Resp: 259639
Ion Ratio Lower Upper
252 100
254 65.8 53.2 79.8
126 10.6 7.8 11.8









#85
Benzo(a)anthracene
Concen: 20.110 ng/ul
RT: 21.556 min Scan# 3123
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:228 Resp: 1157995
Ion Ratio Lower Upper
228 100
229 20.1 15.5 23.3
226 27.4 21.8 32.8



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.485 ng/ul
RT: 21.468 min Scan# 3108
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:149 Resp: 904154
Ion Ratio Lower Upper
149 100
167 27.3 21.5 32.3
279 3.9 3.2 4.8





#87

Chrysene

Concen: 19.868 ng/ul

RT: 21.615 min Scan# 3133

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :





Tgt Ion:228 Resp: 1073390

Ion Ratio Lower Upper

228 100

226 30.1 24.3 36.5

229 19.7 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 25.501 ng/ul




RT: 22.403 min Scan# 3267


Delta R.T. -0.006 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Tgt Ion:149 Resp: 1369375









#90
Benzo(b)fluoranthene
Concen: 20.144 ng/ul
RT: 23.274 min Scan# 3415
Delta R.T. 0.000 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03




Instrument :
BNA_M
ClientSampleId :


Tgt Ion:252 Resp: 1046634
Ion Ratio Lower Upper
252 100
253 22.4 17.6 26.4
125 8.9 8.0 12.0



#91
Benzo(k)fluoranthene
Concen: 20.346 ng/ul
RT: 23.321 min Scan# 3423
Delta R.T. -0.006 min
Lab File: BMCCC180.D
Acq: 21 Dec 2022 20:03

Tgt Ion:252 Resp: 1024517
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.2
125 9.3 8.1 12.1





#93

Benzo(a)pyrene

Concen: 20.011 ng/ul

RT: 23.915 min Scan# 3

Delta R.T. -0.006 min


Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03

Instrument :

BNA_M

ClientSampleId :






Tgt Ion:252 Resp: 914304

Ion Ratio Lower Upper

252 100

253 22.9 17.4 26.0

125 9.5 9.4 14.0



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.234 ng/ul

RT: 26.591 min Scan# 3979

Delta R.T. 0.000 min

Lab File: BMCCC180.D

Acq: 21 Dec 2022 20:03




Tgt Ion:276 Resp: 1190629


Ion Ratio Lower Upper

276 100

138 23.2 19.3 28.9

277 25.2 19.8 29.8









#95
 Dibenzo(a,h)anthracene
 Concen: 20.279 ng/ul
 RT: 26.603 min Scan# 3980
 Delta R.T. 0.000 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03

Instrument :
 BNA_M
 ClientSampleId :

Tgt Ion:278 Resp: 1005963
 Ion Ratio Lower Upper
 278 100
 139 13.8 12.4 18.6
 279 23.7 19.8 29.6



#96
 Benzo(g,h,i)perylene
 Concen: 20.375 ng/ul
 RT: 27.385 min Scan# 4114
 Delta R.T. -0.006 min
 Lab File: BMCCC180.D
 Acq: 21 Dec 2022 20:03

Tgt Ion:276 Resp: 957219
 Ion Ratio Lower Upper
 276 100
 138 20.0 17.1 25.7
 277 23.0 18.2 27.4

