

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BM038156.D
 Acq On : 21 Dec 2022 04:55
 Operator : CG/JU
 Sample : N6008-06
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DBYB3

Quant Time: Dec 21 05:26:40 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.039	152	177728	20.000	ng/u1	0.00
20) Naphthalene-d8	10.857	136	760853	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.668	164	472002	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.409	188	1015733	20.000	ng/u1	0.00
79) Chrysene-d12	21.574	240	809942	20.000	ng/u1	0.00
88) Perylene-d12	24.021	264	744694	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	20548	5.249	ng/uL	0.00
4) Pyridine-d5	3.834	84	179292	15.441	ng/u1	0.00
7) Phenol-d5	7.198	99	334218	23.142	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	228653	26.000	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	317437	26.794	ng/u1	0.00
15) 4-Methylphenol-d8	8.751	113	158932	14.098	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	177165	29.409	ng/u1	0.00
24) 2-Nitrophenol-d4	9.939	143	203843	32.289	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	348073	29.028	ng/u1	0.00
31) 4-Chloroaniline-d4	10.998	131	333516	20.150	ng/u1	0.00
46) Dimethylphthalate-d6	14.080	166	1068182	31.743	ng/u1	0.00
49) Acenaphthylene-d8	14.363	160	1251243	30.424	ng/u1	0.00
54) 4-Nitrophenol-d4	14.868	143	147756	25.764	ng/u1	0.00
60) Fluorene-d10	15.657	176	968303	32.298	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	132915	24.475	ng/u1	0.00
73) Anthracene-d10	17.504	188	1521446	32.094	ng/u1	0.00
81) Pyrene-d10	19.792	212	1736540	35.692	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.868	264	1236923	33.254	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
Data File : BM038156.D
Acq On : 21 Dec 2022 04:55
Operator : CG/JU
Sample : N6008-06
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DBYB3

Quant Time: Dec 21 05:26:40 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 00:11:56 2022
Response via : Initial Calibration

