

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
 Data File : BM038184.D
 Acq On : 21 Dec 2022 23:05
 Operator : CG/JU
 Sample : N6032-12
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COLB4

Quant Time: Dec 21 23:36:27 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 00:11:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	172521	20.000	ng/u1	0.00
20) Naphthalene-d8	10.857	136	710530	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.669	164	434308	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.404	188	891724	20.000	ng/u1	0.00
79) Chrysene-d12	21.568	240	649082	20.000	ng/u1	-0.01
88) Perylene-d12	24.021	264	650952	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	571599	150.436	ng/uL	-0.04
4) Pyridine-d5	3.840	84	53711	4.765	ng/u1	0.00
7) Phenol-d5	7.199	99	76509	5.457	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.363	67	207873	24.351	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	216200	18.799	ng/u1	0.00
15) 4-Methylphenol-d8	8.751	113	136261	12.452	ng/u1	0.00
21) Nitrobenzene-d5	9.210	128	161049	28.628	ng/u1	0.00
24) 2-Nitrophenol-d4	9.934	143	156474	26.541	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.475	165	266747	23.821	ng/u1	0.00
31) 4-Chloroaniline-d4	10.992	131	244615	15.825	ng/u1	-0.01
46) Dimethylphthalate-d6	14.075	166	1009511	32.604	ng/u1	0.00
49) Acenaphthylene-d8	14.363	160	1118641	29.560	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	29021	5.500	ng/u1	0.00
60) Fluorene-d10	15.657	176	880175	31.907	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.774	200	74333	15.591	ng/u1	0.00
73) Anthracene-d10	17.504	188	1448238	34.799	ng/u1	0.00
81) Pyrene-d10	19.786	212	1540380	39.507	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.862	264	1167854	35.919	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122022\
Data File : BM038184.D
Acq On : 21 Dec 2022 23:05
Operator : CG/JU
Sample : N6032-12
Misc :
ALS Vial : 53 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
COLB4

Quant Time: Dec 21 23:36:27 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 00:11:56 2022
Response via : Initial Calibration

