

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122024\
 Data File : BM049094.D
 Acq On : 20 Dec 2024 10:51
 Operator : RC/JU
 Sample : PB165708BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK708

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
 Supervised By :mohammad ahmed 12/24/2024

Quant Time: Dec 20 11:47:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 18 15:33:13 2024
 Response via : Initial Calibration

Supervised By :mohammad
 ahmed
 12/24/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.561	152	5287	0.400	ng/u1	0.00
4) Naphthalene-d8	10.321	136	15023	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.197	164	7942	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.943	188	17004m	0.400	ng/u1	0.00
17) Chrysene-d12	21.149	240	11362m	0.400	ng/u1	0.00
23) Perylene-d12	23.303	264	11143m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.147	96	33258	5.454	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.932	152	5980	0.296	ng/u1	0.00
18) Fluoranthene-d10	18.984	212	11837	0.367	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122024\
Data File : BM049094.D
Acq On : 20 Dec 2024 10:51
Operator : RC/JU
Sample : PB165708BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK708

Quant Time: Dec 20 11:47:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 18 15:33:13 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024
Supervised By :mohammad ahmed 12/24/2024

Supervised By :mohammad
ahmed
12/24/2024

