

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM122024\  
 Data File : BM049138.D  
 Acq On : 21 Dec 2024 15:01  
 Operator : RC/JU  
 Sample : SP6698  
 Misc :  
 ALS Vial : 40 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SP6698

Manual Integrations  
 APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024  
 Supervised By :mohammad ahmed 12/24/2024

Quant Time: Dec 22 23:08:29 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM121824.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Dec 18 15:33:13 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|--------|------|----------|--------|-------|----------|
| -----                       |        |      |          |        |       |          |
| Internal Standards          |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.565  | 152  | 6514     | 0.400  | ng/ul | 0.00     |
| 4) Naphthalene-d8           | 10.321 | 136  | 19671    | 0.400  | ng/ul | 0.00     |
| 9) Acenaphthene-d10         | 14.192 | 164  | 10675    | 0.400  | ng/ul | 0.00     |
| 13) Phenanthrene-d10        | 16.943 | 188  | 21351m   | 0.400  | ng/ul | 0.00     |
| 17) Chrysene-d12            | 21.146 | 240  | 14470m   | 0.400  | ng/ul | 0.00     |
| 23) Perylene-d12            | 23.300 | 264  | 14970m   | 0.400  | ng/ul | 0.00     |
|                             |        |      |          |        |       |          |
| System Monitoring Compounds |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8           | 3.185  | 96   | 109640   | 14.594 | ng/ul | 0.03     |
| 6) 2-Methylnaphthalene-d10  | 11.915 | 152  | 22208    | 0.840  | ng/ul | 0.00     |
| 18) Fluoranthene-d10        | 18.979 | 212  | 36942    | 0.900  | ng/ul | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM122024\  
Data File : BM049138.D  
Acq On : 21 Dec 2024 15:01  
Operator : RC/JU  
Sample : SP6698  
Misc :  
ALS Vial : 40 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SP6698

Quant Time: Dec 22 23:08:29 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM121824.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Dec 18 15:33:13 2024  
Response via : Initial Calibration

Manual Integrations  
APPROVED

Reviewed By :Jagrut Upadhyay 12/23/2024  
Supervised By :mohammad ahmed 12/24/2024

