

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122115\
 Data File : BM003693.D
 Acq On : 22 Dec 2015 01:17
 Operator : UM/SJ
 Sample : G4867-11DL 30X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 G9DA6DL

Quant Time: Dec 22 04:33:31 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM121915.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 22 02:40:18 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	30035	20.00	ng/ul	0.00
18) Naphthalene-d8	10.48	136	122052	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.34	164	66671	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.09	188	143515	20.00	ng/ul	0.00
78) Chrysene-d12	21.29	240	155331	20.00	ng/ul	0.00
86) Perylene-d12	23.54	264	154016	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	0.00	99	0	0.00	ng/ul	
7) Bis-(2-Chloroethyl)ether-d	7.03	67	1216	0.90	ng/ul	0.00
9) 2-Chlorophenol-d4	7.23	132	1387	0.67	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	601	0.27	ng/ul	0.00
19) Nitrobenzene-d5	8.86	128	541	0.61	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	612	0.61	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.11	165	1154	0.63	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	13.76	166	5893	1.07	ng/ul	0.00
47) Acenaphthylene-d8	14.04	160	6973	1.11	ng/ul	0.00
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul	
58) Fluorene-d10	15.34	176	6083	1.33	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
71) Anthracene-d10	17.19	188	8209	1.26	ng/ul	0.00
79) Pyrene-d10	19.49	212	9040	1.33	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.39	264	9370	1.22	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
28) Naphthalene	10.53	128	207409	32.98	ng/ul	100
41) 1,1'-Biphenyl	13.17	154	14958	2.76	ng/ul	100
48) Acenaphthylene	14.07	152	15513	2.26	ng/ul#	94
50) Acenaphthene	14.41	153	38342	8.37	ng/ul	99
59) Fluorene	15.40	166	17175	3.27	ng/ul	98
70) Phenanthrene	17.14	178	29699	3.74	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122115\
Data File : BM003693.D
Acq On : 22 Dec 2015 01:17
Operator : UM/SJ
Sample : G4867-11DL 30X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
G9DA6DL

Quant Time: Dec 22 04:33:31 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM121915.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 22 02:40:18 2015
Response via : Initial Calibration

