

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM122120\
 Data File : BM028239.D
 Acq On : 22 Dec 2020 16:40
 Operator : JU/CG
 Sample : L5144-05
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SVOC-GPC-BLANK

Quant Time: Dec 22 17:46:36 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121520MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Dec 22 16:44:11 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	115073	20.00	ng/ul	0.00
18) Naphthalene-d8	11.03	136	455040	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.82	164	242289	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.54	188	463802	20.00	ng/ul	0.00
78) Chrysene-d12	21.66	240	386557	20.00	ng/ul	0.00
86) Perylene-d12	24.02	264	375607	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL
5) Phenol-d5	0.00	99	0	0.00	ng/ul
7) Bis-(2-Chloroethyl)ether-d	0.00	67	0	0.00	ng/ul
9) 2-Chlorophenol-d4	0.00	132	0	0.00	ng/ul
13) 4-Methylphenol-d8	0.00	113	0	0.00	ng/ul
19) Nitrobenzene-d5	0.00	128	0	0.00	ng/ul
22) 2-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
26) 2,4-Dichlorophenol-d3	0.00	165	0	0.00	ng/ul
29) 4-Chloroaniline-d4	0.00	131	0d	0.00	ng/ul
44) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul
47) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul
52) 4-Nitrophenol-d4	0.00	143	0	0.00	ng/ul
58) Fluorene-d10	0.00	176	0	0.00	ng/ul
63) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul
71) Anthracene-d10	0.00	188	0d	0.00	ng/ul
79) Pyrene-d10	0.00	212	0d	0.00	ng/ul
90) Benzo(a)pyrene-d12	0.00	264	0d	0.00	ng/ul

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM122120\
Data File : BM028239.D
Acq On : 22 Dec 2020 16:40
Operator : JU/CG
Sample : L5144-05
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: Dec 22 17:46:36 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM121520MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Dec 22 16:44:11 2020
Response via : Initial Calibration

