

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122123\
 Data File : BM043517.D
 Acq On : 22 Dec 2023 00:23
 Operator : MA/JU
 Sample : 05912-18
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH3N0

Quant Time: Dec 22 01:16:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 21 21:59:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	191690	20.000	ng/u1	0.00
20) Naphthalene-d8	10.781	136	819079	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.604	164	436160	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	866694	20.000	ng/u1	0.00
79) Chrysene-d12	21.527	240	693464	20.000	ng/u1	0.00
88) Perylene-d12	23.944	264	829831	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.375	96	20548	3.484	ng/uL	0.00
4) Pyridine-d5	3.799	84	146469	8.628	ng/u1	0.00
7) Phenol-d5	7.128	99	92272	4.231	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.310	67	402123	29.118	ng/u1	0.00
11) 2-Chlorophenol-d4	7.498	132	311203	18.691	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	182491	10.785	ng/u1	0.00
21) Nitrobenzene-d5	9.145	128	212346	26.795	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	194261	23.725	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	324028	22.417	ng/u1	0.00
31) 4-Chloroaniline-d4	10.928	131	452702	21.288	ng/u1	0.00
46) Dimethylphthalate-d6	14.027	166	1207368	30.750	ng/u1	0.00
49) Acenaphthylene-d8	14.298	160	1398364	30.023	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	25879	3.482	ng/u1	0.00
60) Fluorene-d10	15.598	176	1045955	32.215	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.715	200	122587	21.386	ng/u1	0.00
73) Anthracene-d10	17.451	188	1689490	34.759	ng/u1	0.00
81) Pyrene-d10	19.739	212	1784969	32.603	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.791	264	1810283	35.351	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122123\
Data File : BM043517.D
Acq On : 22 Dec 2023 00:23
Operator : MA/JU
Sample : 05912-18
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH3N0

Quant Time: Dec 22 01:16:27 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM121523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 21 21:59:32 2023
Response via : Initial Calibration

