


Data Path : Z:\HPCHEM1\BNA M\DATA\BM122315\  
 Data File : BM003724.D  
 Acq On : 23 Dec 2015 00:51  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_M\METHODS\8270-BM120915.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 09 18:52:47 2015



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2450

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 27.7      | 13451   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 212     | PASS             |
| 69          | 198          | 0.00         | 100          | 33.3      | 16135   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 47.1      | 22813   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 48480   | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 3357    | PASS             |
| 275         | 198          | 10           | 60           | 24.6      | 11933   | PASS             |
| 365         | 198          | 1            | 100          | 2.9       | 1415    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.3      | 5774    | PASS             |
| 442         | 198          | 50           | 100          | 77.7      | 37680   | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 7172    | PASS             |