

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
 Data File : BM003754.D
 Acq On : 23 Dec 2015 23:11
 Operator : UM/SJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM122415

Quant Time: Dec 24 10:29:30 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.533	0.453	15.0	101	0.00
3	n-Nitrosodimethylamine	0.532	0.478	10.2	100	0.00
4 S	2-Fluorophenol	1.181	1.035	12.4	99	0.00
5 S	Phenol-d6	1.447	1.237	14.5	100	0.00
6	bis(2-Chloroethyl)ether	1.286	1.149	10.7	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.265	0.231	12.8	101	0.00
9	Nitrobenzene	0.301	0.263	12.6	101	0.00
10	Naphthalene	1.142	1.057	7.4	102	0.00
11	Hexachlorobutadiene	0.176	0.163	7.4	101	0.00
12	2-Methylnaphthalene	0.762	0.693	9.1	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.145	0.114	21.4	103	0.00
15 S	2-Fluorobiphenyl	1.530	1.371	10.4	102	0.00
16	Acenaphthylene	2.275	1.928	15.3	102	0.00
17	Acenaphthene	1.479	1.321	10.7	102	0.00
18	Fluorene	1.754	1.519	13.4	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.03
20	4-Bromophenyl-phenylether	0.281	0.273	2.8	103	0.00
21	Hexachlorobenzene	0.271	0.266	1.8	106	0.00
22	Pentachlorophenol	0.077	0.054	29.9#	108	0.00
23	Phenanthrene	1.689	1.633	3.3	106	0.00
24	Anthracene	1.235	1.119	9.4	106	0.00
25	Fluoranthene	1.605	1.486	7.4	109	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
27	Benzydine	0.268	0.314	-17.2	122	0.00
28	Pvrene	1.751	1.460	16.6	108	0.00
29 S	Terphenyl-d14	0.890	0.742	16.6	111	0.00
30	Benzo(a)anthracene	1.562	1.278	18.2	119	0.00
31	3,3'-Dichlorobenzidine	0.431	0.384	10.9	116	0.00
32	Chrysene	1.561	1.378	11.7	110	0.00
33	Bis(2-ethylhexyl)phthalate	0.984	0.718	27.0#	106	0.00
34	Indeno(1,2,3-cd)pyrene	1.111	0.967	13.0	115	0.01
35 I	Perylene-d12	1.000	1.000	0.0	115	0.01
36	Benzo(b)fluoranthene	1.634	1.445	11.6	113	0.00
37	Benzo(k)fluoranthene	1.507	1.371	9.0	126	0.00
38 C	Benzo(a)pyrene	1.399	1.238	11.5	120	0.00
39	Dibenzo(a,h)anthracene	1.097	0.963	12.2	115	0.00
40	Benzo(a,h,i)perylene	1.148	1.001	12.8	113	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
Data File : BM003754.D
Acq On : 23 Dec 2015 23:11
Operator : UM/SJ
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM122415

Quant Time: Dec 24 10:29:30 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				