

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
 Data File : BM003754.D
 Acq On : 23 Dec 2015 23:11
 Operator : UM/SJ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM122415

Quant Time: Dec 24 10:29:30 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2	1,4-Dioxane	1.000	0.942	5.8	101	0.00
3	n-Nitrosodimethylamine	1.000	0.897	10.3	100	0.00
4 S	2-Fluorophenol	1.000	0.876	12.4	99	0.00
5 S	Phenol-d6	1.000	0.855	14.5	100	0.00
6	bis(2-Chloroethyl)ether	1.000	0.893	10.7	100	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	97	0.00
8 S	Nitrobenzene-d5	1.000	0.870	13.0	101	0.00
9	Nitrobenzene	1.000	0.875	12.5	101	0.00
10	Naphthalene	1.000	0.926	7.4	102	0.00
11	Hexachlorobutadiene	1.000	0.926	7.4	101	0.00
12	2-Methylnaphthalene	1.000	0.909	9.1	101	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	1.000	0.860	14.0	103	0.00
15 S	2-Fluorobiphenyl	1.000	0.896	10.4	102	0.00
16	Acenaphthylene	1.000	0.847	15.3	102	0.00
17	Acenaphthene	1.000	0.947	5.3	102	0.00
18	Fluorene	1.000	0.900	10.0	103	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	102	0.03
20	4-Bromophenyl-phenylether	1.000	1.304	-30.4#	103	0.00
21	Hexachlorobenzene	1.000	1.329	-32.9#	106	0.00
22	Pentachlorophenol	1.000	0.855	14.5	108	0.00
23	Phenanthrene	1.000	1.216	-21.6	106	0.00
24	Anthracene	1.000	0.954	4.6	106	0.00
25	Fluoranthene	1.000	1.006	-0.6	109	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
27	Benزيدine	1.000	1.022	-2.2	122	0.00
28	Pvrene	1.000	0.848	15.2	108	0.00
29 S	Terphenyl-d14	1.000	0.884	11.6	111	0.00
30	Benzo(a)anthracene	1.000	0.848	15.2	119	0.00
31	3,3'-Dichlorobenzidine	1.000	0.960	4.0	116	0.00
32	Chrysene	1.000	0.897	10.3	110	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	0.823	17.7	106	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	0.918	8.2	115	0.01
35 I	Perylene-d12	5.000	5.000	0.0	115	0.01
36	Benzo(b)fluoranthene	1.000	0.884	11.6	113	0.00
37	Benzo(k)fluoranthene	1.000	0.910	9.0	126	0.00
38 C	Benzo(a)pyrene	1.000	0.885	11.5	120	0.00
39	Dibenzo(a,h)anthracene	1.000	0.878	12.2	115	0.00
40	Benzo(a,h,i)perylene	1.000	0.872	12.8	113	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
Data File : BM003754.D
Acq On : 23 Dec 2015 23:11
Operator : UM/SJ
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM122415

Quant Time: Dec 24 10:29:30 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0