

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
 Data File : BM003772.D
 Acq On : 24 Dec 2015 10:49
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 24 16:10:58 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.533	0.441	17.3	89	0.00
3	n-Nitrosodimethylamine	0.532	0.441	17.1	84	0.00
4 S	2-Fluorophenol	1.181	0.979	17.1	86	0.02
5 S	Phenol-d6	1.447	1.168	19.3	86	0.00
6	bis(2-Chloroethyl)ether	1.286	1.069	16.9	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.265	0.214	19.2	85	0.00
9	Nitrobenzene	0.301	0.241	19.9	84	0.00
10	Naphthalene	1.142	1.005	12.0	89	0.00
11	Hexachlorobutadiene	0.176	0.156	11.4	89	0.00
12	2-Methylnaphthalene	0.762	0.661	13.3	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.145	0.110	24.1	91	0.00
15 S	2-Fluorobiphenyl	1.530	1.303	14.8	89	0.00
16	Acenaphthylene	2.275	1.819	20.0	88	0.00
17	Acenaphthene	1.479	1.281	13.4	91	0.00
18	Fluorene	1.754	1.453	17.2	90	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.03
20	4-Bromophenyl-phenylether	0.281	0.246	12.5	88	0.00
21	Hexachlorobenzene	0.271	0.246	9.2	93	0.00
22	Pentachlorophenol	0.077	0.056	27.3#	107	0.00
23	Phenanthrene	1.689	1.554	8.0	96	0.00
24	Anthracene	1.235	1.058	14.3	95	0.00
25	Fluoranthene	1.605	1.409	12.2	98	0.02
26 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
27	Benzydine	0.268	0.264	1.5	105	0.00
28	Pvrene	1.751	1.331	24.0	101	0.00
29 S	Terphenyl-d14	0.890	0.672	24.5	103	0.00
30	Benzo(a)anthracene	1.562	1.215	22.2	116	0.00
31	3,3'-Dichlorobenzidine	0.431	0.349	19.0	108	0.00
32	Chrysene	1.561	1.310	16.1	107	0.00
33	Bis(2-ethylhexyl)phthalate	0.984	0.664	32.5#	100	0.00
34	Indeno(1,2,3-cd)pyrene	1.111	1.025	7.7	125	0.01
35 I	Perylene-d12	1.000	1.000	0.0	125	0.01
36	Benzo(b)fluoranthene	1.634	1.367	16.3	116	0.00
37	Benzo(k)fluoranthene	1.507	1.275	15.4	127	0.00
38 C	Benzo(a)pyrene	1.399	1.170	16.4	123	0.00
39	Dibenzo(a,h)anthracene	1.097	0.963	12.2	125	0.00
40	Benzo(a,h,i)perylene	1.148	1.007	12.3	124	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
Data File : BM003772.D
Acq On : 24 Dec 2015 10:49
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 24 16:10:58 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				