

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
 Data File : BM003774.D
 Acq On : 24 Dec 2015 12:39
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 24 16:13:14 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.533	0.442	17.1	89	0.00
3	n-Nitrosodimethylamine	0.532	0.450	15.4	85	0.00
4 S	2-Fluorophenol	1.181	0.992	16.0	86	0.02
5 S	Phenol-d6	1.447	1.188	17.9	86	0.00
6	bis(2-Chloroethyl)ether	1.286	1.077	16.3	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.265	0.214	19.2	85	0.00
9	Nitrobenzene	0.301	0.241	19.9	84	0.00
10	Naphthalene	1.142	1.010	11.6	89	0.00
11	Hexachlorobutadiene	0.176	0.155	11.9	88	0.00
12	2-Methylnaphthalene	0.762	0.664	12.9	89	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
14 S	2,4,6-Tribromophenol	0.145	0.106	26.9#	88	0.00
15 S	2-Fluorobiphenyl	1.530	1.291	15.6	90	0.00
16	Acenaphthylene	2.275	1.799	20.9	89	0.00
17	Acenaphthene	1.479	1.237	16.4	89	0.00
18	Fluorene	1.754	1.422	18.9	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.281	0.256	8.9	90	0.00
21	Hexachlorobenzene	0.271	0.251	7.4	94	0.00
22	Pentachlorophenol	0.077	0.060	22.1	114	0.00
23	Phenanthrene	1.689	1.576	6.7	96	0.00
24	Anthracene	1.235	1.051	14.9	93	0.00
25	Fluoranthene	1.605	1.430	10.9	98	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
27	Benzydine	0.268	0.073	72.8#	27#	0.00
28	Pvrene	1.751	1.407	19.6	101	0.00
29 S	Terphenyl-d14	0.890	0.709	20.3	103	0.00
30	Benzo(a)anthracene	1.562	1.215	22.2	110	0.00
31	3,3'-Dichlorobenzidine	0.431	0.313	27.4#	92	0.00
32	Chrysene	1.561	1.408	9.8	108	0.00
33	Bis(2-ethylhexyl)phthalate	0.984	0.761	22.7	108	0.00
34	Indeno(1,2,3-cd)pyrene	1.111	1.091	1.8	126	0.01
35 I	Perylene-d12	1.000	1.000	0.0	126	0.00
36	Benzo(b)fluoranthene	1.634	1.404	14.1	120	0.00
37	Benzo(k)fluoranthene	1.507	1.244	17.5	125	0.00
38 C	Benzo(a)pyrene	1.399	1.180	15.7	125	0.00
39	Dibenzo(a,h)anthracene	1.097	0.965	12.0	126	0.00
40	Benzo(a,h,i)perylene	1.148	1.006	12.4	124	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122415\
Data File : BM003774.D
Acq On : 24 Dec 2015 12:39
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 24 16:13:14 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0