

Data Path : Z:\HPCHEM1\BNA_M\Data\BM122416\
 Data File : BM008619.D
 Acq On : 24 Dec 2016 06:23
 Operator : UM/SJ
 Sample : H5995-19
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA885

Quant Time: Dec 25 00:46:19 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 24 04:56:03 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.67	152	550	0.40	ng/ul	-0.08
2) Naphthalene-d8	10.43	136	1687	0.40	ng/ul	-0.07
6) Acenaphthene-d10	14.28	164	963	0.40	ng/ul	-0.03
10) Phenanthrene-d10	17.14	188	90943	0.40	ng/ul	0.05
16) Chrysene-d12	21.27	240	1866	0.40	ng/ul	0.00
20) Perylene-d12	23.48	264	2685	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.07	152	477	0.18	ng/ul	-0.03
14) Fluoranthene-d10	19.08	212	1044	0.00	ng/ul	-0.02
Target Compounds						
9) Fluorene	15.29	166	217	0.06	ng/ul#	41
17) Pyrene	19.45	202	288	0.04	ng/ul#	1
23) Benzo(a)pyrene	23.33	252	537	0.05	ng/ul#	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\Data\BM122416\
Data File : BM008619.D
Acq On : 24 Dec 2016 06:23
Operator : UM/SJ
Sample : H5995-19
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA885

Quant Time: Dec 25 00:46:19 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 24 04:56:03 2016
Response via : Initial Calibration

