

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122416\
 Data File : BM008658.D
 Acq On : 25 Dec 2016 06:38
 Operator : UM/SJ
 Sample : H6040-02DL 10X
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA8T7DL

Manual Integrations
 APPROVED

sohil
 12/26/2016 6:28:23 PM

Quant Time: Dec 26 15:45:13 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 26 00:55:42 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.68	152	379	0.40	ng/ul	-0.05
2) Naphthalene-d8	10.44	136	1186	0.40	ng/ul	-0.04
6) Acenaphthene-d10	14.28	164	701	0.40	ng/ul	-0.03
10) Phenanthrene-d10	17.04	188	1925m	0.40	ng/ul	-0.03
16) Chrysene-d12	21.25	240	1329	0.40	ng/ul	-0.01
20) Perylene-d12	23.48	264	1454	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.03	152	44	0.02	ng/ul	-0.05
14) Fluoranthene-d10	19.08	212	280m	0.05	ng/ul	-0.01
Target Compounds					Ovalue	
3) Naphthalene	10.49	128	785	0.24	ng/ul#	74
5) 2-Methylnaphthalene	12.10	142	1333	0.59	ng/ul	99
7) Acenaphthylene	14.01	152	289	0.09	ng/ul#	38
8) Acenaphthene	14.35	153	252	0.11	ng/ul#	88
9) Fluorene	15.34	166	611	0.22	ng/ul#	63
12) Phenanthrene	17.07	178	2518m	0.43	ng/ul	
13) Anthracene	17.18	178	692m	0.12	ng/ul	
15) Fluoranthene	19.11	202	1552	0.22	ng/ul	80
17) Pyrene	19.48	202	2611	0.51	ng/ul#	90
18) Benzo(a)anthracene	21.23	228	945	0.22	ng/ul#	58
19) Chrysene	21.28	228	1107	0.23	ng/ul#	62
21) Benzo(b)fluoranthene	22.82	252	2375m	0.41	ng/ul	
22) Benzo(k)fluoranthene	22.84	252	795m	0.14	ng/ul	
23) Benzo(a)pyrene	23.38	252	1297	0.24	ng/ul#	82
24) Indeno(1,2,3-cd)pyrene	25.72	276	1278	0.20	ng/ul	94
25) Dibenzo(a,h)anthracene	25.71	278	312	0.06	ng/ul#	1
26) Benzo(a,h,i)perylene	26.41	276	1145	0.21	ng/ul#	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122416\
Data File : BM008658.D
Acq On : 25 Dec 2016 06:38
Operator : UM/SJ
Sample : H6040-02DL 10X
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA8T7DL

Manual Integrations
APPROVED

sohil
12/26/2016 6:28:23 PM

Quant Time: Dec 26 15:45:13 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 26 00:55:42 2016
Response via : Initial Calibration

