

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
 Data File : BM043602.D
 Acq On : 27 Dec 2023 11:06
 Operator : MA/JU
 Sample : 06020-04
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH3P2

Quant Time: Dec 27 21:41:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 26 22:58:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.971	152	5196	0.400	ng/ul	0.00
4) Naphthalene-d8	10.774	136	15591	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.601	164	7085	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.447	188	1436671	0.400	ng/ul #	0.10
17) Chrysene-d12	0.000	240	0	0.000	ng/ul	-21.53
23) Perylene-d12	23.936	264	12961	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	31644	4.820	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.368	152	7301	0.330	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	11820	0.000	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.414	88	567	0.086	ng/ul#	67
12) Fluorene	15.596	166	2768	0.086	ng/ul#	13
26) Benzo(a)pyrene	23.781	252	5569	0.115	ng/ul#	51


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
Data File : BM043602.D
Acq On : 27 Dec 2023 11:06
Operator : MA/JU
Sample : 06020-04
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH3P2

Quant Time: Dec 27 21:41:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 26 22:58:31 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 0.086 ng/ul
RT: 3.414 min Scan# 51
Delta R.T. 0.000 min
Lab File: BM043602.D
Acq: 27 Dec 2023 11:06




Instrument :
BNA_M
ClientSampleId :
BH3P2


Tgt Ion: 88 Resp: 567
Ion Ratio Lower Upper
88 100
43 35.4 27.9 41.9
58 15.5 43.4 65.2#



#12
Fluorene
Concen: 0.086 ng/ul
RT: 15.596 min Scan# 2712
Delta R.T. -0.060 min
Lab File: BM043602.D
Acq: 27 Dec 2023 11:06


Tgt Ion: 166 Resp: 2768
Ion Ratio Lower Upper
166 100
165 8.6 80.3 120.5#
167 34.8 10.8 16.2#





#26
Benzo(a)pyrene
Concen: 0.115 ng/ul
RT: 23.781 min Scan# 4962
Delta R.T. -0.052 min
Lab File: BM043602.D
Acq: 27 Dec 2023 11:06

Instrument :
BNA_M
ClientSampleId :
BH3P2



Tgt Ion:252 Resp: 5569
Ion Ratio Lower Upper
252 100
253 44.5 20.2 30.2#
125 45.2 13.5 20.3#

