

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
 Data File : BM043571.D
 Acq On : 26 Dec 2023 13:53
 Operator : MA/JU
 Sample : 05976-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH7T8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 26 15:02:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 19 23:32:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	5439	0.400	ng/u1	0.00
4) Naphthalene-d8	10.779	136	16274	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.606	164	7418	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.350	188	13933m	0.400	ng/u1	0.00
17) Chrysene-d12	21.530	240	8033m	0.400	ng/u1	0.00
23) Perylene-d12	23.942	264	7803m	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.377	96	26963	3.923	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.368	152	6349	0.275	ng/u1	-0.01
18) Fluoranthene-d10	19.371	212	10324	0.360	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
Data File : BM043571.D
Acq On : 26 Dec 2023 13:53
Operator : MA/JU
Sample : 05976-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BH7T8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2023
Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 26 15:02:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 23:32:46 2023
Response via : Initial Calibration

