

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
 Data File : BM043603.D
 Acq On : 27 Dec 2023 11:42
 Operator : MA/JU
 Sample : 06020-05
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH7W6

Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 27 21:41:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 26 22:58:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.971	152	5030	0.400	ng/ul	0.00
4) Naphthalene-d8	10.774	136	15085	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.601	164	6665	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.346	188	12231m	0.400	ng/ul	0.00
17) Chrysene-d12	21.530	240	6799m	0.400	ng/ul	0.00
23) Perylene-d12	23.936	264	10999	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	29307	4.611	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.368	152	6186	0.289	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	9514	0.392	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.410	88	2234	0.350	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.350 ng/ul
 RT: 3.410 min Scan# 54
 Delta R.T. -0.004 min
 Lab File: BM043603.D
 Acq: 27 Dec 2023 11:42

Instrument :
 BNA_M
 ClientSampleId :
 BH7W6



Tgt Ion: 88 Resp: 2234

Ion	Ratio	Lower	Upper
88	100		
43	41.5	27.9	41.9
58	51.5	43.4	65.2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :mohammad ahmed 12/29/2023

