

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
 Data File : BM043633.D
 Acq On : 28 Dec 2023 09:48
 Operator : MA/JU
 Sample : 06021-09
 Misc :
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BH7Z1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 28 10:17:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 26 22:58:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.967	152	4864	0.400	ng/ul	0.00
4) Naphthalene-d8	10.768	136	14688	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.601	164	6715	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.346	188	12885m	0.400	ng/ul	0.00
17) Chrysene-d12	21.530	240	7531m	0.400	ng/ul	0.00
23) Perylene-d12	23.936	264	6929m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.372	96	29503	4.801	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.363	152	6699	0.322	ng/ul	0.00
18) Fluoranthene-d10	19.371	212	10937	0.406	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.410	88	1613	0.262	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122623\
Data File : BM043633.D
Acq On : 28 Dec 2023 09:48
Operator : MA/JU
Sample : 06021-09
Misc :
ALS Vial : 74 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
BH7Z1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/28/2023
Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 28 10:17:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 26 22:58:31 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.262 ng/ul
 RT: 3.410 min Scan# 54
 Delta R.T. -0.004 min
 Lab File: BM043633.D
 Acq: 28 Dec 2023 09:48

Instrument :
 BNA_M
 ClientSampleId :
 BH7Z1

Tgt Ion: 88 Resp: 161
 Ion Ratio Lower Upper
 88 100
 43 48.6 27.9 41.9
 58 46.8 43.4 65.2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/28/2023
 Supervised By :mohammad ahmed 12/29/2023

