


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122624\
Data File : BM049217.D
Acq On : 26 Dec 2024 09:29
Operator : RC/JU
Sample : DFTPP080
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP080

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Dec 26 11:35:38 2024



AutoFind: Scans 2450, 2451, 2452; Background Corrected with Scan 2442

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	19.2	353223	PASS
68	69	0.00	2	1.3	5268	PASS
69	198	0.00	100	21.2	390876	PASS
70	69	0.00	2	0.4	1394	PASS
127	198	10	80	31.0	570604	PASS
197	198	0.00	2	0.5	8766	PASS
198	198	100	100	100.0	1840469	PASS
199	198	5	9	6.9	126597	PASS
275	198	10	60	27.1	498347	PASS
365	198	1	100	3.6	65643	PASS
441	198	0.01	100	13.1	240469	PASS
442	198	50	100	83.8	1541525	PASS
443	442	15	24	20.1	309632	PASS

m/z	Abundance
35.90	160.0
37.90	293.0
39.00	817.0
40.00	3455.0
41.10	2698.0
42.10	1039.0
43.10	3792.0
44.00	5808.0
45.00	527.0
47.00	302.0
48.00	165.0
49.00	1068.0
49.90	287.0
50.90	495.0
51.90	168.0
52.90	328.0
54.00	503.0
55.10	3175.0
56.00	1420.0
57.10	2609.0
58.00	289.0
59.90	272.0
61.20	152.0
63.00	150.0
65.00	269.0
67.00	791.0
68.00	947.0
69.10	2397.0
70.10	1470.0
71.10	1090.0
73.00	739.0
77.00	401.0
78.10	249.0
79.00	563.0
81.00	636.0
82.10	770.0
83.10	2561.0
84.00	1918.0
85.00	704.0
86.00	607.0
91.00	961.0
91.90	293.0
93.00	229.0
95.00	274.0
96.00	666.0
97.00	1711.0
98.00	750.0
99.10	213.0
102.90	171.0
104.00	237.0
105.00	235.0
106.80	222.0
111.10	892.0
111.90	265.0
115.00	299.0
116.90	223.0
124.00	152.0
125.00	302.0
126.10	300.0
128.90	299.0
131.00	155.0
133.00	173.0
134.90	224.0
137.00	285.0
137.90	169.0
147.10	177.0
149.00	198.0
152.00	192.0
164.80	216.0
167.00	395.0
176.90	223.0
180.10	182.0
191.00	453.0
193.00	307.0
207.00	2141.0
207.90	476.0
208.90	381.0
266.90	159.0
281.10	659.0
282.00	174.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
35.90	230.0
37.00	1365.0
38.00	3748.0
39.10	25184.0
40.00	3235.0
41.00	1377.0
42.00	397.0
43.00	1226.0
44.00	3951.0
44.90	920.0
46.90	411.0
48.10	272.0
49.10	3047.0
50.10	84648.0
51.10	344384.0
52.10	17648.0
53.00	786.0
54.10	278.0
55.00	2548.0
56.00	9737.0
57.00	24512.0
58.00	1363.0
58.90	339.0
60.00	458.0
61.00	4069.0
62.00	4936.0
63.00	13792.0
64.00	2075.0
65.10	7335.0
66.00	637.0
67.00	850.0
68.10	5751.0
69.00	381696.0
70.00	2157.0
71.20	457.0
71.90	227.0
73.10	3134.0
74.10	38576.0
75.00	60672.0
76.10	22120.0
77.10	472128.0
78.10	33536.0
79.00	25504.0
80.00	20808.0
81.00	32800.0
82.00	8175.0
83.00	7786.0
84.00	1675.0
85.10	4838.0
86.00	8253.0
87.00	4022.0
88.00	1929.0
89.10	843.0
89.90	243.0
91.00	7464.0
92.00	7856.0
93.00	48656.0
94.00	3514.0
95.10	1142.0
96.10	2748.0
97.10	1244.0
98.00	36256.0
99.00	32320.0
100.00	3000.0
101.00	17568.0
102.00	1091.0
103.00	5830.0
104.00	10254.0
105.10	10822.0
106.00	3492.0
107.00	147776.0
108.00	22856.0
109.10	3628.0
110.00	264768.0
111.00	40392.0
112.00	5117.0
113.00	1496.0
113.90	342.0
114.80	615.0
115.10	611.0
116.10	6963.0
117.00	124400.0
118.00	9410.0
119.00	1272.0
120.00	1948.0
121.00	526.0
122.00	7871.0
123.00	14183.0
124.00	6675.0
125.00	6537.0
126.10	2870.0
127.00	525952.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

128.00	40320.0
129.00	195072.0
130.00	16488.0
131.00	3441.0
132.10	1412.0
132.90	653.0
134.00	5226.0
135.00	16160.0
136.10	5863.0
137.10	9518.0
138.00	1835.0
139.00	1043.0
140.00	2335.0
141.00	23728.0
142.00	9200.0
143.00	6705.0
144.10	1752.0
145.00	1516.0
146.00	4276.0
147.00	14140.0
148.00	35160.0
149.00	7123.0
150.00	2298.0
151.10	4057.0
151.80	2030.0
153.00	8754.0
154.00	6934.0
155.10	16528.0
156.00	23784.0
157.00	4669.0
158.00	5253.0
159.00	4049.0
160.00	9969.0
161.00	14778.0
162.00	4226.0
163.00	1408.0
164.00	1267.0
165.00	10402.0
166.00	8873.0
167.00	65560.0
168.00	38216.0
169.00	5788.0
170.00	1629.0
171.00	2563.0
172.00	5433.0
173.00	8104.0
174.10	13539.0
175.10	26624.0
176.00	7269.0
177.00	10760.0
178.00	3640.0
179.00	49928.0
180.00	34968.0
181.00	16848.0
182.00	2808.0
183.00	1415.0
184.00	3445.0
185.00	23752.0
186.00	195008.0
187.00	56984.0
188.10	6287.0
189.00	11100.0
190.00	1951.0
191.00	5851.0
192.00	17376.0
193.00	19616.0
194.00	4297.0
195.00	2410.0
196.00	43488.0
196.80	7415.0
198.00	1559552.0
199.00	110664.0
200.00	8562.0
201.50	6025.0
203.00	10042.0
204.00	57552.0
205.00	95392.0
206.10	400448.0
207.10	56488.0
208.00	13411.0
209.00	4942.0
210.10	5547.0
211.00	15067.0
213.00	1206.0
213.90	406.0
215.00	4382.0
216.00	8326.0
217.00	113504.0
218.00	15224.0
219.10	1468.0
221.00	64568.0
221.80	15756.0
223.00	23872.0
224.00	235712.0
225.00	59664.0
226.00	6504.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

227.00	93824.0
228.00	13533.0
229.00	20848.0
230.00	2312.0
231.00	8817.0
232.00	1592.0
233.00	1914.0
234.00	5469.0
235.00	6748.0
236.00	4449.0
237.00	7997.0
238.00	1168.0
239.00	3128.0
240.00	2806.0
241.00	4472.0
242.00	11922.0
243.10	12184.0
244.10	184768.0
245.10	23712.0
246.00	32616.0
247.00	7107.0
247.90	1694.0
249.00	6357.0
250.00	1422.0
251.00	1285.0
252.00	1811.0
253.00	3780.0
254.00	4786.0
255.00	889280.0
256.00	132992.0
257.00	11045.0
258.00	48440.0
259.00	8304.0
260.00	1309.0
261.00	1614.0
261.90	241.0
263.00	502.0
263.90	1445.0
265.00	19264.0
266.00	3410.0
267.10	517.0
268.00	247.0
270.00	861.0
271.00	2944.0
272.00	2463.0
273.00	26208.0
274.00	70080.0
275.00	373312.0
276.00	52152.0
277.00	32520.0
278.00	5698.0
279.00	1184.0
281.00	1034.0
282.10	820.0
283.00	3419.0
284.00	2470.0
285.00	5450.0
286.00	1081.0
288.00	281.0
289.00	1237.0
290.00	1248.0
291.10	746.0
292.10	1616.0
293.00	7154.0
294.00	1894.0
295.00	1837.0
296.00	108280.0
297.00	15484.0
298.00	1009.0
298.90	337.0
300.10	166.0
301.00	1313.0
302.00	1938.0
303.10	11835.0
304.10	3294.0
305.00	397.0
306.80	207.0
308.10	1519.0
309.00	995.0
310.10	1082.0
310.90	277.0
312.00	341.0
313.00	781.0
314.00	4501.0
315.00	11689.0
316.00	5856.0
317.00	1541.0
318.10	210.0
319.00	325.0
320.00	314.0
321.10	2992.0
322.10	1413.0
323.10	31464.0
324.10	6981.0
325.10	602.0
326.00	591.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

327.00	5825.0
328.10	3004.0
329.00	728.0
329.90	159.0
332.00	2312.0
333.00	2621.0
334.00	20712.0
335.00	5836.0
336.10	602.0
339.00	393.0
340.00	380.0
341.10	3782.0
342.00	1278.0
343.00	198.0
346.00	8695.0
347.00	1767.0
348.00	170.0
350.00	327.0
351.10	497.0
352.00	9413.0
353.10	7156.0
354.10	9825.0
355.10	1783.0
355.90	219.0
357.00	158.0
359.10	710.0
361.00	242.0
363.90	298.0
365.00	43968.0
366.00	6447.0
367.10	458.0
370.00	844.0
371.00	2562.0
372.10	15988.0
373.00	4067.0
374.00	485.0
376.90	386.0
383.10	3741.0
384.00	1413.0
385.00	503.0
390.10	2211.0
391.00	1394.0
392.10	1088.0
393.10	352.0
394.90	156.0
397.10	231.0
400.90	1053.0
402.10	6700.0
403.10	8933.0
404.00	3291.0
405.00	712.0
410.10	325.0
415.10	333.0
421.00	7394.0
422.00	8106.0
423.10	51280.0
424.10	12275.0
425.00	1167.0
438.00	354.0
439.10	207.0
440.00	199.0
441.10	152704.0
442.10	972992.0
443.10	201152.0
444.10	18848.0
445.00	1227.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
36.00	281.0
37.00	1538.0
38.10	4297.0
39.10	27720.0
40.00	2740.0
41.00	1219.0
42.00	336.0
43.00	863.0
44.00	3532.0
45.00	989.0
46.10	159.0
47.00	180.0
48.10	244.0
49.10	2917.0
50.10	99952.0
51.10	386816.0
52.10	20856.0
53.00	902.0
54.10	210.0
55.00	2076.0
56.10	10593.0
57.00	27384.0
58.00	995.0
59.10	308.0
60.10	520.0
61.00	4706.0
62.00	5187.0
63.10	15812.0
64.00	2508.0
65.10	8673.0
66.10	503.0
67.10	800.0
68.00	7044.0
69.00	427328.0
70.10	2591.0
71.10	355.0
72.00	276.0
73.00	3300.0
74.10	42424.0
75.00	70296.0
76.10	25272.0
77.10	536960.0
78.10	37376.0
79.00	30056.0
80.00	23144.0
81.00	35744.0
82.00	9022.0
83.00	8940.0
84.00	1280.0
85.10	5814.0
86.00	10164.0
87.00	4875.0
88.00	1738.0
89.10	1106.0
89.90	281.0
91.00	8123.0
92.00	8589.0
93.00	56888.0
94.00	4354.0
95.10	1341.0
96.10	2771.0
97.00	1303.0
98.00	43768.0
99.00	36760.0
100.00	3038.0
101.00	21120.0
102.10	1090.0
103.00	6513.0
104.00	11929.0
105.00	12123.0
106.00	4090.0
107.00	174912.0
108.10	27064.0
109.10	4561.0
110.00	315520.0
111.10	45608.0
112.00	6463.0
113.00	1813.0
114.00	507.0
115.10	808.0
116.10	8714.0
117.00	144384.0
118.00	10223.0
119.10	1463.0
120.00	2202.0
121.00	984.0
122.00	10894.0
123.00	16720.0
124.00	8264.0
125.00	8086.0
126.20	3560.0
127.10	638656.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

128.10	47880.0
129.00	238400.0
130.00	20688.0
131.00	4309.0
132.10	2126.0
133.00	1036.0
134.00	6490.0
135.00	18496.0
136.10	8276.0
137.10	11746.0
138.10	2050.0
139.00	1122.0
140.10	2749.0
141.00	31072.0
142.10	10333.0
143.00	7843.0
144.00	1899.0
145.00	1992.0
146.00	5443.0
147.00	17784.0
148.00	41672.0
149.10	8962.0
150.00	2230.0
151.10	4728.0
152.00	2645.0
153.00	11726.0
154.00	8413.0
155.10	20536.0
156.10	32208.0
157.00	6142.0
158.00	5857.0
159.00	5343.0
160.00	11169.0
161.10	18232.0
162.00	5180.0
163.00	1482.0
164.00	1688.0
165.00	12660.0
166.00	12688.0
167.00	85464.0
168.00	47912.0
169.00	6937.0
170.10	2245.0
171.00	3413.0
172.00	6500.0
173.00	10698.0
174.00	17944.0
175.10	34816.0
176.10	9079.0
177.00	14457.0
178.00	5123.0
179.00	64608.0
180.00	44792.0
181.00	21376.0
182.10	3855.0
182.90	1866.0
184.00	4619.0
185.10	33096.0
186.10	257600.0
187.00	75760.0
188.10	7201.0
189.00	14504.0
190.00	2410.0
191.00	7355.0
192.00	22880.0
193.10	25672.0
194.10	5595.0
195.10	2733.0
196.00	52552.0
196.90	10362.0
198.00	2105856.0
199.00	144832.0
200.00	10502.0
201.50	8718.0
203.10	13722.0
204.10	76568.0
205.10	132160.0
206.10	558464.0
207.10	77792.0
208.00	17056.0
209.00	5894.0
210.20	8745.0
211.10	21848.0
213.00	1339.0
214.00	595.0
215.00	5752.0
216.00	11836.0
217.00	155648.0
218.00	20896.0
219.00	1937.0
221.10	87600.0
221.80	21664.0
223.00	35120.0
224.10	337024.0
225.10	83768.0
226.10	8630.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

227.00	133824.0
228.00	18424.0
229.00	29184.0
230.00	3426.0
231.00	12450.0
232.00	2192.0
233.10	1975.0
234.00	7501.0
235.00	9307.0
236.00	6827.0
237.00	10414.0
238.00	1894.0
239.00	4689.0
240.00	3996.0
241.00	7742.0
242.10	16712.0
243.10	19128.0
244.10	265152.0
245.10	35152.0
246.00	45808.0
247.00	10100.0
248.00	2435.0
249.10	9420.0
250.00	1383.0
251.00	2059.0
252.10	2641.0
253.00	4849.0
254.10	6732.0
255.00	1302528.0
256.00	196864.0
257.00	15874.0
258.00	69320.0
259.00	11913.0
260.10	1990.0
261.10	2041.0
262.00	443.0
263.00	936.0
264.00	2079.0
265.00	29768.0
266.10	4222.0
267.00	567.0
268.00	241.0
269.00	255.0
270.00	1328.0
271.00	3598.0
272.00	3539.0
273.00	38544.0
274.00	103088.0
275.10	577600.0
276.10	78616.0
277.00	48440.0
278.00	7866.0
279.00	1704.0
280.00	295.0
281.10	1067.0
282.00	1221.0
283.00	5172.0
284.10	3855.0
285.10	8660.0
286.10	1505.0
286.90	205.0
288.00	553.0
289.10	1748.0
290.10	1425.0
291.00	1536.0
292.00	2314.0
293.10	12068.0
294.00	3241.0
295.00	2948.0
296.00	162112.0
297.10	22808.0
297.90	1358.0
299.00	340.0
301.00	2071.0
302.00	2828.0
303.10	19352.0
304.10	4731.0
305.00	713.0
306.10	250.0
307.00	202.0
308.10	2133.0
309.00	1272.0
310.00	2203.0
311.10	650.0
311.90	473.0
313.10	1344.0
314.10	7253.0
315.00	17312.0
316.10	10364.0
317.00	1578.0
318.00	346.0
318.90	357.0
320.00	642.0
321.10	5095.0
322.10	2531.0
323.10	52144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

324.10	9708.0
325.20	965.0
326.00	1205.0
327.00	10692.0
328.00	5242.0
329.10	953.0
330.00	368.0
330.90	207.0
332.10	3967.0
333.00	4796.0
334.10	34088.0
335.10	9115.0
336.10	1288.0
337.00	215.0
339.10	889.0
340.10	808.0
341.10	5792.0
342.20	1684.0
343.00	211.0
344.10	184.0
346.00	12707.0
347.10	2517.0
348.10	329.0
350.10	521.0
351.00	691.0
352.10	15326.0
353.10	11194.0
354.10	17216.0
355.10	3471.0
356.10	311.0
357.00	233.0
358.10	382.0
359.00	1152.0
360.00	376.0
361.00	440.0
364.00	348.0
365.00	75976.0
366.10	11146.0
367.00	696.0
370.00	1564.0
371.10	4596.0
372.10	28408.0
373.10	6545.0
374.10	822.0
377.10	829.0
383.10	7914.0
384.00	2045.0
385.10	700.0
390.10	3488.0
391.10	2352.0
392.10	2077.0
393.00	431.0
394.90	323.0
396.90	297.0
401.00	2037.0
402.10	11443.0
403.10	16680.0
404.10	5962.0
405.00	1027.0
407.90	203.0
410.00	466.0
411.10	176.0
415.00	724.0
416.00	268.0
420.10	262.0
421.10	13178.0
422.10	12383.0
423.10	93096.0
424.10	20376.0
425.00	1844.0
438.10	327.0
438.90	392.0
440.00	453.0
441.10	269632.0
442.10	1726464.0
443.10	348352.0
444.10	31872.0
445.10	2055.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

m/z	Abundance
35.90	347.0
37.00	1209.0
38.10	3257.0
39.10	22848.0
40.00	3395.0
41.10	1421.0
42.10	298.0
43.10	1188.0
44.00	4461.0
45.00	845.0
46.90	201.0
48.20	301.0
49.00	2388.0
50.10	80904.0
51.10	329984.0
52.10	16049.0
53.00	938.0
54.10	290.0
55.10	2565.0
56.00	9884.0
57.00	22344.0
58.00	1344.0
59.00	370.0
60.00	532.0
61.00	3661.0
62.00	5148.0
63.10	13093.0
64.10	1934.0
65.10	7731.0
66.10	629.0
67.00	763.0
68.00	5401.0
69.00	369600.0
70.00	2657.0
71.00	477.0
72.00	276.0
73.10	2947.0
74.10	35864.0
75.00	60248.0
76.10	20896.0
77.10	462080.0
78.10	31464.0
79.00	24816.0
80.00	20520.0
81.00	31272.0
82.00	7530.0
83.00	6375.0
84.00	1281.0
85.00	6543.0
86.00	8202.0
87.00	3959.0
88.00	1735.0
88.90	734.0
91.00	7020.0
92.00	7319.0
93.00	48832.0
94.00	3771.0
95.10	898.0
96.00	2378.0
97.10	976.0
98.00	38704.0
99.00	32088.0
100.00	2820.0
101.00	17952.0
102.00	1111.0
103.10	5219.0
104.00	9583.0
105.00	10326.0
106.00	3548.0
107.00	147776.0
108.00	22456.0
109.00	4038.0
110.00	273344.0
111.00	40240.0
112.00	4726.0
113.10	1753.0
113.90	412.0
115.10	597.0
116.10	7043.0
117.00	124760.0
118.00	8965.0
119.00	1231.0
120.10	1737.0
121.00	787.0
122.00	9079.0
123.00	14587.0
124.10	6811.0
125.00	6477.0
126.20	2472.0
127.00	547712.0
128.10	39632.0
129.00	200512.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

130.00	17160.0
131.00	3440.0
132.10	1811.0
133.00	931.0
134.00	5125.0
135.10	16712.0
136.00	7334.0
137.10	10090.0
138.00	1883.0
139.00	820.0
140.00	2834.0
141.00	27624.0
142.10	9261.0
143.00	6909.0
144.00	1747.0
145.00	1393.0
146.00	4691.0
147.00	14734.0
148.00	36920.0
149.00	8030.0
150.00	2245.0
151.10	4105.0
151.90	2105.0
153.00	8921.0
154.10	8192.0
155.10	18424.0
156.10	27264.0
157.10	4640.0
158.00	5209.0
159.00	4540.0
160.00	9838.0
161.00	14708.0
162.00	4312.0
163.10	1620.0
164.00	1599.0
165.00	11542.0
166.00	10098.0
167.00	71640.0
168.00	43136.0
169.00	6133.0
170.00	2035.0
171.00	2585.0
172.00	6037.0
173.00	8208.0
174.00	16139.0
175.10	29680.0
176.10	8120.0
177.00	12212.0
178.00	4116.0
179.00	54528.0
180.00	39616.0
181.00	19144.0
182.00	3341.0
183.00	1753.0
184.00	4620.0
185.10	28792.0
186.00	225536.0
187.00	63160.0
188.10	6920.0
189.00	12830.0
190.00	2745.0
191.00	6543.0
192.00	19656.0
193.00	23064.0
194.00	5089.0
195.00	2561.0
196.00	47320.0
196.90	8521.0
198.00	1856000.0
199.00	124296.0
200.00	9866.0
201.40	7823.0
203.00	12599.0
204.10	69568.0
205.10	112856.0
206.10	488576.0
207.10	64432.0
208.00	15600.0
209.00	5575.0
210.20	7787.0
211.00	18840.0
213.00	1621.0
214.10	428.0
215.00	5326.0
216.00	10898.0
217.00	140928.0
218.00	18192.0
219.00	1962.0
221.10	80984.0
221.90	19824.0
223.00	31760.0
224.10	294912.0
225.10	75200.0
226.00	7150.0
227.00	118592.0
228.00	15270.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

229.00	25840.0
230.00	3145.0
231.10	10744.0
232.10	1878.0
233.10	2169.0
234.00	8098.0
235.00	8825.0
236.00	6927.0
237.00	10724.0
238.00	1434.0
239.00	4837.0
240.00	3474.0
241.00	6608.0
242.00	15410.0
243.10	16648.0
244.10	238464.0
245.10	31696.0
246.00	43896.0
247.00	8980.0
248.00	2394.0
249.00	8669.0
250.00	1700.0
251.00	1731.0
252.00	1926.0
253.10	4133.0
254.10	6640.0
255.00	1211392.0
256.00	173440.0
257.00	12680.0
258.00	68480.0
259.00	10490.0
260.00	2131.0
261.00	2316.0
262.10	358.0
263.00	567.0
264.00	1491.0
265.00	27472.0
266.00	3670.0
267.00	667.0
268.00	441.0
269.00	229.0
270.00	897.0
271.00	3289.0
272.10	2969.0
273.00	38296.0
274.00	98120.0
275.10	544128.0
276.10	72336.0
277.00	46088.0
278.00	7488.0
279.10	1823.0
280.00	190.0
281.00	997.0
282.10	1246.0
283.10	5270.0
284.00	2783.0
285.10	7995.0
286.00	1432.0
287.10	187.0
288.00	476.0
289.10	1984.0
290.10	1510.0
291.00	1339.0
292.10	2291.0
293.00	11376.0
294.00	2601.0
295.00	2868.0
296.00	159232.0
297.00	21664.0
298.00	1382.0
298.90	284.0
300.00	302.0
301.10	2130.0
302.00	2934.0
303.10	18080.0
304.10	4960.0
305.10	826.0
306.10	237.0
307.10	182.0
308.00	1783.0
309.10	1777.0
310.00	2048.0
311.00	577.0
312.00	549.0
313.00	1314.0
314.10	7275.0
315.00	16784.0
316.10	10263.0
317.00	1835.0
318.00	190.0
319.10	361.0
320.00	467.0
321.10	5170.0
322.10	2366.0
323.10	51240.0
324.10	9458.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080

325.10	885.0
326.00	1231.0
327.00	11005.0
328.10	5560.0
329.00	861.0
329.90	247.0
331.00	368.0
332.10	3791.0
333.10	5025.0
334.10	35920.0
335.10	8643.0
336.00	937.0
339.00	792.0
340.10	944.0
341.10	5682.0
342.10	1330.0
343.00	255.0
346.00	13677.0
347.10	2327.0
348.10	296.0
350.10	364.0
351.00	821.0
352.10	16384.0
353.10	10515.0
354.10	18096.0
355.10	3150.0
356.00	339.0
357.10	237.0
358.00	316.0
359.10	1281.0
359.80	184.0
361.20	374.0
362.90	250.0
363.90	447.0
365.00	76984.0
366.10	11877.0
367.00	706.0
370.00	1693.0
371.10	4994.0
372.10	28864.0
373.10	7167.0
374.00	955.0
377.10	781.0
378.10	202.0
381.80	153.0
383.10	8221.0
384.10	2656.0
385.00	742.0
386.00	182.0
390.10	3807.0
391.10	2814.0
392.10	1948.0
393.10	337.0
395.00	364.0
396.00	217.0
397.00	506.0
401.00	2032.0
402.10	13459.0
403.10	17536.0
404.10	5863.0
405.10	1170.0
410.00	559.0
414.90	725.0
416.00	175.0
419.80	249.0
421.10	15880.0
422.10	13438.0
423.10	103928.0
424.10	21344.0
425.10	2242.0
438.10	536.0
439.00	482.0
439.90	292.0
441.10	299072.0
442.10	1925120.0
443.10	379392.0
444.10	33992.0
445.10	1458.0

Instrument :
BNA_M
ClientSampleId :
DFTPP080