

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122716\
 Data File : BM008676.D
 Acq On : 27 Dec 2016 16:39
 Operator : UM/SJ
 Sample : H5995-03
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA864

Manual Integrations
 APPROVED

Sohil
 12/30/2016 8:45:53 AM

Quant Time: Dec 29 12:50:51 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 12:05:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	420m	0.40	ng/ul	-0.20
2) Naphthalene-d8	10.40	136	3873m	0.40	ng/ul	-0.17
6) Acenaphthene-d10	14.33	164	8167m	0.40	ng/ul	-0.02
10) Phenanthrene-d10	17.00	188	157807m	0.40	ng/ul	-0.12
16) Chrysene-d12	21.26	240	8805m	0.40	ng/ul	-0.03
20) Perylene-d12	23.52	264	2179	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.01	152	1660m	0.28	ng/ul	-0.17
14) Fluoranthene-d10	19.10	212	68939m	0.16	ng/ul	-0.05
Target Compounds				Ovalue		
3) Naphthalene	10.45	128	166308m	15.26	ng/ul	
5) 2-Methylnaphthalene	12.08	142	316430m	43.15	ng/ul	
7) Acenaphthylene	14.01	152	24737	0.67	ng/ul#	61
8) Acenaphthene	14.35	153	41500	1.49	ng/ul	97
9) Fluorene	15.34	166	170788m	5.38	ng/ul	
12) Phenanthrene	17.09	178	155948m	0.32	ng/ul	
15) Fluoranthene	19.13	202	54089	0.09	ng/ul	73
17) Pyrene	19.50	202	86302m	2.54	ng/ul	
23) Benzo(a)pyrene	23.42	252	10010	1.26	ng/ul#	30
24) Indeno(1,2,3-cd)pyrene	25.78	276	3977	0.42	ng/ul#	82
26) Benzo(g,h,i)perylene	26.46	276	3858	0.47	ng/ul#	31

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122716\
Data File : BM008676.D
Acq On : 27 Dec 2016 16:39
Operator : UM/SJ
Sample : H5995-03
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA864

Manual Integrations
APPROVED

Sohil
12/30/2016 8:45:53 AM

Quant Time: Dec 29 12:50:51 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 12:05:37 2016
Response via : Initial Calibration

