

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122716\
 Data File : BM008683.D
 Acq On : 27 Dec 2016 21:05
 Operator : UM/SJ
 Sample : H5995-03DL 5X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA864DL

Manual Integrations
 APPROVED

Sohil
 12/30/2016 8:46:00 AM

Quant Time: Dec 29 12:48:44 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 12:05:37 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	344m	0.40	ng/ul	-0.19
2) Naphthalene-d8	10.40	136	1632m	0.40	ng/ul	-0.17
6) Acenaphthene-d10	14.30	164	2805	0.40	ng/ul	-0.05
10) Phenanthrene-d10	16.99	188	23252m	0.40	ng/ul	-0.14
16) Chrysene-d12	21.25	240	3169m	0.40	ng/ul	-0.05
20) Perylene-d12	23.48	264	1452m	0.40	ng/ul	-0.06
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.00	152	271m	0.11	ng/ul	-0.18
14) Fluoranthene-d10	19.08	212	3430m	0.05	ng/ul	-0.07
Target Compounds						
3) Naphthalene	10.45	128	28965m	6.31	ng/ul	
5) 2-Methylnaphthalene	12.07	142	53401m	17.28	ng/ul	
7) Acenaphthylene	14.00	152	3769	0.30	ng/ul#	34
8) Acenaphthene	14.33	153	6922	0.73	ng/ul	93
9) Fluorene	15.34	166	29397	2.70	ng/ul	98
17) Pyrene	19.47	202	14653m	1.20	ng/ul	
23) Benzo(a)pyrene	23.39	252	1842m	0.35	ng/ul	
24) Indeno(1,2,3-cd)pyrene	25.73	276	746m	0.12	ng/ul	
26) Benzo(g,h,i)perylene	26.43	276	702m	0.13	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122716\
Data File : BM008683.D
Acq On : 27 Dec 2016 21:05
Operator : UM/SJ
Sample : H5995-03DL 5X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA864DL

Manual Integrations
APPROVED

Sohil
12/30/2016 8:46:00 AM

Quant Time: Dec 29 12:48:44 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 12:05:37 2016
Response via : Initial Calibration

