

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
 Data File : BM049255.D
 Acq On : 27 Dec 2024 15:16
 Operator : RC/JU
 Sample : P5324-20
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YE644

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :mohammad ahmed 12/30/2024

Quant Time: Dec 27 22:12:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 26 11:35:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.543	152	3670	0.400	ng/ul	0.00
4) Naphthalene-d8	10.303	136	10895	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.181	164	6427	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.929	188	14938m	0.400	ng/ul	0.00
17) Chrysene-d12	21.137	240	12609m	0.400	ng/ul	0.00
23) Perylene-d12	23.285	264	11928m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.138	96	12493	2.952	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.903	152	3104	0.212	ng/ul	0.00
18) Fluoranthene-d10	18.968	212	7491	0.209	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.172	88	1390	0.281	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
Data File : BM049255.D
Acq On : 27 Dec 2024 15:16
Operator : RC/JU
Sample : P5324-20
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YE644

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/30/2024
Supervised By :mohammad ahmed 12/30/2024

Quant Time: Dec 27 22:12:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 26 11:35:38 2024
Response via : Initial Calibration

