

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
 Data File : BM049266.D
 Acq On : 27 Dec 2024 22:25
 Operator : RC/JU
 Sample : P5324-11
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YE5Z8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :mohammad ahmed 12/30/2024

Quant Time: Dec 27 23:06:48 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 26 11:35:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.544	152	4774	0.400	ng/u1	0.00
4) Naphthalene-d8	10.303	136	15200	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.177	164	9042	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.929	188	21138m	0.400	ng/u1	0.00
17) Chrysene-d12	21.137	240	17812m	0.400	ng/u1	0.00
23) Perylene-d12	23.288	264	16772m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.139	96	17097	3.105	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.909	152	4033	0.197	ng/u1	0.00
18) Fluoranthene-d10	18.968	212	10556	0.209	ng/u1	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.172	88	1594	0.248	ng/u1	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
Data File : BM049266.D
Acq On : 27 Dec 2024 22:25
Operator : RC/JU
Sample : P5324-11
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
YE5Z8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/30/2024
Supervised By :mohammad ahmed 12/30/2024


Quant Time: Dec 27 23:06:48 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 26 11:35:38 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.248 ng/ul
 RT: 3.172 min Scan# 41
 Delta R.T. -0.004 min
 Lab File: BM049266.D
 Acq: 27 Dec 2024 22:25

Instrument :
 BNA_M
 ClientSampleId :
 YE5Z8



Tgt Ion: 88 Resp: 1594

Ion	Ratio	Lower	Upper
88	100		
43	54.1	45.4	68.2
58	44.5	43.6	65.4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :mohammad ahmed 12/30/2024

