

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
 Data File : BM049282.D
 Acq On : 28 Dec 2024 08:33
 Operator : RC/JU
 Sample : P5325-04
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YE682

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :mohammad ahmed 12/30/2024

Quant Time: Dec 29 21:55:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 26 11:35:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.539	152	6672	0.400	ng/ul	0.00
4) Naphthalene-d8	10.299	136	21856	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.174	164	12953	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.926	188	29531m	0.400	ng/ul	0.00
17) Chrysene-d12	21.134	240	25409m	0.400	ng/ul	0.00
23) Perylene-d12	23.285	264	25497m	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.139	96	53011	6.889	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.899	152	13587	0.463	ng/ul	0.00
18) Fluoranthene-d10	18.965	212	31383m	0.435	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.172	88	829	0.092	ng/ul#	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM122724\
Data File : BM049282.D
Acq On : 28 Dec 2024 08:33
Operator : RC/JU
Sample : P5325-04
Misc :
ALS Vial : 43 Sample Multiplier: 1


Instrument :
BNA_M
ClientSampleId :
YE682

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/30/2024
Supervised By :mohammad ahmed 12/30/2024


Quant Time: Dec 29 21:55:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM121824.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 26 11:35:38 2024
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.092 ng/ul
 RT: 3.172 min Scan# 41
 Delta R.T. -0.004 min
 Lab File: BM049282.D
 Acq: 28 Dec 2024 08:33

Instrument :
 BNA_M
 ClientSampleId :
 YE682



Tgt Ion: 88 Resp: 829

Ion	Ratio	Lower	Upper
88	100		
43	61.0	45.4	68.2
58	26.5	43.6	65.4

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/30/2024
 Supervised By :mohammad ahmed 12/30/2024

