

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM122816\
 Data File : BM008696.D
 Acq On : 28 Dec 2016 16:38
 Operator : UM/SJ
 Sample : H6067-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DA7G6

Manual Integrations
 APPROVED

mohammad
 12/30/2016 8:53:30 AM

Quant Time: Dec 29 03:43:11 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 29 03:29:39 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	427m	0.40	ng/ul	-0.17
2) Naphthalene-d8	10.43	136	1223	0.40	ng/ul	-0.13
6) Acenaphthene-d10	14.28	164	830m	0.40	ng/ul	-0.07
10) Phenanthrene-d10	17.04	188	1612m	0.40	ng/ul	-0.07
16) Chrysene-d12	21.27	240	1500m	0.40	ng/ul	-0.01
20) Perylene-d12	23.50	264	1263m	0.40	ng/ul	-0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.06	152	530	0.28	ng/ul	-0.10
14) Fluoranthene-d10	19.10	212	1501m	0.34	ng/ul	-0.04
Target Compounds				Qvalue		
12) Phenanthrene	17.07	178	278m	0.06	ng/ul	
15) Fluoranthene	19.12	202	442m	0.07	ng/ul	
18) Benzo(a)anthracene	21.26	228	276m	0.06	ng/ul	
19) Chrysene	21.31	228	249m	0.05	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM122816\
Data File : BM008696.D
Acq On : 28 Dec 2016 16:38
Operator : UM/SJ
Sample : H6067-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DA7G6

Manual Integrations
APPROVED

mohammad
12/30/2016 8:53:30 AM

Quant Time: Dec 29 03:43:11 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM121916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 29 03:29:39 2016
Response via : Initial Calibration

