

Data Path : Z:\HPCHEM1\BNA M\Data\BM122915\
 Data File : BM003825.D
 Acq On : 28 Dec 2015 13:04
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 29 01:30:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.533	0.446	16.3	112	0.00
3	n-Nitrosodimethylamine	0.532	0.456	14.3	108	0.00
4 S	2-Fluorophenol	1.181	1.081	8.5	117	0.02
5 S	Phenol-d6	1.447	1.334	7.8	121	0.00
6	bis(2-Chloroethyl)ether	1.286	1.133	11.9	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	111	-0.02
8 S	Nitrobenzene-d5	0.265	0.231	12.8	116	0.00
9	Nitrobenzene	0.301	0.263	12.6	115	0.00
10	Naphthalene	1.142	1.045	8.5	115	0.00
11	Hexachlorobutadiene	0.176	0.164	6.8	117	0.00
12	2-Methylnaphthalene	0.762	0.700	8.1	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.145	0.115	20.7	119	0.00
15 S	2-Fluorobiphenyl	1.530	1.383	9.6	118	0.00
16	Acenaphthylene	2.275	2.051	9.8	125	0.00
17	Acenaphthene	1.479	1.268	14.3	112	0.00
18	Fluorene	1.754	1.493	14.9	117	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.281	0.260	7.5	112	0.00
21	Hexachlorobenzene	0.271	0.247	8.9	112	0.00
22	Pentachlorophenol	0.077	0.031	59.7#	70	0.00
23	Phenanthrene	1.689	1.531	9.4	113	0.00
24	Anthracene	1.235	1.068	13.5	115	0.00
25	Fluoranthene	1.605	1.302	18.9	109	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
27	Benzydine	0.268	0.460	-71.6#	142	0.00
28	Pvrene	1.751	1.810	-3.4	107	0.00
29 S	Terphenyl-d14	0.890	0.846	4.9	101	0.00
30	Benzo(a)anthracene	1.562	1.377	11.8	102	0.00
31	3,3'-Dichlorobenzidine	0.431	0.454	-5.3	109	0.00
32	Chrysene	1.561	1.343	14.0	85	0.00
33	Bis(2-ethylhexyl)phthalate	0.984	1.083	-10.1	126	0.00
34	Indeno(1,2,3-cd)pyrene	1.111	1.235	-11.2	117	0.01
35 I	Perylene-d12	1.000	1.000	0.0	95	0.01
36	Benzo(b)fluoranthene	1.634	1.406	14.0	90	0.00
37	Benzo(k)fluoranthene	1.507	1.383	8.2	104	0.00
38 C	Benzo(a)pyrene	1.399	1.319	5.7	105	0.00
39	Dibenzo(a,h)anthracene	1.097	1.192	-8.7	117	0.00
40	Benzo(a,h,i)perylene	1.148	1.280	-11.5	119	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM122915\
Data File : BM003825.D
Acq On : 28 Dec 2015 13:04
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 29 01:30:04 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				