

Data Path : Z:\HPCHEM1\BNA M\Data\BM122915\
 Data File : BM003825.D
 Acq On : 28 Dec 2015 13:04
 Operator : UM/SJ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Dec 29 01:30:04 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 24 10:08:49 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	0.00
2	1,4-Dioxane	1.000	0.926	7.4	112	0.00
3	n-Nitrosodimethylamine	1.000	0.857	14.3	108	0.00
4 S	2-Fluorophenol	1.000	0.915	8.5	117	0.02
5 S	Phenol-d6	1.000	0.922	7.8	121	0.00
6	bis(2-Chloroethyl)ether	1.000	0.881	11.9	111	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	111	-0.02
8 S	Nitrobenzene-d5	1.000	0.873	12.7	116	0.00
9	Nitrobenzene	1.000	0.873	12.7	115	0.00
10	Naphthalene	1.000	0.915	8.5	115	0.00
11	Hexachlorobutadiene	1.000	0.930	7.0	117	0.00
12	2-Methylnaphthalene	1.000	0.918	8.2	117	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	1.000	0.866	13.4	119	0.00
15 S	2-Fluorobiphenyl	1.000	0.904	9.6	118	0.00
16	Acenaphthylene	1.000	0.902	9.8	125	0.00
17	Acenaphthene	1.000	0.910	9.0	112	0.00
18	Fluorene	1.000	0.885	11.5	117	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	117	0.00
20	4-Bromophenyl-phenylether	1.000	1.221	-22.1	112	0.00
21	Hexachlorobenzene	1.000	1.201	-20.1	112	0.00
22	Pentachlorophenol	1.000	0.649	35.1#	70	0.00
23	Phenanthrene	1.000	1.120	-12.0	113	0.00
24	Anthracene	1.000	0.914	8.6	115	0.00
25	Fluoranthene	1.000	0.894	10.6	109	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
27	Benzydine	1.000	1.369	-36.9#	142	0.00
28	Pvrene	1.000	1.060	-6.0	107	0.00
29 S	Terphenyl-d14	1.000	1.017	-1.7	101	0.00
30	Benzo(a)anthracene	1.000	0.914	8.6	102	0.00
31	3,3'-Dichlorobenzidine	1.000	1.090	-9.0	109	0.00
32	Chrysene	1.000	0.871	12.9	85	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.144	-14.4	126	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	1.171	-17.1	117	0.01
35 I	Perylene-d12	5.000	5.000	0.0	95	0.01
36	Benzo(b)fluoranthene	1.000	0.860	14.0	90	0.00
37	Benzo(k)fluoranthene	1.000	0.918	8.2	104	0.00
38 C	Benzo(a)pyrene	1.000	0.943	5.7	105	0.00
39	Dibenzo(a,h)anthracene	1.000	1.086	-8.6	117	0.00
40	Benzo(a,h,i)perylene	1.000	1.115	-11.5	119	0.00

Data Path : Z:\HPCHEM1\BNA M\Data\BM122915\
Data File : BM003825.D
Acq On : 28 Dec 2015 13:04
Operator : UM/SJ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Dec 29 01:30:04 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM122415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 24 10:08:49 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				