

Data Path : Z:\HPCHEM1\BNA M\DATA\BM122916\
 Data File : BM008713.D
 Acq On : 29 Dec 2016 09:59
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.441

Quant Time: Dec 30 00:55:22 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-SIM-BM121916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 29 03:29:39 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
3	Naphthalene	1.125	1.052	6.5	116	0.00
4 SURR	2-Methylnaphthalene-d10	0.613	0.580	5.4	120	0.00
5	2-Methylnaphthalene	0.757	0.649	14.3	107	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.00
7	Acenaphthylene	1.809	1.987	-9.8	143	0.00
8 C	Acenaphthene	1.361	1.397	-2.6	132	0.00
9	Fluorene	1.553	1.490	4.1	122	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	125	0.00
11	Pentachlorophenol	0.124	0.113	8.9	104	0.00
12	Phenanthrene	1.221	1.256	-2.9	129	0.00
13	Anthracene	1.186	1.233	-4.0	135	0.00
14 SURR	Fluoranthene-d10	1.101	1.261	-14.5	143	0.00
15 C	Fluoranthene	1.498	1.683	-12.3	143	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	129	0.00
17	Pyrene	1.546	1.703	-10.2	144	0.00
18	Benzo(a)anthracene	1.313	1.369	-4.3	142	0.00
19	Chrysene	1.438	1.553	-8.0	137	0.00
20 I	Perylene-d12	1.000	1.000	0.0	134	0.00
21	Benzo(b)fluoranthene	1.576	1.529	3.0	132	0.00
22	Benzo(k)fluoranthene	1.537	1.510	1.8	130	0.00
23 C	Benzo(a)pyrene	1.461	1.502	-2.8	139	0.00
24	Indeno(1,2,3-cd)pyrene	1.741	1.711	1.7	129	0.00
25	Dibenzo(a,h)anthracene	1.391	1.347	3.2	127	0.00
26	Benzo(a,h,i)perylene	1.493	1.477	1.1	129	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0